



CL MICROMED PVT. LTD.

T-17, OKHLA INDUSTRIAL AREA PHASE-2, NEW DELHI-110 020

CIN : U74899DL1987PTC027941

TEL : 91-11-26387262, 26387263

FAX : 91-11-26387264

e-mail : info@clmicromed.com

www.clmicromed.com

TECHNICAL / SERVICE REPORT

Date : 24-09-2021.....

INSTALLATION WARRANTY PERIOD CALL SERVICE CONTRACT CALL CHARGED CALL COURTESY

Customer's Name : Fatima Hospital (Laboratory).....

Address : Mother Teresa Road, Padri Bagar
Gosakhpur (U.P.).....

Contact Person : Laboratory I/c...... Telephone : 0551-2281862.....

Instrument : Electrolyte Analyser.....

Model : COMBILINE..... Make : Eschweiler GmbH & Co...... Serial No.: CL 0030
Germany

Nature of Fault : Routine maintenance & calibration service......

Description of Job : Preventive maintenance carried out. Performed cleaning & activation
of all the sensors. Calibration done and sensor parameters found OK.
checked Q.C. results, found correct & precise. The equipment
working satisfactorily.

1. Fault Rectified... yes..... 2. Job to be continued... N/A..... 3. Preventive Maintenance... yes.....

PART REPLACED Form Our Stock From Customer's Inventory

Sr. No.	Description	Qty.
1.	Ref. Membrane (FOC)	1
2.	Set tubing	1

Engineer's Remarks..... /

Customer's Comments (if any).....

Instrument is working satisfactorily.

Engineers Name : Alok Verma

Contact No. : 9415021528

Engineer's Signature A.V.

Customer's Signature

Date :