

Patient Name: Mr. Sushant L

SPLIT TESTING RECORD

NORMAL SAMPLES

Date	Patients ID	Test Parameter	Expected Result	Observed Result	Acceptable CV %	Observed CV %	Comment	Section Head Sign/Date
27/12	46500	Quantity	Low	Low	NA	NA	Satisfactory	[Signature]
		Colour	Pale yellow	Pale yellow	NA	NA	Satisfactory	[Signature]
		Appearance	Clear	Clear	NA	NA	Satisfactory	[Signature]
		RH	6.5	6.5	NA	NA	Satisfactory	[Signature]
		SH	1.000	1.000	NA	NA	Satisfactory	[Signature]
		PROTEIN	Absent	Absent	NA	NA	Satisfactory	[Signature]
		SUGAR	Absent	Absent	NA	NA	Satisfactory	[Signature]
		KETONE	Absent	Absent	NA	NA	Satisfactory	[Signature]
		URICACID	Normal	Normal	NA	NA	Satisfactory	[Signature]
		OB	Negative	Negative	NA	NA	Satisfactory	[Signature]
		URICATE	Negative	Negative	NA	NA	Satisfactory	[Signature]
		BILE SALT	Negative	Negative	NA	NA	Satisfactory	[Signature]

Acceptable Criteria: As mentioned in QSP

PERFORMED BY: MANISHA [Signature] APPROVED BY: ANKITA [Signature] CHECKED BY: KARUNA [Signature]

Document No: GEN-69	Document Name: Split Testing Record	Status: Controlled
Issue No: 01	Issue Date: 01.02.2020	Page 1 of 1
Amendment No: 00	Amendment Date: NA	
ISSUED BY: QA DEPARTMENT		
APPROVED BY: HEAD OF DEPARTMENT		

SPLIT TESTING RECORD

Patient Name: Ms ANURADHA ANANDKAR

MS ANURADHA ANANDKAR

Date	Patients ID	Test Parameter	SPUT 1 Expected Result	SPUT 2 Observed Result	Acceptable CV %	Observed CV%	Comment	Section Head Sign/Date
2/7/21	46502	Quantity	20 ml	25 ml	NA	NA	Satisfactory	[Signature]
		Colour	yellow	yellow.	NA	NA	Satisfactory	[Signature]
		Appearance	cloudy	cloudy	NA	NA	Satisfactory	[Signature]
		PH	6	6	NA	NA	Satisfactory	[Signature]
		Sp	1.030	1.030	NA	NA	Satisfactory	[Signature]
		protein	Absent	Absent	NA	NA	Satisfactory	[Signature]
		Sugar	+++	+++	NA	NA	Satisfactory	[Signature]
		ketone	Present	Present	NA	NA	Satisfactory	[Signature]
		Urobilinogen	Normal	Normal	NA	NA	Satisfactory	[Signature]
		OB	Negative	Negative	NA	NA	Satisfactory	[Signature]
		Mixite	Negative	Negative	NA	NA	Satisfactory	[Signature]
		Bile salt.	Negative	Negative	NA	NA	Satisfactory.	[Signature]

Acceptable Criteria: As mentioned in QSP Performed by: MANIKA [Signature] And ANVITA [Signature]

Checked by: KANAKA [Signature]

Document No: GEN-69	Document Name: Split Testing Record	Status: Controlled
Issue No: 01	Issue Date: 01.02.2020	Page 1 of 1
Amendment No: 00	Amendment Date: NA	APPROVED BY: HEAD OF DEPARTMENT
ISSUED BY: QA DEPARTMENT		

July - 2021.

Normal.
Split - 1 - Mahira

DekaPHAN LAURA
Seq.No:0013
ID:SWAPNIL
COLOR:PALE-Y
CLARITY:CLEAR
08.07.2021 17:36

BLD NEG
LEU NEG
BIL NEG
UBG NORM
KET NEG
GLU NEG
PRO NEG
PH 6
NIT NEG
SG 1.030

Performed by
Mahi
8/07/2021

Normal
Split - 2 - Ankita

DekaPHAN LAURA
Seq.No:0014
ID:SWAPNIL-2
COLOR:PALE-Y
CLARITY:CLEAR
08.07.2021 17:39

BLD NEG
LEU NEG
BIL NEG
UBG NORM
KET NEG
GLU NEG
PRO NEG
PH 6
NIT NEG
SG 1.030

Performed by
AN
8/07/2021

Abnormal
Split 1 - mahira

DekaPHAN LAURA
Seq.No:0001
ID:ANURADHA
COLOR:YELLOW
CLARITY:CLOUDY
08.07.2021 16:49

BLD NEG
*LEU 25 Leu/ μ l 1+
BIL NEG
UBG NORM
*KET 5.2 mg/dl +-
*GLU 300 mg/dl 3+
PRO NEG
PH 6
NIT NEG
SG 1.030

Performed by
ANURADHA
8/07/2021

Abnormal
Split - 2 - Ankita

DekaPHAN LAURA
Seq.No:0006
ID:ANURADHA
COLOR:YELLOW
CLARITY:CLOUDY
08.07.2021 17:01

BLD NEG
*LEU 25 Leu/ μ l 1+
BIL NEG
UBG NORM
*KET 5.2 mg/dl +-
*GLU 1000 mg/dl 4+
PRO NEG
PH 6
NIT NEG
SG 1.030

Performed by
AN
08/07/2021

PERFORMED BY : MAHIRA

CHECKED BY : KARUNA

PERFORMED BY : ANKITA

CHECKED BY : KARUNA

Patient names : BHAVARWATI.

SPLIT TESTING RECORD

NORMAL SAMPLES

Date	Patients ID	Test Parameter	Split 1 Expected Result	Split 2 Observed Result	Acceptable CV %	Observed CV %	Comment	Section Head Sign/Date
10/8/21	56269	Quantity.	15 ml	15 ml	NA	NA	Satisfactory	Rohini
		Colour	Pale yellow	Pale yellow	NA	NA	Satisfactory	Rohini
		Appearance	Clear	Clear	NA	NA	Satisfactory	Rohini
		PH	5	5	NA	NA	Satisfactory	Rohini
		SG	1.030	1.030	NA	NA	Satisfactory	Rohini
		Protein	Absent	Absent	NA	NA	Satisfactory	Rohini
		Sugar	Absent	Absent	NA	NA	Satisfactory	Rohini
		Ketone	Absent	Absent	NA	NA	Satisfactory	Rohini
		Urobilinogen	Normal	Normal	NA	NA	Satisfactory	Rohini
		DB	Negative	Negative	NA	NA	Satisfactory	Rohini
		Nitrite	Negative	Negative	NA	NA	Satisfactory	Rohini
		Billi salt	Negative	Negative	NA	NA	Satisfactory.	Rohini

Acceptable Criteria: As mentioned in QSP

Performed by : Ankita Mahina

Checked by : Kamna Jadhav

Document No: GEN-69	Document Name: Split Testing Record	Status: Controlled
Issue No: 01	Issue Date: 01.02.2020	Page 1 of 1
Amendment No: 00	Amendment Date: NA	
ISSUED BY : QA DEPARTMENT		APPROVED BY : HEAD OF DEPARTMENT

10/8/21

AUGUST-2021.

NORMAL

SPLIT - 1

DekaPHAN LAURA
Seq.No:0016
ID:BHAVARNATH-1
COLOR:PALE-Y
CLARITY:CLEAR
10.08.2021 11:00

BLD NEG
LEU NEG
BIL NEG
UBG NORM
KET NEG
GLU NEG
PRO NEG
pH 5
NIT NEG
SG 1.030

PERFORMED BY
AN
10/8/21

CHECKED BY
Ramesh
10/8/21

NORMAL

SPLIT-2

DekaPHAN LAURA
Seq.No:0017
ID:BHAVARNATH-2
COLOR:PALE-Y
CLARITY:CLEAR
10.08.2021 11:02

BLD NEG
LEU NEG
BIL NEG
UBG NORM
KET NEG
GLU NEG
PRO NEG
pH 5
NIT NEG
SG 1.030

PERFORMED
By
AN
10/8/2021

CHECKED BY
Ramesh
10/8/21

ABNORMAL

SPLIT - 1

DekaPHAN LAURA
Seq.No:0018
ID:HIRAMANI
COLOR:PALE-Y
CLARITY:TURBID
10.08.2021 11:05

*BLD 50 Ery/ μ l 2+
*LEU 500 Leu/ μ l 3+
BIL NEG
UBG NORM
KET NEG
*GLU 50 mg/dl 1+
*PRO 100 mg/dl 2+
pH 6
NIT NEG
SG 1.015

PERFORMED BY
AN
10/8/21

CHECKED BY
Ramesh
10/8/21

ABNORMAL

SPLIT-2

DekaPHAN LAURA
Seq.No:0019
ID:HIRAMANI-2
COLOR:PALE-Y
CLARITY:TURBID
10.08.2021 11:08

*BLD 50 Ery/ μ l 2+
*LEU 500 Leu/ μ l 3+
BIL NEG
UBG NORM
KET NEG
*GLU 50 mg/dl 1+
*PRO 100 mg/dl 2+
pH 6
NIT NEG
SG 1.015

PERFORMED BY
AN
10/8/2021

CHECKED BY
Ramesh
10/8/21