

### CALIBRATION CERTIFICATE

NAME OF CUSTOMER:

DATE: 25.03.2021

**ABIRAMI KIDNEY CARE**  
NO: 580, 581, BROUGH ROAD,  
NEAR RAVI THEATRE,  
ERODE-638 011.

DESCRIPTION OF THE EQUIPMENT: CENTRIFUGE

MODEL: R8C - (REMI)

SL.NO: 06887

INSTRUMENT USED FOR CALIBRATION: DIGITAL TACHOMETER

MODEL NO:KUSAM-MECO&amp;KM-2234BL

SL.NO: S1049947

CERTIFIED BY: MSIR INDIA PVT.LTD

INSTRUMENT VALIDITY: 22-02-2022

SL.NO.	DISPLAY RPM	TACHOMETER READING	ALLOWABLE ERROR	ERROR OBSERVED ±	RESULT
1.	1000	1014	±100	14	OK
2.	2500	2459	±100	41	OK
3.	3000	3044	±100	44	OK
4.	4500	4478	±100	26	OK

NEXT CALIBRATION DUE ON: 24.03.2022

TESTED BY: G DINESH

APPROVED BY: G DINESH



NOTE: 1) Calibration Points & due date is given as per customer request 2) This certificate refer only to particular item submitted for certificate 3) The calibration results reported are valid at the time of and under the stated conditions of the measurements 4) Any error in the certificate should be brought to our lab within 45 days from the date of issue of certificate.