

## VIEW LAB MONTHLY SUMMARY

Lab Name	ABIRAMI KIDNEY CARE PVT LTD	Details About Robust Analysis
Lab No	7711	Detail About Monthly Summary
Month	March	Detail about SDI
Year	2021	
Constituent Group	Chemistry I	

Click on the analyte to view Graphical Data

[All Analyser Result](#)[Print](#)

Date of Result Entered : 15/03/2021

Date of Report Published : 01/04/2021

S/No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE	Hexokinase	Roche C111, C311, Integra 400 plus, C501, C502, C702	277	172.03	3.90	6.71	180.9 mg/dl	1.32	0.81
2	UREA	Urease UV / GLDH	Roche C111, C311, Integra 400 plus, C501, C502, C702	292	120.99	5.31	6.42	121.8 mg/dl	0.13	0.75
3	CREATININE	Enzymatic Colorimetric	Roche C111, C311, Integra 400 plus, C501, C502, C702	37	6.23	5.19	0.32	6.7 mg/dl	1.46	0.11
4	T.BILIRUBIN	Diazonium salt ( Colorimetric ) / Jendrassik	Roche C111, C311, Integra 400 plus, C501, C502, C702	272	2.60	5.46	0.14	2.5 mg/dl	-0.70	0.02
5	T-PROTEIN	Biuret - Colorimetric	Roche C111, C311, Integra 400 plus, C501, C502, C702	280	4.90	4.72	0.23	4.5 g/dl	-1.73	0.03
6	ALBUMIN	BCG - colorimetric	Roche C111, C311, Integra 400 plus, C501, C502, C702	267	3.19	4.41	0.14	3.2 g/dl	0.07	0.02
7	CALCIUM	5-Nitro-5-methyl-BAPTA	Roche C111, C311, Integra 400 plus, C501, C502, C702	217	11.51	3.94	0.45	11.7 mg/dl	0.42	0.06
8	PHOSPHORUS	Molybdate UV / Phosphomolybdate complex	Roche C111, C311, Integra 400 plus, C501, C502, C702	205	3.41	4.49	0.15	3.4 mg/dl	-0.07	0.02
9	URIC ACID	Enzymatic / Uricase Colorimetric	Roche C111, C311, Integra 400 plus, C501, C502, C702	273	6.23	4.51	0.28	6.3 mg/dl	0.25	0.03
10	CHOLESTEROL	CHOD-PAP	Roche C111, C311, Integra 400 plus, C501, C502, C702	291	109.14	4.11	4.49	112.2 mg/dl	0.68	0.53
11	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Roche C111, C311, Integra 400 plus, C501, C502, C702	258	181.57	4.11	7.46	177.1 mg/dl	-0.60	0.93
12	HDL CHO	Direct method / Enzymatic colorimetric	Roche C111, C311, Integra 400	255	19.93	5.72	1.14	19.9 mg/dl	-0.03	0.14

			plus, C501, C502, C702							
13	SODIUM	ISE - Indirect	Roche C111, C311, Integra 400 plus, C501, C502, C702	310	131.24	3.18	4.18	133 mmol/L	0.42	0.47
14	POTASSIUM	ISE - Indirect	Roche C111, C311, Integra 400 plus, C501, C502, C702	312	3.92	4.15	0.16	3.9 mmol/L	-0.12	0.02
15	CHLORIDE	ISE - Indirect	Roche C111, C311, Integra 400 plus, C501, C502, C702	244	100.59	4.10	4.12	96 mmol/L	-1.11	0.53
16	BICARBONATE	Dry Chemistry	Fuji Dry Chemistry series	25	16.68	16.05	2.68	20.6 mmol/L	1.46	1.07
17	AST	UV kinetic with PLP ( P-5-P )	Roche C111, C311, Integra 400 plus, C501, C502, C702	106	161.69	7.42	11.99	167 U/L	0.44	2.33
18	ALT	UV kinetic with PLP ( P-5-P )	Roche C111, C311, Integra 400 plus, C501, C502, C702	92	88.38	8.51	7.52	112.9 U/L	3.26	1.57
19	ALP	PNP AMP kinetic	Roche C111, C311, Integra 400 plus, C501, C502, C702	249	196.74	7.29	14.33	198 U/L	0.09	1.82

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between $\pm 1.0$ to $\pm 2.0$	Good.
Between $\pm 2.0$ to $\pm 3.0$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

#### Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

## LAB MONTHLY SUMMARY



Lab Name ABIRAMI KIDNEY CARE PVT LTD Lab No 7711  
 Month March Year 2021  
 Constituent Group HbA1c



Date of Result Entered : 15/03/2021

Date of Report Published : 01/04/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participant's CV	SD	Your Value	SDI	U
1	HbA1c	Turbidimetric Inhibition Immunoassay	Roche C111, C311, Integra 400 plus, C501, C502, C702	103	5.41	6.73	0.36	5.2 %	-0.58	0.07

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between $\pm 1.0$ to $\pm 2.0$	Good.
Between $\pm 2.0$ to $\pm 3.0$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss  
 CMC EQAS Co-Ordinator

Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*

## LAB MONTHLY SUMMARY



Lab Name **ABIRAMI KIDNEY CARE PVT LTD** Lab No **7711**  
 Month **March** Year **2021**  
 Constituent Group **Urine Chemistry**



Date of Result Entered : 15/03/2021

Date of Report Published : 01/04/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	UR.UREA	Ur. Urease UV / GLDH	Roche C111, C311, Integra 400 plus, C501, C502, C702	47	1,778.59	6.51	115.86	1787.5 mg/dL	0.08	33.80
2	UR.CREATININE	Ur. Enzymatic Colorimetric	Any Analyser (Automation / Semi Automation)	29	106.73	6.27	6.70	102.6 mg/dL	-0.62	2.49
3	UR.CALCIUM	Ur. 5-Nitro-5-methyl-BAPTA	Roche C111, C311, Integra 400 plus, C501, C502, C702	47	10.12	13.10	1.33	11.2 mg/dL	0.81	0.39
4	UR.PHOSPHORUS	Ur. Molybdate UV / Phosphomolybdate Complex	Roche C111, C311, Integra 400 plus, C501, C502, C702	46	14.08	6.97	0.98	15.7 mg/dL	1.65	0.29
5	UR.URIC ACID	Ur. Enzymatic / Uricase Colorimetric	Roche C111, C311, Integra 400 plus, C501, C502, C702	51	17.16	7.77	1.33	8.2 mg/dL	-6.72	0.37
6	UR.SODIUM	Ur. ISE Indirect	Roche C111, C311, Integra 400 plus, C501, C502, C702	37	94.52	3.05	2.88	96 mmol/L	0.51	0.95
7	UR.POTASSIUM	Ur. ISE Indirect	Roche C111, C311, Integra 400 plus, C501, C502, C702	37	36.12	4.78	1.72	36.3 mmol/L	0.10	0.57

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between $\pm 1.0$ to $\pm 2.0$	Good.
Between $\pm 2.0$ to $\pm 3.0$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)

Contact Number: 0416-2283102

Dr. Pamela Christudoss  
 CMC EQAS Co-Ordinator  
 Christian Medical College,  
 Vellore

\*\*\*\*\* End of Report \*\*\*\*\*