

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN. 33AAHCA8053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140

CALIBRATION CERTIFICATE



Customer Name & Address:	M/S. New Green Lab& Biopsy Centre, #1/786, Bharathi Nagar 7 th Street, Ramanathapuram – 623 503	Certificate No:	AMS/LAB/2889808
		Issuing Date:	25/09/2021
		Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Bio Chemistry Analyser	Id. No:	NGC/LAB /BCA/001
Make:	MISPA VIVA	Department:	Laboratory
Model:	V1.190907	Calibrated On:	24/09/2021
Sl. No:	212030355476	Calibration Due Suggested:	23/09/2022
Type:	Semi Auto	DUC Condition on Receipt:	OK
Range & Least Count:	---		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electrical Safety Analyser	RigelSafeTest60	PSG/CIRD/BMCC/5507/2021	18/01/2022
2	Thermo Hygrometer	Beetech/088	PC/01/21/ATL/0194-01	22/01/2022

Room Temperature: 25°C±2 Room Humidity: 30-60% (At the time of Calibration)

ESA Test Results:

SL.No	Parameter	ESA Reading	Unit	Normal Range	Condition
1	Current	0.228	mA	500	Normal
2	Frequency	50.1	Hz	49 -50	Normal
3	Mains Voltage Line - Neutral	225.6	Volts	220 - 240	Normal
4	Mains Voltage Line - Earth	114.9	Volts	220 - 240	Abnormal
5	Mains Voltage Neutral - Earth	111.2	Volts	< 7	No Earth

*Electrical Safety: Please check the attached log sheet

SL.No	Parameter	Measured	Standard	Correction	Unit
1	Lamp Voltage	11.83	12.00	-0.17	Volts
2	Detector Sensitivity	-0.00	0.0	0.0	% Absorbance
3	Sipper Volume	365	400	-35	µl
4	Peristaltic Pump pressure	-13.69	-15.00	+1.31	cmH ₂ O

Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:
(Calibration Engineer)

B. S. S.



Approved by:
(Technical Manager)

B. S. S.