


CALIBRATION CERTIFICATE

NAME OF CUSTOMER:

SRI VIRUDHAI MEDICAL CENTRE,
 103/49, SOUTHFORT STREET,
 VIRUDHACHALAM-606001
 PH: 04143231000, CELL: 9344865129

DESCRIPTION OF THE EQUIPMENT: CENTRIFUGE
MODEL: C854/8 - (REMI)
SL.NO: 5688
INSTRUMENT USED FOR CALIBRATION: DIGITAL TACHOMETER
MODEL NO:KUSAM-MECO&KM-2234BL
SL.NO: S1049947
CERTIFIED BY: MSIR INDIA PVT.LTD
INSTRUMENT VALIDITY: 22-02-2022

SL.NO.	UNIT RPM	TACHOMETER READING	ALLOWABLE ERROR	ERROR OBSERVED ±	RESULT
1.	1000	1014	±100	14	OK
2.	2500	2459	±100	41	OK
3.	3500	3471	±100	79	OK

MAXIMUM SPEED: 3500 RPM
NEXT CALIBRATION DUE ON: 01.08.2022
TESTED BY: G DINESH
APPROVED BY: G DINESH


NOTE: 1) Calibration Points & due date is given as per customer request 2) This certificate refer only to particular item submitted for certificate 3) The calibration results reported are valid at the time of and under the stated conditions of the measurements 4) Any error in the certificate should be brought to our lab within 45 days from the date of issue of certificate.