


# BHAGWATI SCIENTIFIC

SHOP NO. 5537, PLOT NO. 23 BASTI HARPHOOL SINGH  
SADAR THANA ROAD, DELHI-110006  
Mob No. +91-8800995541

CUSTOMER SERVICE REPORT		
CSR NO.	DATE: 19-May-2021	
Customer Name: DHAM DIAGNOSTIC CENTRE		
Address: No, 33 S6/23, Arya Samaj Road Koga Pura, New Delhi-110005		
City: New Delhi	State:	ZIP Code: 110005
Status of Call: Warranty or AMC (Please circle)	Instruction From Dr. Mr. Jaiprakash Kaur On	
NATURE OF PROBLEM		
Problem Reported: RBC & PLT Issue @ check the continuity RBC Bath. found faulty.		
System Down: Yes / No (Please circle)		Equipment Type:
Make: MINDRAY	Model: BC3600	Serial No. TB53000364
Call Reported by: MR SHIVAM	Date:	Time:
SERVICE DETAILS		
Service Rendered:		
Engineer's Remarks: Repairing RBC Bath. Check the continuity. Machine is working ok.	Status after Service: Machine is working fine & satisfactory.	
Events: (Date & Time)	Start of Service:	End of service:
CUSTOMER FEEDBACK		
Remarks: Machine is OK & under observation.		
Name: SHIVAM	Designation: Service Tech	Phone/Fax: 8800995541
Email:		
Signature: 	Date: 28 May-2021	Place:





# Dr. Dham's Diagnostic Centre

A Specialist Centre

3856/23 Reghar Pura, Arya Samaj Road, Karol Bagh,  
New Delhi-110005 Ph. : 011-25852890, 25852891

Date	15/09/2021	Srl No.	55	Ref. No	
Name	MRS. SARIKA	Age	38 Yrs.	Sex	F
Ref. By	Dr. S.R.ELAWADI	Address	#064/37 REHGAR PURA KAROL BAGH		

Test Name	Value	Unit	Normal Value
<b>COMPLETE HAEMOGRAM</b>			
HAEMOGLOBIN (Hb)	12.3	gm/dl	11.5 - 16.5
TLC	9300	/cumm	4000 - 11000
<b>DIFFERENTIAL LEUCOCYTE COUNT (DLC)</b>			
NEUTROPHIL	75	%	40 - 75
LYMPHOCYTE	20	%	20 - 45
EOSINOPHIL	05	%	1 - 6
MONOCYTE	00	%	1 - 3
BASOPHIL	00	%	0 - 01
ESR (WESTEGREN's METHOD)	12	mm/1st hr.	0 - 20
R B C COUNT	4.10	Millions/cmm	3.8 - 4.8
P.C.V / HAEMATOCRIT	38.4	%	35 - 45
M C V	91.7	fl.	80 - 100
M C H	29.4	Picogram	27.0 - 31.0
M C H C	32.0	gm/dl	33 - 37
PLATELET COUNT	3.60	Lacs/cmm	1.50 - 4.50

\*\*\*\* End Of Report \*\*\*\*

  
Dr. JASPREE KAUR

MBBS, MD (DMC No. 15396)

Consultant Biochemist

# GAGAN PATHOLOGY & IMAGING PVT. LTD.

(Previously Gagan Pathology Centre)

**NABH (MIS) Accredited CT / MRI Imaging**

MRI, CT, ULTRASOUND, ECHO, TMT, EEG, EMG  
MAMMOGRAPHY, NCV, DEXA SCAN, OPG, ECG  
PFT, BERA, VEP, DIGITAL & PORTABLE X-RAYS

Date 15/09/2024

Lsh Report  
SN No. 67

Ref. No

Name MRS. MRS. SARIKA

Age 35 Yrs.

Sex F

Ref. By SELF

Address

Test Name	Value	Unit	Normal Value
<b>COMPLETE HAEMOGRAM</b>			
HAEMOGLOBIN (Hb)	12.3	gm/dl	11.5 - 16.5
TLC	9100	/cumm	4000 - 11000
<b>DIFFERENTIAL LEUCOCYTE COUNT (DLC)</b>			
NEUTROPHIL	72	%	40 - 75
LYMPHOCYTE	24	%	20 - 45
EOSINOPHIL	04	%	1 - 6
MONOCYTE	00	%	1 - 3
BASOPHIL	00	%	0 - 01
R B C COUNT	4.00	Millions/cmm	3.8 - 4.8
P.C.V / HAEMATOCRIT	36.0	%	35 - 45
M C V	88.2	fl.	80 - 100
M C H	27.4	Picogram	27.0 - 31.0
M C H C	32.0	gm/dl	33 - 37
PLATELET COUNT	3.40	Lacs/cmm	1.50 - 4.50

\*\*\*\* End Of Report \*\*\*\*

**Dr. R.K. Garg**

MBBS, MD  
Senior Consultant Pathologist

Page 1 of 1

**Dr. Ankur Garg**

MBBS, MD  
Consultant Pathologist

*Expert and Reliable Opinion for over 30 Years*

H.O. : F-26/21-22, Sector-7, Rohini, Delhi-110085 Email : gaganlab@yahoo.com website : www.gaganlab.com