

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN. 33AAHCA8053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140

CALIBRATION CERTIFICATE



Customer Name & Address:	M/S. New Green Lab& Biopsy Centre, #1/786, Bharathi Nagar 7 th Street, Ramanathapuram – 623 503	Certificate No:	AMS/LAB/2889815
		Issuing Date:	25/09/2021
		Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Micro Pipette	Id. No:	NGC/LAB /VMP/005
Make:	Dragon Lab	Department:	Laboratory
Model:	Variable	Calibrated On:	24/09/2021
Sl. No:	YE19BAL0612848	Calibration Due Suggested:	23/09/2022
Type:	Variable	DUC Condition on Receipt:	OK
Range & Least Count:	100 to 1000µl		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electronic Weighing Scale	EWC/SF400C	PC/01/21/ATL/0339-01	22/01/2022
2	Thermo Hygrometer	Beetech/088	PC/01/21/ATL/0194-01	22/01/2022

Room Temperature: 25°C±2 Room Humidity: 30-60% (At the time of Calibration)


Test Results:

SL. No.	DUC Reading (µl)	Standard Reading (µl)		Correction (µl)
		Trial 1	Trial 2	
1	100.0	100.0	100.0	+0.0
2	200.0	200.0	200.0	+0.0
3	500.0	500.0	500.0	+0.0
4	800.0	800.0	800.0	+0.0
5	1000.0	1020.0	1020.0	+20.0

Allowable Tolerance: User Defined

Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.



Calibrated By:
(Calibration Engineer)

B. Jey.



Approved by:
(Technical Manager)

B. Jey.