

Lab Monthly Summary

Lab Name	DR PATH EXPERT DIAGNOSTIC PVT LTD	Lab No	14839
Month	June	Year	2021
Constituent Group	Chemistry II		

Date of Result Entered : 20/06/2021

Date of Report Published : 03/07/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Mindray	146	197.88	10.47	20.72	219.3 mg/dl	1.03	3.43
2	UREA II	Urease UV / GLDH II	Mindray	114	77.16	9.25	7.14	76.37 mg/dl	-0.11	1.34
3	CREATININE II	Jaffes End point II	Any Analyser (Automation / Semi Automation)	583	4.32	13.82	0.60	4.1 mg/dl	-0.37	0.05
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Mindray	129	3.86	15.41	0.59	4.3 mg/dl	0.74	0.10
5	T-PROTEIN II	Biuret - Colorimetric II	Mindray	125	5.31	8.52	0.45	6.12 g/dl	1.79	0.08
6	ALBUMIN II	BCG - Colorimetric II	Mindray	130	3.29	10.15	0.33	3.31 g/dl	0.06	0.06
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Mindray	115	4.59	22.58	1.04	4.98 mg/dl	0.38	0.19
8	CHOLESTROL II	CHOD-PAP II	Mindray	134	107.35	10.66	11.44	94.38 mg/dl	-1.13	1.98
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Mindray	128	139.58	10.12	14.12	115.2 mg/dl	-1.73	2.50

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcevellore.ac.in

Contact Number: 0416-2283102