

Catalog

PIPETTE BILL.pdf	1
pipette cal sheet.pdf	2

Tax Invoice

(ORIGINAL FOR RECIPIENT)

SKYMED SYSTEMS 37/A, New No.11/1 Veerapandian Street Vysarpadi, Chennai-600039 PH:9003287853/8608337744 GSTIN/UIN: 33ADVFS8854J1Z5 State Name : Tamil Nadu, Code : 33 E-Mail : skymedsys@gmail.com	Invoice No.	Dated
	SS/2794/21-22	4-Oct-2021
	Delivery Note	Mode/Terms of Payment
	Supplier's Ref.	Other Reference(s)
	SS/2794/21-22	
	Buyer's Order No.	Dated
Despatch Document No.	Delivery Note Date	
Despatched through	Destination	
Terms of Delivery		

Buyer
Sri Virudhai Medical Centre
 103/49, South Fort Street
 Virudhachalam- 606 001
 9344865129
 GSTIN/UIN : 33AQAPR3539D1Z4
 PAN/IT No :
 State Name : Tamil Nadu, Code : 33

Sl No.	Description of Goods	HSN/SAC	Quantity	Rate	per	Disc. %	Amount	
1	Dragon Pipette (5ul - 50ul)	8479	1 no	1,600.00	no		1,600.00	
2	Micropipette Variable 100-1000ul Dragon	84798999	1 no	1,600.00	no		1,600.00	
3	Dargon Pipette (10- 100ul)	8479	1 no	1,600.00	no		1,600.00	
							4,800.00	
							SGST	432.00
							CGST	432.00
Total			3 no				₹ 5,664.00	

Amount Chargeable (in words)

INR Five Thousand Six Hundred Sixty Four Only

E. & O.E

HSN/SAC	Taxable Value	Central Tax		State Tax		Total Tax Amount
		Rate	Amount	Rate	Amount	
8479	3,200.00	9%	288.00	9%	288.00	576.00
84798999	1,600.00	9%	144.00	9%	144.00	288.00
Total	4,800.00		432.00		432.00	864.00

Tax Amount (in words) : **INR Eight Hundred Sixty Four Only**

Company's PAN : **ADVFS8854J**

Declaration
 We declare that this invoice shows the actual price of the goods described and that all particulars are true and correct.

Company's Bank Details

Bank Name : **ALLAHABAD BANK**

A/c No. : **50479831585**

Branch & IFS Code : **Jawahar Nagar & IDIB000J587**

for SKYMED SYSTEMS

Authorised Signatory



QUALITY CONTROL CERTIFICATE
QUALITÄTSKONTROLLZERTIFIKAT
CERTIFICAT DE CONTRÔLE QUALITÉ
CERTIFICADO DE CONTROL DE CALIDAD
CERTIFICATO DI CONTROLLO QUALITÀ

Pipette 10-100ul

Serial Number/Serien-Nr/N° de série/Número de serie/Numero di Soric: YE19BAM0061042

Mode/Programm/Modo/Funzione: P Volume/Volumen: 25.00 ul
Mean/Mittelwert/Moyenne/Significado/Medio: 25.11 ul
Inaccuracy/Unrichtigkeit/Erreur de justesse/Inexatitud/Inaccuratezza: 0.42 %
Imprecision/Unpräzision/Erreur de répétabilité/Imprecisión/Imprecisione: 0.36 %

Date/Datum/Date/Fecha/Date: 2019-11-26 13:42:05
According to/Nach/Selon/Según/In base a: ISO8655 DIN 12650; 21 °C

Tested by/Prüfer/Testée par/Comprobado por/Testado da: May

CE ISO 9001/13485
CERTIFIED

QUALITY CONTROL CERTIFICATE
QUALITÄTSKONTROLLZERTIFIKAT
CERTIFICAT DE CONTRÔLE QUALITÉ
CERTIFICADO DE CONTROL DE CALIDAD
CERTIFICATO DI CONTROLLO QUALITÀ

Pipette 5-50ul

Serial Number/Serien-Nr/N° de série/Número de serie/Numero di Soric: YE217AS0257085

Mode/Programm/Modo/Funzione: P Volume/Volumen: 20.00 ul

Mean/Mittelwert/Moyenne/Significado/Medio: 20.11 ul

Inaccuracy/Unrichtigkeit/Erreur de justesse/Inexatitud/Inaccuratezza: 0.54 %

Imprecision/Unpräzision/Erreur de répétabilité/Imprecisión/Imprecisione: 0.27 %

Date/Datum/Date/Fecha/Data: 2021-07-21 15:07:38

According to/Nach/Selon/Según/In base a: ISO8655 DIN 12650; 22°C

Tested by/Prüfer/Testée par/Comprobado por/Testado da: Lily

CE ISO 9001/13485
CERTIFIED

QUALITY CONTROL CERTIFICATE
QUALITÄTSKONTROLLZERTIFIKAT
CERTIFICAT DE CONTRÔLE QUALITÉ
CERTIFICADO DE CONTROL DE CALIDAD
CERTIFICATO DI CONTROLLO QUALITÀ

Pipette 100-1000ul

Serial Number/Serien-Nr/N° de série/Número de serie/Numero di Soric: YE217AS0257545

Mode/Programm/Modo/Funzione: P Volume/Volumen: 200.00 ul
Mean/Mittelwert/Moyenne/Significação/Medio: 199.84 ul
Inaccuracy/Unrichtigkeit/Erreur de justesse/Inexatitud/Inaccuratezza: -0.08 %
Imprecision/Unpräzision/Erreur de répétabilité/Imprecisión/Imprecisione: 0.21 %

Date/Datum/Date/Fecha/Data: 2021-07-22 06:44:40
According to/Nach/Selon/Según/In base a: ISO8655 DIN 12650; 22°C

Tested by/Prüfer/Testée par/Comprobado por/Testado da: Snow

CE ISO 9001/13485
CERTIFIED