



03rd September, 2021

To,
Shreeji Pathology Laboratory
Geeta Malkan
A- Wing, 1st Floor
Radha Vishweshwar Apartment
Narsing Lane
Opp N.L High School
Malad, Mumbai

Dear Madam,

This is to certify that, the Instrument Model PCE-525 Blood Cell Counter bearing Serial No. 610620-000031 has been Calibrated and checked with CBC- 3K Normal control print out is given.

Kindly find attached service report of our Engineer for your reference.

Thanking you and assuring you of our attention & prompt services at all times.

Sincerely,

For AGD Biomedicals (P) Ltd.


Authorized Signatory

AGD Biomedicals (P) Ltd.
Mehta Trade Center,
Sir M. V. Road, Andheri East
Mumbai - 400 099 INDIA

P : +91-22-28231061 / 66
+91-22-28257999
E : sales@agdbio.com
W : www.agdbio.com

ISO : 9001 : 2015
ISO : 13485 : 2016

SER/F/001/00

INSTALLATION REPORT



CUSTOMER DETAILS : (* TO BE FILLED IN CAPITAL LETTERS)

Customer Name : Dr. Geeta -N. Malkan (Shreeji Pathology Lab)
 Address : Radha Vishweshwar apartment, 1st floor, Narsing lane, opp N.L. High School, Malad (w) Mumbai
 Pin code : 400064 City : Mumbai State : Maharashtra
 Contact Person : Dr. Geeta N. Malkan
 Mobile/ Tel : 91-22-2808069 Email : www.shreejipathlab.com

INSTRUMENT DETAILS :

Model : PCE - 525 Date of Installation : 29/01/2020
 Instrument Serial No. : 611219-000017 Warranty Expiry Date : 28/01/2021
 Software Version : V0 - 72007 Invoice No. : _____

Invoiced from AGD/Distributor (If through distributor, please mention the name)

Distributor Name : origin health care Place : Goregaon (E)

PHYSICAL INSPECTION & PRE-INSTALLATION CHECKS :

Instrument received in good condition : Yes No Vibration free Platform : Yes No
 CVT/Online UPS available : Yes No Dust Free : Yes No
 Is room Air-conditioned : Yes No Away from Sunlight : Yes No
 Mains AC supply : L & N : 230 VAC N & E : 0.1V VAC L & E : 230 VAC

ENGINEER REMARKS :

Instrument ~~set~~ working Satisfactory.

TO BE FILLED BY THE CUSTOMER :

We agree that the above mentioned instrument has been received in good condition and installed to my satisfaction by the Engineer. I/We have been given operational and maintenance training by the Engineer to our satisfaction.

Customer's Comments, If any _____

G.N. Malkan
 Signature of Customer :
 Seal & Date

Ajay Sharma Ajas
 Name and Signature of Engineer
 Date and Place (Malad)
29/1/2020 Mumbai

SHREEJI PATHOLOGICAL LABORATORY

A-Wing, 1st Floor,
 Vishweshwar Apartment, Narsing Lane,
 N.L. High School, Malad (W), Mumbai - 64.
 2882 7000. Fax : 81-22-2885 1189.

AGD Biomedicals (P) Ltd.
 Mehta Trade Center,
 Sir M. V. Road, Andheri East
 Mumbai - 400 099 INDIA

P : +91-22-28231061 / 66 ISO : 9001 : 2015
 +91-22-28257999 ISO : 13485 : 2016
 E : sales@agdbio.com
 W : www.agdbio.com

SER/F/007/00

AGD Biomedicals Pvt. Ltd.

Mehta Trade Centre, No.1 Shivaji Colony, Sir M. V. Road, Andheri East, Mumbai 400099. India
TEL: 91 22 2823 1061 / 2823 1066 FAX: 91 22 28224297 email: ho-service@agdbio.com
visit us at: www.agdbio.com



Sr. No 35402

SERVICE REPORT

Customer Name: Shreeji Pathology Lab		Instrument Model: PCE 525	
Address/Area: Malad (W)		Version: V0.74-005	Serial No: 610620-000031
Name Contact Person: MS. babita Lab Tech		Repair Carried Out: Workshop <input type="checkbox"/> Onsite <input checked="" type="checkbox"/>	
Instrument status	Warranty <input type="checkbox"/> Free AMC <input type="checkbox"/> AMC <input checked="" type="checkbox"/> Not in AMC <input type="checkbox"/>	Courtesy <input type="checkbox"/>	Paid Visit <input type="checkbox"/>
Type Of Call	Installation <input type="checkbox"/> Preventative Maintenance <input type="checkbox"/> Break Down <input type="checkbox"/>	APPLICATION <input type="checkbox"/>	

CALL DETAILS

CALL RECEIVED		RESPONSE		CALL COMPLETED	
DATE: 11/9/21	TIME: 11:00	DATE: 11/9/21	TIME: 11:30	DATE: 11/9/21	TIME: 12:00

Problem Reported: **Calibration Certificate**
 Observation: **CBC 3K Run Normal level Control**
 Engineer Remark: **check all path found, Run Control for calibration certificate, Run Normal level control LOT:- KK103, and result found with in range, instrument working properly.**

Hemat Reagent: CCR	Voltage Details	L-N: 220	E-N: 0.1V
Lot No:		LE: 219	CVT/UPS Y(4)N(1)

<input type="checkbox"/> FOLLOWING PARTS HAVE BEEN REPLACED	Spare Billing Details (If Any)
<input type="checkbox"/> FOLLOWING PARTS NEED TO BE REPLACED PLEASE APPROVE	Service: Customer Remarks
SPARES/AMC / CMC / VISIT CHARGE	Poor <input type="checkbox"/> Average <input type="checkbox"/> Good <input type="checkbox"/> Excellent <input type="checkbox"/>
Description	Comments:
Run PCE 525	
Normal Control	
	WE HEREBY AGREED TO PAY SUM OF Rs: -----

FOLLOW UP ACTION (If Require):
 Payment Collected: Cheque / DD / Others

Machine is working satisfactorily.

Engineer Name: Ajay Sharma	Customer Signature: Babita
Engineer Signature: Ajay	Date: 11/9/21
Date: 11/9/21	Stamp:

FOR OFFICE USE ONLY