


LAB MONTHLY SUMMARY



Lab Name DR. POTDAR LABORATORIES Lab No 1959
 Month July Year 2021
 Constituent Group Chemistry I



Date of Result Entered : 14/07/2021

Date of Report Published : 02/08/2021

SI.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE	Hexokinase	Roche C111, C311, Integra 400 plus, C501, C502, C702	288	100.02	4.77	4.78	99.3 mg/dl	-0.15	0.56
2	UREA	Urease UV / GLDH	Roche C111, C311, Integra 400 plus, C501, C502, C702	302	138.31	5.49	7.59	137.35 mg/dl	-0.13	0.87
3	CREATININE	Jaffes Kinetic - Alkaline picrate	Roche C111, C311, Integra 400 plus, C501, C502, C702	249	8.21	4.93	0.40	8.5 mg/dl	0.72	0.05
4	T.BILIRUBIN	Diazonium salt (Colorimetric) / Jendrassik	Roche C111, C311, Integra 400 plus, C501, C502, C702	289	5.44	4.80	0.26	5.5 mg/dl	0.23	0.03
5	T-PROTEIN	Biuret - Colorimetric	Roche C111, C311, Integra 400 plus, C501, C502, C702	299	4.68	4.79	0.22	4.4 g/dl	-1.25	0.03
6	ALBUMIN	BCG - colorimetric	Roche C111, C311, Integra 400 plus, C501, C502, C702	282	3.13	4.38	0.14	3.2 g/dl	0.51	0.02
7	CALCIUM	OCPC (O - Cresolphalein Compleazone)	Any Analyser (Automation / Semi Automation)	177	10.20	8.73	0.89	10.4 mg/dl	0.22	0.13
8	PHOSPHORUS	Molybdate UV / Phosphomolybdate complex	Any Analyser (Automation / Semi Automation)	319	6.39	12.37	0.79	6.3 mg/dl	-0.11	0.09
9	URIC ACID	Enzymatic / Uricase Colorimetric	Roche C111, C311, Integra 400 plus, C501, C502, C702	285	9.04	4.82	0.44	9.3 mg/dl	0.60	0.05
10	CHOLESTEROL	CHOD-PAP	Roche C111, C311, Integra 400 plus, C501, C502, C702	302	101.94	4.46	4.55	104.33 mg/dl	0.53	0.52
11	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Roche C111, C311, Integra 400 plus, C501, C502, C702	272	233.89	3.77	8.82	255.45 mg/dl	2.44	1.07
12	HDL CHO	Direct method / Enzymatic colorimetric	Roche C111, C311, Integra 400 plus, C501, C502, C702	272	21.51	5.17	1.11	24.09 mg/dl	2.32	0.13
13	SODIUM	ISE - Indirect	Roche C111, C311, Integra 400 plus, C501, C502, C702	308	145.88	4.05	5.90	158 mmol/L	2.05	0.67
14	POTASSIUM	ISE - Indirect	Roche C111, C311, Integra 400 plus, C501, C502, C702	309	3.66	5.85	0.21	4 mmol/L	1.59	0.02
15	CHLORIDE	ISE - Indirect	Roche C111, C311, Integra 400 plus, C501, C502, C702	238	100.30	4.07	4.09	101 mmol/L	0.17	0.53

16	AST	UV kinetic with PLP (P-5-P)	Roche C111, C311, Integra 400 plus, C501, C502, C702	110	108.16	13.93	15.07	118.19 U/L	0.67	2.87
17	ALT	UV kinetic with PLP (P-5-P)	Roche C111, C311, Integra 400 plus, C501, C502, C702	79	34.47	19.32	6.66	35.5 U/L	0.15	1.50
18	ALP	PNP AMP kinetic	Roche C111, C311, Integra 400 plus, C501, C502, C702	281	140.96	8.93	12.59	136 U/L	-0.39	1.50
19	AMYLASE	CNPG3	Any Analyser (Automation / Semi Automation)	393	111.19	17.02	18.93	117 U/L	0.31	1.91
20	IRON	Ferro Zine (No Protein Removal)	Roche C111, C311, Integra 400 plus, C501, C502, C702	116	200.91	4.78	9.61	298.4 ug/dl	10.14	1.78
21	MAGNESIUM	Xylidyl Blue	Roche C111, C311, Integra 400 plus, C501, C502, C702	100	3.76	5.40	0.20	2.34 mg/dl	-7.00	0.04

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Co-Ordinator

Christian Medical College, Vellore

***** End of Report *****



DR. POTDAR LABORATORIES

PATHOLOGY • MICROBIOLOGY • IMMUNOLOGY • MOLECULAR BIOLOGY

■ Pathologist

Dr. Nilkanth V. Potdar M.D.(Path.)
Reg.No. 62880

■ Microbiologist

Dr. Mrs. Shubhangi N. Potdar M.D.(Micro.)
Reg.No. 76109

Lab-1 : "Shashwat Heights" Office No.4, First Floor, 519, Shukrawar Peth, Manik Chowk,
Solapur - 413 002. Ph.: (0217) 2320999, 2320994, Cell : 9822023451

Lab-2 : "Golden Heights" 1st Floor, 1, Anand Nagar, Near I.T.I., Vijapur Road, Solapur.
Tel.: 0217 - 2346127, Cell : 9359955487 ■ Website : www.drpotdarlabs.com

Date - 18/08/2021

Pt Name - Hanmantrai

Ref. by - Dr. Self

Lab No. - 45

Age / Sex - Yrs. / M

SERUM IRON STUDIES

SERUM IRON : 163.50 ug/dl

NORMAL RANGE :


FEMALE : 28 - 170 ug/dl

MALE : 45 - 182 ug/dl

BIOCHEMISTRY

Test	Result	Unit	Normal Range
SERUM MAGNESIUM	: 1.98	mg/dl	1.20 - 2.50

Dr. Mrs. Shubhangi N. Potdar
M. D. (Microbiology)


Dr. Nilkanth V. Potdar



Reg. No. SUR/GNL/107/Dt. - 27/12/1966

SHRI MARKANDEY SOLAPUR SAHAKARI RUGNALAYA & RESEARCH CENTRE NIYAMIT

Final Plot No. 19, Pachha Peth, SOLAPUR - 413 005. (Maharashtra) Phone No. (0217) 2452000, 2452096

Email : mssrugsol@gmail.com • Website www.markandeyarugnalaya.com



No.

Patient's Name : EQAS OUTLIER (ID 1835)

DATE : 16/08/2021

Ref.by : Dr.Potdar laboratories

Blood Chemistry report

Test	Results	Unit	Reference range
Sr. Magnesium (Mg)	2.1	mg/dl	1.3 - 2.5
Sr. Iron (Fe)	172.1	ug/dl	Female : 30-170 Male : 45-185

YEARS OF HEALTHCARE SERVICES
1966 - 2016

SRK
M.D.