



CERTIFICATE OF CALIBRATION

JK Medicals
Polo View, M.A Road, Srinagar, 190001
Tel: +91-7006508935.
calibration@jkmedicals.com

Certificate Number:- JK2021-SR-0000146
Date of Issue:- 12TH October
Date of Calibration: 12TH October
Issue Number:- 1
Issued By:- JK Medicals
Page:- 1/1

Customer:- BIOCHEMISTRY DIAGNOSTIC LABORATORY, GMC, Srinagar and associated SMHS hospital
Equipment: - DEEP FREEZER thermometer reading
Manufacturer:- REMI
Serial Number:- BDI-3299
Model:-
Procedure Number:- SOP/thermal/08

Test Conditions:-
Temperature:- $-25 \pm 10^{\circ}\text{C}$
Humidity:- $50\%RH \pm 20\%RH$

Calibration Results:-

As Found Results	Measured	Displayed	Uncertainty
Display	-2.0 °C	-1.8 °C	$\pm 0.043^{\circ}\text{C}$
Display	-4.0 °C	-3.9 °C	$\pm 0.038^{\circ}\text{C}$
Display	-6.0 °C	-6.4 °C	$\pm 0.067^{\circ}\text{C}$

Authorized Signatory

The 'Measured' and 'Displayed' results are the means of 5 measurements each, taken from calibration equipment and unit under test displayed respectively.



CERTIFICATE OF CALIBRATION

JK Medicals

Polo View, M.A Road, Srinagar, 190001

Tel: +91-7006508935.

calibration@jkmedicals.com

Certificate Number:- JK2021-SR-0000145

Date of Issue:- 12TH October

Date of Calibration:- 12TH October

Issue Number:- 1

Issued By:- JK Medicals

Page:- 1/1

Customer:- BIOCHEMISTRY DIAGNOSTIC LABORATORY, GMC, Srinagar and associated SMHS hospital

Equipment:- Refrigerator Thermometer Reading

Manufacturer:- Samsung

Serial Number:- RS51K56H02A

Model:- RS51K56H02A

Job Number:-

Procedure Number:- SOP/thermal/08

Test Conditions:-

Temperature:- $-25 \pm 10^{\circ}\text{C}$

Humidity:- $50\% \text{RH} \pm 20\% \text{RH}$

Calibration Results:-

As Found Results	Measured	Displayed	Uncertainty
Display	2.0 °C	2.0 °C	$\pm 0^{\circ}\text{C}$
Display	4.0 °C	4.0 °C	$\pm 0^{\circ}\text{C}$
Display	6.0 °C	6.0 °C	$\pm 0^{\circ}\text{C}$

The 'Measured' and 'Displayed' results are the means of 5 measurements each, taken from calibration equipment and unit under test displayed respectively.

