

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

MICROCRAFT PATHOLOGY LLP (546)

Outlier And Analyte Summary Report

Result Details For Cycle No 210207 and Sample No 02


Report Date : 19/07/2021

Instrument : Sysmex XP-100

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
# Absolute Eosinophils Count.	/c.mm				
# Absolute Lymphocyte Count.	/c.mm				
@ Absolute Neutrophils Count.	/c.mm	5005.00	4,517.00	0.27	1.43
# Eosinophils Differential.	%				
@ Erythrocyte (RBC) Count.	mill/cu.mm	5.27	5.18	0.21	0.78
@ Haemoglobin.	g/dL	16.90	16.96	-0.04	0.00
# Lymphocytes Differential.	%				
@ MCH (Mean Corpuscular Hb).	pg	32.07	32.63	-0.67	-0.51
@ MCHC (Mean Corpuscular Hb Concentration).	g/dL	29.04	31.23	-0.37	-0.73
@ MCV (Mean Corpuscular Volume).	fL	110.44	104.80	0.50	0.53
@ Neutrophils Differential.	%	55.00	50.75	-1.48	-1.43
@ Platelet Count.	1000/ µL	298.00	280.14	0.89	1.17
@ Total Leucocytes (WBC) count.	/c.mm	9100.00	8,519.05	0.51	0.71

Legend @ : z score ≤ 2.0 - Acceptable
! : 2.0 < z score < 3.0 - Warning
X : z score ≥ 3 - Unacceptable
: Not Evaluated
🕒 : Delayed Result Entry
* : Not considered for evaluation.

Total Parameters	13
Not Evaluated Parameters	4
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC WB	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416,
Commercial Building - 1A,
Kohinoor Mall, Kiroli Road, Kuria (W),
Mumbai - 400070

End of Report