

CALIBRATION CERTIFICATE

Customer name And Address	A3 Hormone Lab, Salem-636016	Certificate No.	SB/2021/0071
		Date of calibration	04.10.2021
		Valid up to	03.10.2022
		Date of issue	08.10.2021
		Calibrated at	On site

<u>Details of Test Instrument</u>			
Instrument name	Refrigerator	S.no	-
Make	Samsung	Model	RT34T454258/HL/2020
Range/Size	2 ^o C to 8 ^o C	Location	Laboratory
Least Count	1 ^o C	Accuracy	As per manual
I.D No.	-	Visual Inspection	Ok

<u>Details of standards used</u>	
Name	RTD with Indicator
Certificate No.	PC-10-20/ATL/0782-01
Valid up to	23/10/2021
Traceability	yes

Environmental Condition	Temperature	24 ^o C	-	-
	Relative Humidity	60%	-	-

<u>Calibration Results</u>			
Serial No.	UUC Reading in ^o C	Standard Reading in ^o C	Uncertainty At 95% C.L. (coverage factor K=2)
1	5	5.2	2

Remarks:

- ❖ The above UUC was calibrated at On-site
- ❖ UUC is defined as Unit under Calibration.
- ❖ This result of calibration refers to only to the particular item submitted for calibration.
- ❖ The above results are valid at the time of and under the stated conditions measurement.

Note: This machine is working fine and measured values are within the range



Chief Technical Coordinator
Karthick Madheswaran B.E, MBA,