



CALIBRATION CERTIFICATE

Certificate No	: UAS/CAL/2021/08/57/07	Certificate Issue Date	: 04.08.2021
Date Of Calibration	: 02.08.2021	Next Calibration Due Date	: 01.02.2022
ULR Code	: ULR-CC300521000000614F	Calibration Method No:	UAS-CP-333
Work Order No	: 57/07	Calibrated At:	site
Date Of Receipt	: 02.08.2021		
Customer Name	: M/s.DR.POTDAR LABORATORIES		
Customer Address	: Shashwat Heights, Office No.4 First Flore, 519,Shukrawar peth, Manik Chowk, Solapur- 413002		
Temperature	: 25 ±3°C	Humidity:	40 to 60%RH

Description of Unit under Calibration (UUC)

Name	: Digital Thermometer	Range	: -50 to 300
ID No	: DPL/DTM-01	L.C.	: 0.1°C
Make	: Sigma	Accuracy	: ±1°C
Type	: PT 100	Sr. No	: 1.81E+08
Location	: On Site	Model	: LTT-ST9263

Standards & Master Equipment Used & Their Traceability

Standard Used	ID No	Traceability	Certificate No	Valid Upto
6½ Digit Multimeter	UAS/ET/301	CC-2282	SCL/20-21/01/13-003	01.02.2022
RTD Sensor (4 Wire)	UAS/TH/102	CC-2294	NI/2102/011/005	05.02.2022

Calibration Results

Cal Points	Standard Reading	UUC Reading	Error	Expanded Uncertainty
°C	°C	°C	°C	(±)°C
50	50.4	50.0	-0.4	0.4
150	150.5	150.0	-0.5	0.4
200	200.5	200.1	-0.4	0.4

Note :

The reported Expanded Uncertainty is stated as a standard Uncertainty multiplied by a coverage factor k =2 at 95% confidence level.

Remarks :-

- 1) Result are related only to the item calibrated .
- 2) This certificate refers only to the particular items submitted for calibration .
- 3) This certificate shall not be reproduced except in full without our prior permission in writing.
- 4) Standard used for calibration were traceable to National / International standard.

Approved By :

(Pramod Markale)
(Managing Director)

Authorized Signatory :

(Pawan Kokil)
(Technical Manager)

