

CERTIFICATE OF CALIBRATION

CUSTOMER DETAILS:

M/S.ALPHA LIFECARE
NO 17,BASHYAM STREET,
MANJAKUPPAM ,,CUDDALORE-607001

Calibrated at : Lab
Date of Receipt :14.5.21
Cond. On Receipt:Satisfactory
Date of issue :14.5.21

Certificate No : AL/2021-22/CAL/27
Asset No: AL/2021-22/AN/27

Date Of calibration : 14.5.21
Recom. Due Date:13.5.22

DETAILS OF UUC:

Description : REFRIGERATOR
Range : -
Least Count: -
Location : -
Cal\Working Range : 2 to 8

Identification No : RL/0001
Make : WHIRLPOOL
Serial No: -
Accuary : -
Model/Type: DC205

Environmental Details : Temperature $25 \pm 5^{\circ}\text{C}$ Relative Humidity : 40-70%RH

THERMAL CALIBRATION

CALIBRATION RESULT

SL.NO	TEST VALUE	STANDARD	UNCERTAINTY
		VALUE ($^{\circ}\text{C}$)	$\pm(^{\circ}\text{C})$
1	Refrigerator(2 to 8°C)	3.23	0.19

The above mentioned instrument has been calibrated using standards manufacturer recommended protocols, using equipments having traceability to National/International Standards. Devices for which there are no National / International Standards calibration standards are measured against in-house performance standard using accepted test procedure. Partial reproduction of the certificate is not permitted.

Calibration Done by



Approved By

