

## LAB MONTHLY SUMMARY



Lab Name SURYA DIAGNOSTIC CENTRE Lab No 14213  
 Month September Year 2021  
 Constituent Group Chemistry II



Date of Result Entered : 16/09/2021

Date of Report Published : 05/10/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Biosystems	283	401.70	9.14	36.71	393 mg/dl	-0.24	4.36
2	UREA II	Others ( Urease Berthelot / Ned Dye ) II	Any Analyser (Automation / Semi Automation )	820	86.31	15.24	13.16	81 mg/dl	-0.40	0.92
3	CREATININE II	Jaffes End point II	Any Analyser (Automation / Semi Automation )	608	3.23	14.23	0.46	3.4 mg/dl	0.37	0.04
4	T.BILIRUBIN II	Diazonium Salt ( Colorimetric ) / Jendrassik II	Any Analyser (Automation / Semi Automation )	830	2.22	17.99	0.40	2.2 mg/dl	-0.05	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Biosystems	235	5.24	8.88	0.46	4.8 g/dl	-0.95	0.06
6	ALBUMIN II	BCG - Colorimetric II	Biosystems	244	3.17	8.91	0.28	2.5 g/dl	-2.38	0.04
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Biosystems	247	8.70	10.85	0.94	8.35 mg/dl	-0.37	0.12
8	CHOLESTROL II	CHOD-PAP II	Biosystems	261	114.91	9.41	10.81	114 mg/dl	-0.08	1.34
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Biosystems	234	229.21	9.87	22.63	273 mg/dl	1.93	2.96

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between $\pm 1.0$ to $\pm 2.0$	Good.
Between $\pm 2.0$ to $\pm 3.0$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)

Contact Number: 0416-2283102

Dr. Pamela Christudoss  
 CMC EQAS Co-Ordinator

Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*