

## LAB MONTHLY SUMMARY



Lab Name **COSMOS CLINICS AND DIAGNOSTICS** Lab No **10077**  
 Month **August** Year **2021**  
 Constituent Group **HbA1c**

Date of Result Entered : **18/08/2021**Date of Report Published : **01/09/2021**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	Turbidimetric Inhibition Immunoassay	Any Analyser	219	7.52	11.72	0.88	6.7 %	-0.93	0.12
<b>SDI Range</b>		<b>Interpretation</b>								
Within -1.0 to +1.0		Excellent.								
Between ±1.0 to ±2.0		Good.								
Between ±2.0 to ±3.0		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)

Contact Number: 0416-2283102

Dr. Pamela Christudoss  
 CMC EQAS Co-Ordinator  
 Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*