

## VIEW LAB MONTHLY SUMMARY

Lab Name	COSMOS CLINICS AND DIAGNOSTICS		<a href="#">Details About Robust Analysis</a>							
Lab No	10077		<a href="#">Detail About Monthly Summary</a>							
Month	September		<a href="#">Detail about SDI</a>							
Year	2021									
Constituent Group	Chemistry I									
<a href="#">Click on the analyte to view Graphical Data</a>			<a href="#">All Analyser Result</a>		<a href="#">Print</a>					
			Date of Result Entered : 16/09/2021							
			Date of Report Published : 05/10/2021							
SI.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	<a href="#">GLUCOSE</a>	GOD-POD	Any Analyser (Automation / Semi Automation )	709	81.33	9.78	7.96	78 mg/dl	-0.42	0.60
2	<a href="#">UREA</a>	Urease UV / GLDH	Any Analyser (Automation / Semi Automation )	578	93.58	10.00	9.36	131 mg/dl	4.00	0.78
3	<a href="#">CREATININE</a>	Jaffes Kinetic - Alkaline picrate	Any Analyser (Automation / Semi Automation )	479	7.01	11.04	0.77	7.9 mg/dl	1.15	0.07
4	<a href="#">T.BILIRUBIN</a>	Diazonium salt ( Colorimetric ) / Jendrassik	Any Analyser (Automation / Semi Automation )	541	1.75	16.28	0.28	4 mg/dl	8.04	0.02
5	<a href="#">T-PROTEIN</a>	Biuret - Colorimetric	Any Analyser (Automation / Semi Automation )	666	5.37	9.63	0.52	4.1 g/dl	-2.46	0.04
6	<a href="#">CALCIUM</a>	Arsenazo III	Any Analyser (Automation / Semi Automation )	504	8.67	9.82	0.85	9.4 mg/dl	0.86	0.08
7	<a href="#">PHOSPHORUS</a>	Molybdate UV / Phosphomolybdate complex	Any Analyser (Automation / Semi Automation )	351	3.35	17.69	0.59	4 mg/dl	1.10	0.06
8	<a href="#">URIC ACID</a>	Enzymatic / Uricase Colorimetric	Any Analyser (Automation / Semi Automation )	633	7.81	11.90	0.93	9.68 mg/dl	2.01	0.07
9	<a href="#">CHOLESTEROL</a>	CHOD-PAP	Any Analyser (Automation / Semi Automation )	667	111.25	8.59	9.55	117 mg/dl	0.60	0.74
10	<a href="#">TRIGLYCERIDE</a>	GPO-PAP / Enzymatic Colorimetric / End Point	Any Analyser (Automation / Semi Automation )	615	245.32	10.30	25.26	350 mg/dl	4.14	2.04
11	<a href="#">HDL CHO</a>	Direct method / Enzymatic colorimetric	Any Analyser (Automation / Semi Automation )	443	27.77	16.06	4.46	24.1 mg/dl	-0.82	0.42
12	<a href="#">SODIUM</a>	ISE - Direct	Any Analyser (Automation / Semi Automation )	605	127.77	3.62	4.62	136 mmol/L	1.78	0.38
13	<a href="#">POTASSIUM</a>	ISE - Direct	Any Analyser (Automation / Semi Automation )	608	5.30	5.64	0.30	5.6 mmol/L	1.00	0.02
14	<a href="#">CHLORIDE</a>	ISE - Direct	Any Analyser (Automation / Semi Automation )	472	103.50	4.05	4.19	101 mmol/L	-0.60	0.39

			Automation )							
15	AST	UV kinetic without PLP (P-5-P )	Any Analyser (Automation / Semi Automation )	394	86.33	11.92	10.29	105 U/L	1.81	1.04
16	ALT	UV kinetic without PLP (P-5-P )	Any Analyser (Automation / Semi Automation )	355	257.48	16.34	42.08	296 U/L	0.92	4.47
17	ALP	PNP AMP kinetic	Any Analyser (Automation / Semi Automation )	452	183.43	18.11	33.22	240 U/L	1.70	3.13
18	AMYLASE	CNPG3	Any Analyser (Automation / Semi Automation )	390	73.03	16.72	12.21	52.3 U/L	-1.70	1.24
19	MAGNESIUM	Xylidyl Blue	Any Analyser (Automation / Semi Automation )	155	2.40	15.43	0.37	3 mg/dl	1.62	0.06

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

### Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>