



## CALIBRATION CERTIFICATE

DIAGNOSTICNAME	COSMOSDIAGNOSTIC
ASSRESS	#503, SMV ARCADE ANDRAHALLI MAIN ROAD BANGALOREKARNATAKA
ANALYSER	ICUBIO (iMagic - M7)
DEPARTMENT	BIOCHEMISTRY DEPT.
DATE	APRIL18,2021

This is to certify that the above mentioned analyzer was calibrated by the following procedure as per manufacturer \*s recommendations.

1. Analyser Rotor Init position check
2. Check Abs sensitivity100%
3. Check Init Position Accuracy
4. Check Pump Motor condition
5. Air/watercalibration

All the checks and calibration were done and system is being used under recommended conditions  
Next calibration is due on April18, 2022

Euro Diagnostic Systems Pvt. Ltd.

Gopi SL  
Technical Specialist  
Bangalore