



Multitech Technofab Enterprises

475, First Floor, F.I.E., Patparganj Industrial Area
Delhi - 110092

Phone No. : 09810943334 011-41764006, Email:
multitech.technofab@gmail.com, mtedelhi2020@gmail.com,
Web. : www.multitechtechnofab.com



CALIBRATION CERTIFICATE

Calibration Certificate No.: MTE/204/CC/1614 Certificate Issued on: 03-08-2021

Job / Identification No. : MTE/CAL/JOB/2469	Date of Calibration : 02-08-2021
Job Request Date: 29-07-2021	Recommended Date of Next Calibration: 02-08-2022

Calibration Certificate for : Centrifuge Item Location : NA

CALIBRATED FOR	HARSHIT MEDICAL RESEARCH CENTRE Tiraha, near Varun Hotel, Wazidpur, , Jaunpur, Uttar Pradesh, 222002	Reference: Order thru Telephonic
DISCIPLINE	Mechanical	GROUP Acceleration & Speed

Item's Description

Name of the Item	Make / Model	Serial / I.D. No.	Working / Inst. Range	Unit	Item Condition
Centrifuge	REMI / NA	NA / CHARK/CF/1	0 to 4000	RPM	Working
ENVIRONMENTAL CONDITION		Temperature (in °C):	25.8 ± 2 °C	Humidity (in % RH):	51 ± 10% RH

STANDARD REFERENCE DETAILS Traceable to National / International Standard

Name of the Standard	Make / Model	Certificate No.	Calibration Agency / Traceability	Date of Calibration	Valid up to
Digital Tachometer(Non contact)	Lutron / DT-2234C	2101/19/L/001	QTEX Consumer Products, Fbd	18-01-2021	18-01-2022

PRINCIPLE/METHODOLOGY OF CALIBRATION: Lab Procedure No.: MTE-2/RPM/2.50 St. Procedure No.: NA

RESULTS: On-Site Calibration
All Readings are in RPM, Specified Otherwise

U.U.C. READING(Set At)	Average	UNCERTAINTY [^] in (%)
Knob1	984.3	± 0.57
Knob2	1427.3	± 1.23
Knob3`	2217.3	± 0.48
Knob4	3218	± 0.48
Knob5`	3820	± 0.48


[^]The reported uncertainty in measurement is stated as the standard uncertainty multiplied by the coverage factor k = 2, which for a normal distribution corresponds to a coverage probability of approximately 95% confidence level.

REMARKS:

ULR : CC322521000001061F

Calibrate By
(Parveen Kumar) , parveen
Deputy Technical Manager

For and on behalf of
Multitech Technofab Enterprises


Laboratory Head



Multitech Technofab Enterprises

475, First Floor, F.I.E., Patparganj Industrial Area
Delhi - 110092

Phone No. : 09810943334 011-41764006, Email:
multitech.technofab@gmail.com, mtedelhi2020@gmail.com,
Web. : www.multitechtechnofab.com



CALIBRATION CERTIFICATE

Calibration Certificate No.: MTE/204/CC/1614

Certificate Issued on: 03-08-2021

- The recommended date for next calibration is mentioned, as asked by the customer.

Note

1. This report is not to be reported wholly or in part and cannot be used as evidence in any court of law and not to be used in any advertising media without our special permission in writing.
2. The result listed refers only to the calibrated samples and applicable parameters. Endorsement of products is neither inferred nor implied.
3. Total liability of our organization is limited to the invoiced amount.
4. Samples will be destroyed after one month from the date of issue of calibration Certificate unless otherwise specified.
5. In case any reconfirmation of contents of Calibration Certificate is required, please contact our office

***** END OF REPORT *****

ULR : CC322521000001061F

Calibrate By
(Parveen Kumar) , parveen
Deputy Technical Manager

For and on behalf of
Multitech Technofab Enterprises

Laboratory Head