



# Avi Scientific (India)

CALIBRATION LABORATORY



Certificate No. CC - 2645

## CALIBRATION CERTIFICATE

AVI/Certificate/F-12D		Calibration Certificate for Centrifuge			Page No: 01 of 01	
Date of Calibration:		Next Cal. due on:	Date of Issue:	ULR No.:	CC26452100000214F	
21.07.2021		20.07.2022	22.07.2021	Certificate No:	AVI2107070	
SRF No.	2100087	SRF Date	21.07.2021	Job No.	2107070	Job Date   21.07.2021
<b>Discipline : Mechanical calibration</b>			<b>Group : Acceleration Speed</b>			
Customer's Name & Address:			<b>Microcraft Pathology LLP</b> 106, First Floor , Navjeevan Society Dr Dadasaheb Bhandarkar Marg, Mumbai Central , Mumbai 40008, Maharashtra			
Calibration Carried out:			ONSITE			
Description of item Calibrated:						
Make			Remi			
Model			R-303			
Serial No.			ZHDN-13288			
Identification No.			-			
Range			0 to 3500 rpm			
Least Count			10 rpm			
Location			-			
Work Instruction No.			AVI/WI-16			
Environmental Condition			Temp. (°C): (25±4) Relative Humidity (% RH) (40 - 60)			
Standard(s) used for Calibration:			Digital Tachometer, Sr. No. - S1074030 Certificate No. -M-200818-16-1/A Due date-18.08.2021 Calibrated By- Godrej & Boyce Mfg.Co.ltd, CC-2634			

## OBSERVATION TABLE


Sl. No.	Range rpm	Standard Reading rpm	UUC Reading rpm	Error rpm	Expanded Uncertainty (At 95% CL, K=2), (±rpm)
1	0 to 3500 rpm	500.6	500	-0.6	4.8
2		1002	1000	-2	7.1
3		2003	2000	-3	7.1
4		3504	3500	-4	7.1

Remark : Calibration points taken as per Customer request.

Calibrated By

(Prathmesh Mestri)

Calibration Engineer



Authorised Signatory

(Janardan N. Chavan)

(Technical Manager)

Please Refer Note Backside of the Page