



Avi Scientific (India)

CALIBRATION LABORATORY



Certificate No. CC-2645

CALIBRATION CERTIFICATE

AVI/Certificate/F-12D		Calibration Certificate for Micropipette			Page No: 01 of 01	
Date of Calibration:		Next Cal. due on:	Date of Issue	ULR No :	CC26452100000213F	
21.07.2021		20.07.2022	22.07.2021	Certificate No.:	AVI2107069	
SRF No.	2100087	SRF Date	21.07.2021	Job No	2107069	Job Date 13.07.2021
Discipline : Mechanical calibration			Group : Mass & Volume			
Customer's Name & Address:			Microcraft Pathology LLP 106, First Floor , Navjeevan Society Dr Dadasaheb Bhandarkar Marg, Mumbai Central , Mumbai 40008, Maharashtra			
Calibration Carried out:			In-House			
Description of item Calibrated:						
Make			Labtech			
Model			---			
Serial No.			---			
Identification No.			MPL/MP/03			
Range			10 to 100 µl			
Location			---			
Work instruction No.			AVI/WI/12			
Environmental Condition			Temp. (°C): (25 ± 2) Relative Humidity (% RH) (40 - 60) Atmospheric Pressure (hpa) 970-1030			
Standard(s) used for Calibration:			Digital Weighing Balance, Sr No. B125154050 , Cert No.- CC/M&V/0699/20-21, CC 2287, Due date :15/09/2021 ,calibratd by IDEMI , Mumbai.			

OBSERVATION TABLE

Volume at 27°C

Sl. No.	Range	UUC Reading (µl)	Standard Reading (µl)	Error (µl)	Expanded Uncertainty (At 95% CL, K=2), (±µl)
1	10 to 100 µl	10	10.24	-0.24	0.24
2		50	50.36	-0.36	0.50
3		100	100.48	-0.48	0.50

Calibrated By


(Prathmesh Mestri)

Calibration Engineer

Authorised Signatory

(Janardan N. Chavan)

(Technical Manager)



Please Refer Note Backside of the Page