



Avi Scientific (India)

CALIBRATION LABORATORY



Certificate No. CC -2645



CALIBRATION CERTIFICATE

AVI/Certificate/F-12D	Calibration Certificate for Micropipette			Page No: 01 of 01
Date of Calibration:	Next Cal. due on:	Date of Issue	ULR No :	CC26452100000230F
05.08.2021	04.08.2022	06.08.2021	Certificate No.:	AVI2108004
SRF No. 2100095	SRF Date	05.08.2021	Job No	2108004
			Job Date	05.08.2021
Discipline : Mechanical calibration		Group : Mass & Volume		
Customer's Name & Address:		Microcraft Pathology LLP 106, First Floor , Navjeevan Society Dr Dadasaheb Bhandarkar Marg, Mumbai Central , Mumbai 40008, Maharashtra		
Calibration Carried out:		In-House		
Description of item Calibrated:				
Make		-		
Model		-		
Serial No.		MLP/MP/08		
Identification No.		-		
Range		10 to 100 µl		
Location		---		
Work Instruction No.		AVI/WI/12		
Environmental Condition		Temp. (°C): (25 ± 2) Relative Humidity (% RH) (40 - 60) Atmospheric Pressure (hpa) 970-1030		
Standard(s) used for Calibration:		Digital Weighing Balance, Sr No. B125154050 .Cert No.- CC/M&V/0699/20-21, CC 2287, Due date :15/09/2021 ,calibratd by IDEMI , Mumbai.		


OBSERVATION TABLE

Volume at 27°C

Sl. No.	Range	UUC Reading (µl)	Standard Reading (µl)	Error (µl)	Expanded Uncertainty (At 95% CL, K=2), (±µl)
1	10 to 100 µl	10	10.28	-0.28	0.24
2		50	50.32	-0.32	0.24
3		100	100.46	-0.46	0.50

Calibrated By

(Prathmesh Mestri)
Calibration Engineer



Authorised Signatory

(Janardan N. Chavan)
(Technical Manager)

Please Refer Note Backside of the Page