



# Avi Scientific (India)

CALIBRATION LABORATORY



## CALIBRATION CERTIFICATE

AVI/Certificate/F-12D		Calibration Certificate for		Micropipette	Page No: 01 of 01
Date of Calibration:		Next Cal. due on:	Date of Issue	ULR No :	CC26452100000231F
05.08.2021		04.08.2022	06.08.2021	Certificate No.:	AVI2108005
SRF No.	2100095	SRF Date	05.08.2021	Job No	2108005
				Job Date	05.08.2021
<b>Discipline : Mechanical calibration</b>			<b>Group : Mass &amp; Volume</b>		
Customer's Name & Address:			<b>Microcraft Pathology LLP</b> 106, First Floor , Navjeevan Society Dr Dadasaheb Bhandarkar Marg, Mumbai Central , Mumbai 40008, Maharashtra		
Calibration Carried out:			In-House		
Description of item Calibrated:					
Make			---		
Model			---		
Serial No.			MLP/MP/09		
Identification No.			---		
Range			100 to 1000 µl		
Location			---		
Work Instruction No.			AVI/WI/12		
Environmental Condition			Temp. (°C): (25 ± 2) Relative Humidity (% RH) (40 - 60) Atmospheric Pressure (hpa) 970-1030		
Standard(s) used for Calibration:			Digital Weighing Balance, Sr No. B125154050 , Cert No.- CC/M&V/0699/20-21, CC 2287, Due date :15/09/2021 ,calibratd by IDEMI , Mumbai.		

### OBSERVATION TABLE

#### Volume at 27°C


Sl. No.	Range	UUC Reading ( µl )	Standard Reading ( µl )	Error ( µl )	Expanded Uncertainty (At 95% CL, K=2), (±µl)
1	100 to 1000 µl	100	100.21	-0.21	0.24
2		500	500.25	-0.25	0.50
3		1000	1000.36	-0.36	0.50

Calibrated By

*Prathmesh Mestri*

(Prathmesh Mestri)

Calibration Engineer



Authorised Signatory

*Janardan N. Chavan*

(Janardan N. Chavan)

(Technical Manager)

Please Refer Note Backside of the Page