




Avi Scientific (India)

CALIBRATION LABORATORY



Certificate No. CC - 2645



CALIBRATION CERTIFICATE

AVI/Certificate/F-12D	Calibration Certificate for Micropipette			Page No: 01 of 01
Date of Calibration:	Next Cal. due on:	Date of Issue	ULR No :	CC26452100000232F
05.08.2021	04.08.2022	06.08.2021	Certificate No.:	AVI2108006
SRF No. 2100095	SRF Date	05.08.2021	Job No 2108006	Job Date 05.08.2021
Discipline : Mechanical calibration Group : Mass & Volume				
Customer's Name & Address:			Microcraft Pathology LLP 106, First Floor , Navjeevan Society Dr Dadasaheb Bhandarkar Marg, Mumbai Central , Mumbai 40008, Maharashtra	
Calibration Carried out:			In-House	
Description of item Calibrated:				
Make			---	
Model			---	
Serial No.			MLP/MP/10	
Identification No.			---	
Range			5 to 50 µl	
Location			---	
Work Instruction No.			AVI/WI/12	
Environmental Condition			Temp. (°C): (25 ± 2) Relative Humidity (% RH) (40 - 60) Atmospheric Pressure (hpa) 970-1030	
Standard(s) used for Calibration:			Digital Weighing Balance, Sr No. B125154050 , Cert No.- CC/M&V/0699/20-21, CC 2287, Due date :15/09/2021 ,calibratd by IDEMI , Mumbai.	

OBSERVATION TABLE

Volume at 27°C

Sl. No.	Range	UUC Reading	Standard Reading	Error	Expanded Uncertainty (At 95% CL, K=2), (±µl)
		(µl)	(µl)	(µl)	
1	5 to 50 µl	5	5.23	-0.23	0.24
2		25	25.28	-0.28	0.24
3		50	50.34	-0.34	0.24

Calibrated By

(Prathmesh Mestri)

Calibration Engineer



Approved By

(Janardan N. Chavan)

(Technical Manager)

Please Refer Note Backside of the Page