



Avi Scientific (India)

CALIBRATION LABORATORY



Certificate No. CC -2645


CALIBRATION CERTIFICATE


AVI/certificate/F-12D		Calibration Certificate for Digital Thermometer			Page No: 01 of 01	
Date of Calibration:		Next Cal. due on:	Date of Issue:	URL No.:	CC264521000000210F	
21.07.2021		20.07.2022	22.07.2021	Certificate No.:	AVI 2107066	
SRF No.	2100087	SRF Date	21.07.2021	Job No.	2107066	Job Date 21.07.2021
Customer's Name & Address:				Microcraft Pathology LLP 106, First Floor, Navjeevan Society Dr Dadasaheb Bhandarkar Marg, Mumbai Central, Mumbai 40008, Maharashtra		
Calibration Carried out:				In-House		
Description of item Calibrated:						
Make				-		
Model				-		
Serial No.				-		
Identification No.				MPL/DHT/02		
Range				(-40 to 150) °C		
Least Count/Res				0.1°C		
Location				-		
Accuracy						
Work Instruction No.				AVI/WI/14		
Environmental Condition				Temp. (°C):		(25±2)
				Relative Humidity (% RH)		(40 - 60)
Standard(s) used for Calibration:				Digital Indicator with Temperature Sensor (4 wire), Sr. No. O19/420, 1677, Due Date 07/06/2022. Cal Certificate No. - - QATM-21060100, Cal By QA Tech, CC-2881		


OBSERVATION TABLE

Sl. No.	Range °C	Standard °C	UUC Reading °C	Error °C	Expanded Uncertainty (At 95% CL, K=2), (±°C)
1	(-40 to 150)	-39.82	-40.0	-0.18	0.47
2		25.27	25.0	-0.27	0.47
3		50.33	50.0	-0.33	0.47
4		100.36	100.0	-0.36	0.47
5		150.45	150.0 *	-0.45	0.47

Remark : * Marked Readings are not cover under NABL Scope.

Calibrated By

 (Prathmesh Mestri)
 Calibration Engineer



Authorised Signatory

 (Janardan N. Chavan)
 (Technical Manager)

Please Refer Note Backside of the Page