




Avi Scientific (India)

CALIBRATION LABORATORY



CALIBRATION CERTIFICATE

AVI/certificate/F-12D		Calibration Certificate for Digital Thermometer			Page No: 01 of 01	
Date of Calibration:		Next Cal. due on:	Date of Issue:	URL No.:	CC264521000000207F	
21.07.2021		20.07.2022	22.07.2021	Certificate No.:	AVI 2107063	
SRF No.	2100087	SRF Date	21.07.2021	Job No.	2107063	Job Date 21.07.2021
Customer's Name & Address:				Microcraft Pathology LLP 106, First Floor , Navjeevan Society Dr Dadasaheb Bhandarkar Marg, Mumbai Central , Mumbai 40008, Maharashtra		
Calibration Carried out:				In-House		
Description of item Calibrated:						
Make				Mextech		
Model				PM10		
Serial No.				-		
Identification No.				MPL/DHT/04		
Range				(-50 to 110) °C		
Least Count/Res				0.1°C		
Location				-		
Accuracy				-		
Work Instruction No.				AVI/WI/14		
Environmental Condition				Temp. (°C): (25±2) Relative Humidity (% RH) (40 - 60)		
Standard(s) used for Calibration:				Digital Indicator with Temperature Sensor (4 wire), Sr. No. O19/420, 1677, Due Date 07/06/2022. Cal Certificate No. - - QATM-21060100, Cal By QA Tech, CC-2881		

OBSERVATION TABLE

Sl. No.	Range °C	Standard °C	UUC Reading °C	Error °C	Expanded Uncertainty (At 95% CL, K=2), (±°C)
1	(-50 to 110)	-39.86	-40.0	-0.14	0.47
2		10.23	10.0	-0.23	0.47
3		30.30	30.0	-0.30	0.47
4		50.41	50.0	-0.41	0.47
5		100.49	100.0 *	-0.49	0.47

Remark : * Marked Readings are not cover under NABL Scope.

Calibrated By (Prathmesh Mestri) Calibration Engineer		Authorised Signatory (Janardan N. Chavan) (Technical Manager)
---	--	---

Please Refer Note Backside of the Page