

# Jiangsu XinKang Medical Instrument Co., Ltd.

No. 88 Ningyan Road, Qintong Town, Jiangyan District,

Taizhou City, Jiangsu Province, China 225538

Tel:0086-523-88655555 Fax: 0086-523-88650088 Web:www.cnxinkang.com



## CALIBRATION CERTIFICATE

CUSTOMER : A3 HORMONE LAB (A Unit of Diaplus Clinic), Krishnammal Nagar,  
Salem-636106

CERTIFICATE NO. :

DESCRIPTION OF ITEM : LAB TECH Digital Centrifuge Machine Brushless

CALIBRATION DATE : 21.05.2021

MODEL NO.: 802 B1

M/C. Sr.No: 3264

RANGE :

### DETAILS OF EQUIPMENTS USED FOR CALIBRATION :-

SR. NO.	DESCRIPTION	CERTIFIED BY	CERTIFICATE NO.	VALIDITY
1	DIGITAL TACHOMETER			

### OBSERVATION :

R.P.M. CALIBRATION \* [ For Centrifuges only]

POINT	INSTRUMENT R.P.M. DISPLAYED	CALIBRATED TACHO METER READING	ERROR OBSERVED	ACCEPTED ERROR	RESULT
1	1000	1000	NIL	+/-100	OK
2	2500	2500	NIL	+/-100	OK
3	4000	4000	NIL	+/-100	OK

NEXT CALIBRATION DUE : 20.05.2022

APPROVED BY : DAIJO GEORGE

INSPECTION DEPARTMENT

\* : RPM & temperature readings are taken on standard testing panels and instruments. All readings are taken under ideal test condition.

\*\* : As required by the user.



C-poulo®

**Labtech**  
Medico (P) Ltd.

(An ISO 9001:2015 CERTIFIED COMPANY)

CIN : U52311KL2005PTC018791

Office: XI-434, Chakkiath Joseph Memorial Building  
Parappuram Jn., Karukutty P.O,  
Ernakulam Dist., Kerala - 683 576, INDIA  
Ph : +91 759 3840 603 to 610 Mob : +91 94464 25864  
Email : info.labtech@yahoo.in ; labtechmedico@gmail.com  
Website : www.labtechmedico.com  
GST IN : 32AABCL1876D1ZV

## CERTIFICATE OF CALIBRATION

CALIBRATION DATE : 21.05.2021	NEXT CALIBRATION: 20.05.2022
-------------------------------	------------------------------

NAME OF INSTRUMENT	:	DIGITAL CENTRIFUGE MACHINE BRUSHLESS
MODEL	:	12 TUBE , 80-2B1
MAKE	:	CHINA
SERIAL NUMBER	:	3264
CUSTOMER ADDRESS	:	DIAPLUS CLINIC ,KRISHNAMMAL NAGAR,SALEM-636106

CALIBRATED BY VISHNU

REVIEWED BY: DAIJO GEORGE



Labtech  
The labware people