

Parashar Micro Measurement Pvt. Ltd.



The way to Technology
(18 Years of Excellence)

B-59, SECTOR-64, NOIDA (U.P.) 201 307
Ph.: 0120-4252360-65, Email : pmmpl.india@gmail.com
Website : www.pmmpl.in, www.parashar.info

Calibration Certificate

Page 1 of 1

Certificate Number : PMM/171252		Date of Request	20.09.2021
Calibrated For	M/s. Diagnostic Point Path Labs 109 Agarwal Tower Pocket O & P Market, Dilshad Garden.	Calibrated on	21.09.2021
Description of Equipments			
Nomenclature	Incubator (Dig. Thermometer With Sensor)	Accuracy/Class Condition of UUC	Not Mentioned Physically Ok
Range	-50 to 300 ° C	Serial No.	_____
Resolution	0.1° C	Party ID Mark No.	DPPL - 19
Make & Model No.	Aceteq	Location	Lab
Master Equipment / Standard Used			
Sr.N.	Nomenclature	Make	Serial No. / I/D Mark
1	Dig. Temp Indicator With RTD Sensor	TempSens YCT	PMM/RTD/02 PMM/RTDTHM/01
Calibration Performed at		Site	Environmental Condition(s)
Calibration Procedure No.		CP/T/S/04	Temperature = 25 ± 5 ° C Humidity = 30 to 75 % RH
Reference Standard / Guideline			
IS:2806 - 1992			
Thermal Calibration			
CALIBRATION RESULTS			
S.N.	Nominal Value in UUC (in ° C)	Actual Observed Value in Std. (in ° C)	Error (in ° C)
1	30.1	30.05	0.05
2	37.3	37.18	0.12
3	40.5	40.32	0.18
4	50.4	50.18	0.22
5	60.7	60.32	0.38

UUC - Unit Under Calibration

Std. - Standard Instrument

Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor $k = 2$) = ± 0.7 ° C

Calibrated By : Sudhir Kumar
Designation : Calibration Engineer



Approved By : Moh. Iqbal
Designation : Tech. Manager / Auth. Sign.

- Conditions : 1. This certificate refers only to the particular item submitted for calibration.
2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.
3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.