



CALIBRATION CERTIFICATE

Certificate No.	MES/20-21/THQHNM/CEN/D03	Date of Issue : 28/12/2020
		Calibration Done On : 22/12/2020
		Calibration Due On : 21/12/2021
Customer Name	TALUK HEAD QUARTERS HOSPITAL, NEDUMKANDAM, IDUKKI.	

DETAILS OF DEVICE UNDER CALIBRATION

Equipment Name	Make / Model	Serial No.	Asset ID	Location
Centrifuge	Remi / R-8c PlusDX	DHLC-5674	68060963	ICTC Laboratory

DETAILS OF MASTER EQUIPMENT USED

Sl No.	Master Equipment	Make / Model	Serial No.	Reference ID	Cal. Due
1	Digital Tachometer	Equinox / EQ-801B	S916577	ACS/20-21/102671-01	24/09/2021
2	Digital Thermo Hygrometer	Equinox / EQ-308-CTH	20191602	ACS/MC/20-21/2593-01	18/06/2021

The instruments used for calibration are under valid calibration and are traceable to National Standards.


Environmental Conditions	Temperature	27°C	Relative Humidity	55 %
---------------------------------	-------------	------	-------------------	------

PERFORMANCE DATA

Set Reading	Display Reading	Measured Reading	Tolerance
Minimum	1000 rpm	1000 rpm	± 5 rpm
Medium	2500 rpm	2501 rpm	± 5 rpm
Maximum	4520 rpm	4522 rpm	± 5 rpm

Calibration Status	Pass	✓	Fail
---------------------------	------	---	------

Calibrated By


 Mahesh MC (Service Engineer)

Approved By


 Binoj Thomas (Quality Manager)