



# Perfect Calibration Centre Pvt. Ltd.

No.40, 71st Cross, 5th Block, Rajaji Nagar, Bangalore - 560 010  
 Ph : 080-2315 5522, 2315 5577, 2315 5588, 99726 97733  
 E-mail : infobl@perfectcallab.in / perfectblr@yahoo.in  
 Website : www.perfectcallab.in



**NABL LAB FOR : MECHANICAL • THERMAL • ET • AIR / FLUID FLOW • FORCE • MASS • OPTICAL**

## CERTIFICATE OF CALIBRATION

ULR-CC224721100011590F  
 ISSUE DATE:10/07/2021

REPORT NO. : PC-07-21/ATS/0366-03  
 CAL DATE : 08/07/2021  
 DUE DATE : 07/07/2022  
 PAGE NO. : 1/3

CUSTOMER : M/s. COSMOS CLINIC & DIAGNOSTICS,  
 ANDRAHALLI, MAIN ROAD, BANGALORE - 560 091.

CUST REF : VERBAL  
 DEPT. : AT SITE

### DETAILS OF DEVICE UNDER CALIBRATION

ELECTRONIC BALANCE		RANGE	Max 500 g
MAKE	----	L.C.	0.01 g
MODEL	----	LOCATION	LAB
SL. NO.	----	TYPE	DIGITAL
ID NO	CCD/DIGITALSCALE/52		

### ENVIRONMENTAL CONDITIONS:

TEMPERATURE : 25±5°C  
 HUMIDITY : 30 to 75 %RH

STANDARD USED	CERTIFICATE NUMBER	VALIDITY
MASTER WEIGHTS	WI/June/19/002	10/06/2022
MASTER WEIGHTS	3001020-V01	20/10/2021

The standards used are traceable to National / International Standards.

REFERENCE STANDARD : O.I.M.L.-R-76 RECOMMENDATION

CALIBRATION PROCEDURE : PCCPL/CAL/WB/002(S)

### CALIBRATION RESULTS

SL. NO.	NOMINAL WEIGHT VALUE	STD. WEIGHT READING	TEST BALANCE READING	OBSERVED DEVIATION
1	50 mg	0.050002 g	0.05 g	0.00 g
2	100	0.100004	0.10	0.00
3	200	0.200007	0.20	0.00
4	500	0.500004	0.50	0.00
5	1 g	1.000011	1.00	0.00
6	2	1.999978	1.99	-0.01
7	5	4.999978	4.99	-0.01
8	10	9.999974	9.98	-0.02
9	20	19.999974	19.98	-0.02
10	50	49.999974	49.96	-0.04
11	100	99.999974	99.94	-0.06
12	200	199.999974	199.94	-0.06
13	500	499.999974	499.93	-0.07



S. SRINIVASAN



# Perfect Calibration Centre Pvt. Ltd.

No.40, 71st Cross, 5th Block, Rajaji Nagar, Bangalore - 560 010.  
 Ph : 080-2315 5522, 2315 5577, 2315 5588, 99726 97733  
 E-mail : infoblr@perfectcallab.in / perfectblr@yahoo.in  
 Website : www.perfectcallab.in



**NABL LAB FOR : MECHANICAL • THERMAL • ET • AIR / FLUID FLOW • FORCE • MASS • OPTICAL**

## CERTIFICATE OF CALIBRATION


PAGE NO. :2/3

ULR-CC224721100011590F  
 ISSUE DATE:10/07/2021

REPORT NO. : PC-07-21/ATS/0366-03


### CORNER LOAD TEST ASCENDING

POSITION	STD. WEIGHT READING	TEST BALANCE READING
R0	200 g	199.94 g
R1	200	199.94
R2	200	199.91
R3	200	199.93
R4	200	199.94



### CORNER LOAD TEST DESCENDING

POSITION	STD. WEIGHT READING	TEST BALANCE READING
R0	200 g	199.94 g
R1	200	199.94
R2	200	199.91
R3	200	199.93
R4	200	199.94




### CORNER LOAD TEST AVERAGE

ASCENDING in g		DESCENDING in g	
R0 -R1	0.00	R0 -R1	0.00
R0 -R2	0.03	R0 -R2	0.03
R0 -R3	0.01	R0 -R3	0.01
R0 -R4	0.00	R0 -R4	0.00

SK



KARTHIK.P



S SRINIVASA  
 MANAGING DIR  
 AUTHORIZED SIGN





# Perfect Calibration Centre Pvt. Ltd.

No. 40, 71st Cross, 8th Block, Rajaj Nagar, Bangalore - 560 011.  
 Ph: 080-2315 5522, 2315 5577, 2315 5599, 99726 97700  
 E-mail: info@perfectcalib.in, perfectcal@perfectcalib.in  
 Website: www.perfectcalib.in



CC-2287

**NABL LAB FOR : MECHANICAL • THERMAL • ET • AIR / FLUID FLOW • FORCE • MASS • OPTICAL**

## CERTIFICATE OF CALIBRATION

ULR-CC224721100011590F  
 ISSUE DATE: 10/07/2021

REPORT NO. : PC-07-21/ATS/0366-03

PAGE NO. : 3/3

<b>REPERATABILITY OF FULL LOAD TEST</b>	<b>REPERATABILITY OF HALF LOAD TEST</b>
---	---

TEST RANGE - 500g			
TRIAL	SL.NO.	STD	DUC
TRIAL - 01	1	0	0.00
	2	500	499.93
	3	500	499.93
	4	0	0.00
TRIAL - 02	1	0	0.00
	2	500	499.93
	3	500	499.93
	4	0	0.00
TRIAL - 03	1	0	0.00
	2	500	499.93
	3	500	499.93
	4	0	0.00
TRIAL - 04	1	0	0.00
	2	500	499.93
	3	500	499.93
	4	0	0.00
TRIAL - 05	1	0	0.00
	2	500	499.93
	3	500	499.93
	4	0	0.00

TEST RANGE - 250g			
TRIAL	SL.NO.	STD	DUC
TRIAL - 01	1	0	0.00
	2	250	249.94
	3	250	249.94
	4	0	0.00
TRIAL - 02	1	0	0.00
	2	250	249.94
	3	250	249.94
	4	0	0.00
TRIAL - 03	1	0	0.00
	2	250	249.94
	3	250	249.94
	4	0	0.00
TRIAL - 04	1	0	0.00
	2	250	249.94
	3	250	249.94
	4	0	0.00
TRIAL - 05	1	0	0.00
	2	250	249.94
	3	250	249.94
	4	0	0.00

**Conclusion Remarks:**

UNCERTAINTY OF MEASUREMENTS = ± 6.56mg

THE REPORTED EXPANDED UNCERTAINTY OF MEASUREMENT IS STATED AS THE STANDARD UNCERTAINTY OF MEASUREMENT MULTIPLIED BY THE COVERAGE FACTOR K SUCH THAT THE COVERAGE PROBABILITY CORRESPONDS TO APPROXIMATELY 95% AND K = 2

**NOTE :** THIS CALIBRATION IS VALID FOR SCIENTIFIC AND INDUSTRIAL PURPOSE ONLY. HOWEVER, IF USED FOR COMMERCIAL TRADING ADDITIONAL RECOGNITION APPROVAL SHALL BE COMPILED AS REQUIRED BY DEPT. OF LEGAL METROLOGY, REGULATORY BODIES ETC.

*SK*  
 SANTHOSH KUMAR V  
 CALIBRATION ENGINEER



*[Signature]*  
 KARTHIK P  
 CALIBRATION ENGINEER



*[Signature]*  
 S SRINIVASAN  
 MANAGING DIRECTOR  
 AUTHORISED SIGNATORY