



# Avi Scientific (India)

CALIBRATION LABORATORY



Certificate No. CC - 2545



## CALIBRATION CERTIFICATE

AVI/certificate/F-12D		Calibration Certificate for Digital Thermometer			Page No: 01 of 01	
Date of Calibration:		Next Cal. due on:	Date of Issue:	URL No.:	CC264521000000209F	
21.07.2021		20.07.2022	22.07.2021	Certificate No.:	AVI 2107065	
SRF No.	2100087	SRF Date	21.07.2021	Job No.	2107065	Job Date 21.07.2021
Customer's Name & Address:			Microcraft Pathology LLP 106, First Floor , Navjeevan Society Dr Dadasaheb Bhandarkar Marg, Mumbai Central , Mumbai 40008, Maharashtra			
Calibration Carried out:			In-House			
Description of item Calibrated:						
Make			-			
Model			-			
Serial No.			-			
Identification No.			MPL/DHT/01			
Range			(-40 to 150) °C			
Least Count/Res			0.1°C			
Location			-			
Accuracy			-			
Work Instruction No.			AVI/WI/14			
Environmental Condition			Temp. (°C): (25±2) Relative Humidity (% RH) (40 - 60)			
Standard(s) used for Calibration:			Digital Indicator with Temperature Sensor (4 wire), Sr. No. O19/420, 1677, Due Date 07/06/2022. Cal Certificate No.- - QATM-21060100, Cal By QA Tech, CC-2881			

### OBSERVATION TABLE

Sl. No.	Range °C	Standard °C	UUC Reading °C	Error °C	Expanded Uncertainty (At 95% CL, K=2), (±°C)
1	(-40 to 150)	-39.85	-40.0	-0.15	0.47
2		25.24	25.0	-0.24	0.47
3		50.31	50.0	-0.31	0.47
4		100.39	100.0	-0.39	0.47
5		150.46	150.0 *	-0.46	0.47

Remark : \* Marked Readings are not cover under NABL Scope.

Calibrated By  (Prathmesh Mestri) Calibration Engineer	 601	Authorised Signatory  (Janardan N. Chavan) (Technical Manager)
---	---------	---

Please Refer Note Backside of the Page