



# Perfect Calibration Centre Pvt. Ltd.

No.40, 71st Cross, 5th Block, Rajaji Nagar, Bangalore - 560 010.  
Ph : 080-2315 5522, 2315 5577, 2315 5588, 99726 97733  
E-mail : infobl@perfectcallab.in / perfectblr@yahoo.in  
Website : www.perfectcallab.in



**NABL LAB FOR : MECHANICAL • THERMAL • ET • AIR / FLUID FLOW • FORCE • MASS • OPTICAL**

## CERTIFICATE OF CALIBRATION

ULR-CC224721100011592F  
ISSUE DATE:10/07/2021

|  |  |
|--|--|
| <b>CUSTOMER :</b> M/s. COSMOS CLINIC & DIAGNOSTICS,<br>ANDRAHALLI, MAIN ROAD, BANGALORE - 560 091. | <b>REPORT NO. :</b> PC-07-21/ATS/0366-05 |
| <b>CUST REF :</b> E-MAIL   | <b>CAL DATE :</b> 08/07/2021             |
| <b>DEPT. :</b> AT SITE   | <b>DUE DATE :</b> 07/07/2022             |
|  | <b>PAGE NO. :</b> 1/1                    |

### DETAILS OF DEVICE UNDER CALIBRATION

#### CENTRIFUGE

|                |                      |                 |              |
|----------------|----------------------|-----------------|--------------|
| <b>MAKE</b>    | REMI                 | <b>RANGE</b>    | Max 3600 rpm |
| <b>MODEL</b>   | ----                 | <b>L.C.</b>     | ----         |
| <b>SL. NO.</b> | ----                 | <b>LOCATION</b> | LAB          |
| <b>ID NO</b>   | CCD/CENTRIFUGE(B)/20 | <b>TYPE</b>     | ANALOG       |

### ENVIRONMENTAL CONDITIONS:

**TEMPERATURE :** 25±5°C  
**HUMIDITY :** 30 to 75%RH

| NO. | STANDARD USED       | ID.NO./SL.NO.  | CERT.NO             | VALIDITY   |
|-----|---------------------|----------------|---------------------|------------|
| 1   | DIGITAL TACHOMETER. | PCCPL/S/DTM/19 | HT/CC/201009-11/001 | 19/09/2021 |

The standards used are traceable to National / International Standards.

**Cal Procedure No:** PCCPL/CAL/S&A/001(S) & 002(S)

### COMPARISON METHOD


**MECHANICAL CALIBRATION (SPEED)**

### CALIBRATION RESULTS


| SL. NO | DUC NOB POSITION | STANDARD READING IN rpm | Measurement Uncertainty (±) % |
|--------|------------------|-------------------------|-------------------------------|
| 1      | 1                | 1499.5                  | 0.12                          |
| 2      | 2                | 2300.9                  | 0.12                          |
| 3      | 3                | 2819.9                  | 0.12                          |
| 4      | 4                | 3119.8                  | 0.12                          |
| 5      | 5                | 3609.3                  | 0.12                          |

### Conclusion Remarks:

- The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95% and k = 2
- Standard & DUC readings is an average of Five repeated readings.

  
SANTHOSHKUMAR.V  
CALIBRATION ENGINEER  
CALIBRATED BY



  
KARTHIK.P  
CALIBRATION ENGINEER  
CHECKED BY



  
S. MADESHWARAN  
DEPUTY QUALITY MANAGER  
AUTHORISED SIGNATORY

THIS REPORT RELATES ONLY TO THE ITEM(S) SUBMITTED. THIS REPORT SHALL NOT BE REPRODUCED EXCEPT IN FULL WITHOUT THE WRITTEN APPROVAL OF M/s. PERFECT CALIBRATION CENTRE PVT. LTD. ALL REFERENCE STANDARDS USED ARE TRACEABLE TO NATIONAL / INTERNATIONAL STANDARDS AS PER ISO / IEC 17025 - 2017 . THE RESULTS REPORTED ARE VALID AT THE TIME AND UNDER STATED CONDITIONS OF MEASUREMENTS. THE CALIBRATION INTERVAL IS TO BE DECIDED BY THE "USER" QF/7.8/01

BLR 6783



# Perfect Calibration Centre Pvt. Ltd.

No.40, 71st Cross, 5th Block, Rajaji Nagar, Bangalore - 560 010.  
Ph : 080-2315 5522, 2315 5577, 2315 5588, 99726 97733  
E-mail : infobl@perfectcallab.in / perfectblr@yahoo.in  
Website : www.perfectcallab.in



CC - 2247

**NABL LAB FOR : MECHANICAL • THERMAL • ET • AIR / FLUID FLOW • FORCE • MASS • OPTICAL**

## CERTIFICATE OF CALIBRATION

ULR-CC224721100011591F  
ISSUE DATE:10/07/2021

|  |  |
|--|--|
| <b>CUSTOMER</b> : M/s. COSMOS CLINIC & DIAGNOSTICS,<br>ANDRAHALLI, MAIN ROAD, BANGALORE - 560 091. | <b>REPORT NO.</b> : PC-07-21/ATS/0366-06 |
| <b>CUST REF</b> : E-MAIL   | <b>CAL DATE</b> : 08/07/2021             |
| <b>DEPT.</b> : AT SITE   | <b>DUE DATE</b> : 07/07/2022             |
|  | <b>PAGE NO.</b> : 1/1                    |

### DETAILS OF DEVICE UNDER CALIBRATION

#### CENTRIFUGE

|                |                      |                 |              |
|----------------|----------------------|-----------------|--------------|
| <b>MAKE</b>    | REMI                 | <b>RANGE</b>    | Max 5000 rpm |
| <b>MODEL</b>   | R-8C                 | <b>L.C.</b>     | 10 rpm       |
| <b>SL. NO.</b> | ZEBN - 08179         | <b>LOCATION</b> | LAB          |
| <b>ID NO</b>   | CCD/CENTRIFUGE(A)/19 | <b>TYPE</b>     | DIGITAL      |

### ENVIRONMENTAL CONDITIONS:

**TEMPERATURE** : 25±5°C  
**HUMIDITY** : 30 to 75%RH

| NO. | STANDARD USED       | ID.NO./SL.NO.  | CERT.NO             | VALIDITY   |
|-----|---------------------|----------------|---------------------|------------|
| 1   | DIGITAL TACHOMETER. | PCCPL/S/DTM/19 | HT/CC/201009-11/001 | 19/09/2021 |

The standards used are traceable to National / International Standards.

Cal Procedure No: PCCPL/CAL/S&A/001(S) & 002(S)

### COMPARISON METHOD

### MECHANICAL CALIBRATION (SPEED)

### CALIBRATION RESULTS

| SL. NO | DUC READING IN rpm | STANDARD READING IN rpm | OBSERVED DEVIATION IN rpm | MEASUREMENT UNCERTAINTY (±) % |
|--------|--------------------|-------------------------|---------------------------|-------------------------------|
| 1      | 1000               | 999.5                   | 0.5                       | 0.12                          |
| 2      | 1500               | 1499.8                  | 0.2                       | 0.12                          |
| 3      | 2000               | 2000.2                  | -0.2                      | 0.12                          |
| 4      | 2500               | 2500.4                  | -0.4                      | 0.12                          |
| 5      | 3000               | 3000.3                  | -0.3                      | 0.12                          |
| 6      | 4000               | 4000.3                  | -0.3                      | 0.12                          |
| 7      | 5000               | 5000.7                  | -0.7                      | 0.12                          |

### Conclusion Remarks:

- The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95% and k = 2
- Standard & DUC readings is an average of Five repeated readings.

*sk*  
SANTHOSHKUMAR.V  
CALIBRATION ENGINEER  
CALIBRATED BY



*K*  
KARTHIK.P  
CALIBRATION ENGINEER  
CHECKED BY



*Madeshwar*  
MADESHWARAN  
DEPUTY QUALITY MANAGER  
AUTHORISED SIGNATORY

THIS REPORT RELATES ONLY TO THE ITEM(S) SUBMITTED. THIS REPORT SHALL NOT BE REPRODUCED EXCEPT IN FULL WITHOUT THE WRITTEN APPROVAL OF M/s. PERFECT CALIBRATION CENTRE PVT. LTD. ALL REFERENCE STANDARDS USED ARE TRACEABLE TO NATIONAL / INTERNATIONAL STANDARDS AS PER ISO / IEC 17025 - 2017 . THE RESULTS REPORTED ARE VALID AT THE TIME AND UNDER STATED CONDITIONS OF MEASUREMENTS. THE CALIBRATION INTERVAL IS TO BE DECIDED BY THE "USER"  
QF/7.8/01

BLR 6782





# Perfect Calibration Centre Pvt. Ltd.

No.40, 71st Cross, 5th Block, Rajaji Nagar, Bangalore - 560 010.  
Ph : 080-2315 5522, 2315 5577, 2315 5588, 99726 97733  
E-mail : infobl@perfectcallab.in / perfectblr@yahoo.in  
Website : www.perfectcallab.in



NABL LAB FOR : MECHANICAL • THERMAL • ET • AIR / FLUID FLOW • FORCE • MASS • OPTICAL

## CERTIFICATE OF CALIBRATION

ULR-CC224721100011593F  
ISSUE DATE:10/07/2021

|  |  |
|--|--|
| <b>CUSTOMER</b> : M/s. COSMOS CLINIC & DIAGNOSTICS,<br>ANDRAHALLI, MAIN ROAD, BANGALORE - 560 091. | <b>REPORT NO.</b> : PC-07-21/ATS/0366-04 |
| <b>CUST REF</b> : E-MAIL   | <b>CAL DATE</b> : 08/07/2021             |
| <b>DEPT.</b> : AT SITE   | <b>DUE DATE</b> : 07/07/2022             |
|  | <b>PAGE NO.</b> : 1/1                    |

### DETAILS OF DEVICE UNDER CALIBRATION

| CENTRIFUGE |                      |          |              |
|------------|----------------------|----------|--------------|
| MAKE       | REMI                 | RANGE    | Max 3600 rpm |
| MODEL      | ----                 | L.C.     | ----         |
| SL. NO.    | ----                 | LOCATION | LAB          |
| ID NO      | CCD/CENTRIFUGE(C)/21 | TYPE     | ANALOG       |

### ENVIRONMENTAL CONDITIONS:

TEMPERATURE : 25±5°C  
HUMIDITY : 30 to 75%RH

| NO. | STANDARD USED       | ID.NO./SL.NO.  | CERT.NO             | VALIDITY   |
|-----|---------------------|----------------|---------------------|------------|
| 1   | DIGITAL TACHOMETER. | PCCPL/S/DTM/19 | HT/CC/201009-11/001 | 19/09/2021 |

The standards used are traceable to National / International Standards.

Cal Procedure No: PCCPL/CAL/S&A/001(S) & 002(S)

### COMPARISON METHOD


MECHANICAL CALIBRATION (SPEED)

### CALIBRATION RESULTS

| SL. NO | DUC NOB POSITION | STANDARD READING IN rpm | Measurement Uncertainty (±) % |
|--------|------------------|-------------------------|-------------------------------|
| 1      | 1                | 1240.8                  | 0.12                          |
| 2      | 2                | 2050.6                  | 0.12                          |
| 3      | 3                | 2690.3                  | 0.12                          |
| 4      | 4                | 3040.6                  | 0.12                          |
| 5      | 5                | 3601.5                  | 0.12                          |

### Conclusion Remarks:


- The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95% and k = 2
- Standard & DUC readings is an average of Five repeated readings.

  
SANTHOSHKUMAR.V  
CALIBRATION ENGINEER  
CALIBRATED BY



  
KARTHIK.P  
CALIBRATION ENGINEER  
CHECKED BY



  
S. MADESHWARAN  
DEPUTY QUALITY MANAGER  
AUTHORISED SIGNATORY

THIS REPORT RELATES ONLY TO THE ITEM(S) SUBMITTED. THIS REPORT SHALL NOT BE REPRODUCED EXCEPT IN FULL WITHOUT THE WRITTEN APPROVAL OF M/s. PERFECT CALIBRATION CENTRE PVT. LTD. ALL REFERENCE STANDARDS USED ARE TRACEABLE TO NATIONAL / INTERNATIONAL STANDARDS AS PER ISO / IEC 17025 - 2017. THE RESULTS REPORTED ARE VALID AT THE TIME AND UNDER STATED CONDITIONS OF MEASUREMENTS. THE CALIBRATION INTERVAL IS TO BE DECIDED BY THE "USER" QF/7.8/01

BLR 6784