

LAB MONTHLY SUMMARY



Lab Name **INDIAN MEDICAL DIAGNOSTIC CENTRE** Lab No **13670**
 Month **July** Year **2021**
 Constituent Group **Chemistry II**



Date of Result Entered : 18/07/2021

Date of Report Published : 02/08/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	986	147.82	9.98	14.75	170.4 mg/dl	1.53	0.94
2	UREA II	Others (Urease Berthelot / Ned Dye) II	Any Analyser (Automation / Semi Automation)	853	34.55	15.04	5.20	35.2 mg/dl	0.13	0.36
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	720	2.75	11.18	0.31	3.1 mg/dl	1.14	0.02
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	775	2.96	15.26	0.45	2.7 mg/dl	-0.58	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	742	5.53	9.90	0.55	5.5 g/dl	-0.05	0.04
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	711	3.31	9.70	0.32	3.6 g/dl	0.90	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	834	5.17	19.21	0.99	4.5 mg/dl	-0.68	0.07
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	984	115.76	13.00	15.04	109 mg/dl	-0.45	0.96
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	818	177.61	12.70	22.56	169 mg/dl	-0.38	1.58

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept,with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****