

LAB MONTHLY SUMMARY



Lab Name **INDIAN MEDICAL DIAGNOSTIC CENTRE** Lab No **13670**
 Month **September** Year **2021**
 Constituent Group **Chemistry II**



Date of Result Entered : 18/09/2021

Date of Report Published : 05/10/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	1006	386.50	11.89	45.96	398.4 mg/dl	0.26	2.90
2	UREA II	Others (Urease Berthelot / Ned Dye) II	Any Analyser (Automation / Semi Automation)	820	86.31	15.24	13.16	80.4 mg/dl	-0.45	0.92
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	718	3.30	11.92	0.39	3.5 mg/dl	0.51	0.03
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	830	2.22	17.99	0.40	2.3 mg/dl	0.20	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	751	5.37	9.93	0.53	5.2 g/dl	-0.32	0.04
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	733	3.19	9.97	0.32	3 g/dl	-0.60	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	799	8.43	13.92	1.17	10.7 mg/dl	1.93	0.08
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	973	115.65	11.78	13.62	107.5 mg/dl	-0.60	0.87
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	806	223.24	11.93	26.62	235.4 mg/dl	0.46	1.88

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****