

LAB MONTHLY SUMMARY



Lab Name DR DHAM S DIAGNOSTIC CENTRE Lab No 11089
 Month March Year 2021
 Constituent Group Chemistry II



Date of Result Entered : 16/03/2021

Date of Report Published : 01/04/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Mindray	146	162.75	8.45	13.76	185 mg/dl	1.62	2.28
2	UREA II	Urease UV / GLDH II	Mindray	114	42.71	10.49	4.48	56 mg/dl	2.97	0.84
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Mindray	92	1.61	11.65	0.19	1.7 mg/dl	0.48	0.04
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Mindray	106	3.66	12.39	0.45	3.9 mg/dl	0.53	0.09
5	T-PROTEIN II	Biuret - Colorimetric II	Mindray	127	5.66	6.81	0.38	6.7 g/dl	2.70	0.07
6	ALBUMIN II	BCG - Colorimetric II	Mindray	129	3.49	10.22	0.36	4.3 g/dl	2.27	0.06
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Mindray	131	7.30	12.71	0.93	6.8 mg/dl	-0.54	0.16
8	CHOLESTROL II	CHOD-PAP II	Mindray	136	121.09	8.58	10.39	154 mg/dl	3.17	1.78
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Mindray	129	121.97	12.35	15.06	123 mg/dl	0.07	2.65

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****

SUMMARY



Lab Name DR DHAM S DIAGNOSTIC CENTRE Lab No 11089
 Month April Year 2021
 Constituent Group Chemistry II



Date of Result Entered : 17/04/2021

Date of Report Published : 01/05/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Mindray	125	321.37	9.20	29.58	397 mg/dl	2.56	5.29
2	UREA II	Urease UV / GLDH II	Mindray	105	26.68	11.98	3.20	27 mg/dl	0.10	0.62
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	Mindray	87	6.59	11.03	0.73	7 mg/dl	0.56	0.16
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Mindray	122	3.47	16.03	0.56	3.4 mg/dl	-0.13	0.10
5	T-PROTEIN II	Biuret - Colorimetric II	Mindray	99	5.40	7.56	0.41	4 g/dl	-3.41	0.08
6	ALBUMIN II	BCG - Colorimetric II	Mindray	126	3.34	9.44	0.32	2.4 g/dl	-2.98	0.06
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Mindray	98	4.30	16.80	0.72	4.2 mg/dl	-0.14	0.15
8	CHOLESTROL II	CHOD-PAP II	Mindray	142	115.60	9.91	11.46	152.5 mg/dl	3.22	1.92
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Mindray	118	233.19	9.59	22.36	304 mg/dl	3.17	4.12

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****

SUMMARY



Lab Name DR DHAM S DIAGNOSTIC CENTRE Lab No 11089
 Month May Year 2021
 Constituent Group Chemistry II



Date of Result Entered : 18/05/2021

Date of Report Published : 05/06/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Mindray	129	133.59	8.67	11.58	130 mg/dl	-0.31	2.04
2	UREA II	Urease UV / GLDH II	Mindray	107	64.54	11.57	7.47	73 mg/dl	1.13	1.44
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Mindray	82	1.81	9.34	0.17	2 mg/dl	1.12	0.04
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Mindray	110	2.52	17.62	0.44	2.34 mg/dl	-0.40	0.08
5	T-PROTEIN II	Biuret - Colorimetric II	Mindray	120	5.42	11.27	0.61	5.84 g/dl	0.69	0.11
6	ALBUMIN II	BCG - Colorimetric II	Mindray	119	3.36	11.48	0.39	2.87 g/dl	-1.27	0.07
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Mindray	116	6.89	14.60	1.01	6.8 mg/dl	-0.09	0.19
8	CHOLESTROL II	CHOD-PAP II	Mindray	136	112.55	11.60	13.06	137 mg/dl	1.87	2.24
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Mindray	123	223.20	9.92	22.13	257 mg/dl	1.53	3.99

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****

SUMMARY



Lab Name DR DHAM S DIAGNOSTIC CENTRE Lab No 11089
 Month June Year 2021
 Constituent Group Chemistry II



Date of Result Entered : 09/06/2021

Date of Report Published : 03/07/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Mindray	146	197.88	10.47	20.72	217 mg/dl	0.92	3.43
2	UREA II	Urease UV / GLDH II	Mindray	114	77.16	9.25	7.14	114.5 mg/dl	5.23	1.34
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Mindray	83	4.35	10.51	0.46	5.42 mg/dl	2.34	0.10
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Mindray	129	3.86	15.41	0.59	4.31 mg/dl	0.76	0.10
5	T-PROTEIN II	Biuret - Colorimetric II	Mindray	125	5.31	8.52	0.45	6.44 g/dl	2.50	0.08
6	ALBUMIN II	BCG - Colorimetric II	Mindray	130	3.29	10.15	0.33	3.37 g/dl	0.24	0.06
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Mindray	115	4.59	22.58	1.04	4.6 mg/dl	0.01	0.19
8	CHOLESTROL II	CHOD-PAP II	Mindray	134	107.35	10.66	11.44	127 mg/dl	1.72	1.98
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Mindray	128	139.58	10.12	14.12	150.3 mg/dl	0.76	2.50

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****

SUMMARY



Lab Name DR DHAM S DIAGNOSTIC CENTRE Lab No 11089
 Month July Year 2021
 Constituent Group Chemistry II



Date of Result Entered : 16/07/2021

Date of Report Published : 02/08/2021

Sl.No	Analyte	Method /Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Mindray	142	151.78	7.48	11.36	163 mg/dl	0.99	1.91
2	UREA II	Urease UV / GLDH II	Mindray	119	31.60	13.28	4.20	38 mg/dl	1.53	0.77
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Mindray	94	2.80	10.45	0.29	3.25 mg/dl	1.54	0.06
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Mindray	125	2.92	12.35	0.36	2.5 mg/dl	-1.16	0.06
5	T-PROTEIN II	Biuret - Colorimetric II	Mindray	125	5.39	8.15	0.44	5.9 g/dl	1.16	0.08
6	ALBUMIN II	BCG - Colorimetric II	Mindray	135	3.35	9.01	0.30	3.25 g/dl	-0.33	0.05
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Mindray	130	4.85	19.12	0.93	5.3 mg/dl	0.48	0.16
8	CHOLESTROL II	CHOD-PAP II	Mindray	144	113.36	10.13	11.48	129.5 mg/dl	1.41	1.91
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Mindray	136	175.27	10.64	18.64	188.7 mg/dl	0.72	3.20

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss

CMC EQAS Co-Ordinator

Christian Medical College, Vellore

***** End of Report *****

SUMMARY

Lab Name DR DHAM S DIAGNOSTIC CENTRE Lab No 11089
 Month August Year 2021
 Constituent Group Chemistry II



Date of Result Entered : 18/08/2021

Date of Report Published : 01/09/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Mindray	131	257.89	9.56	24.66	294 mg/dl	1.46	4.31
2	UREA II	Urease UV / GLDH II	Mindray	112	44.17	9.62	4.25	47 mg/dl	0.67	0.80
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Mindray	88	1.87	14.43	0.27	2.1 mg/dl	0.85	0.06
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Mindray	118	4.15	10.91	0.45	4.2 mg/dl	0.11	0.08
5	T-PROTEIN II	Biuret - Colorimetric II	Mindray	129	5.35	8.73	0.47	5.4 g/dl	0.11	0.08
6	ALBUMIN II	BCG - Colorimetric II	Mindray	127	3.33	9.26	0.31	3.3 g/dl	-0.10	0.05
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Mindray	121	3.42	24.47	0.84	3.7 mg/dl	0.33	0.15
8	CHOLESTROL II	CHOD-PAP II	Mindray	133	110.62	9.25	10.24	128 mg/dl	1.70	1.78
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Mindray	135	133.53	15.27	20.39	140 mg/dl	0.32	3.51

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****

VIEW LAB MONTHLY SUMMARY

Lab Name	DR DHAM S DIAGNOSTIC CENTRE	Details About Robust Analysis	
Lab No	11089	Detail About Monthly Summary	
Month	September	Detail about SDI	
Year	2021		
Constituent Group	Chemistry II		
Click on the analyte to view Graphical Data		All Analyser Result	Print
		Date of Result Entered : 20/09/2021	
		Date of Report Published : 05/10/2021	

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	GLUCOSE II	GOD-POD II	Mindray	143	403.99	11.17	45.12	490 mg/dl	1.91	7.55
2	UREA II	Urease UV / GLDH II	Mindray	112	91.16	8.16	7.44	94 mg/dl	0.38	1.41
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	Mindray	87	3.35	10.93	0.37	2.4 mg/dl	-2.60	0.08
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Mindray	125	2.12	13.90	0.30	2.3 mg/dl	0.61	0.05
5	T-PROTEIN II	Biuret - Colorimetric II	Mindray	121	5.13	7.50	0.38	5.5 g/dl	0.96	0.07
6	ALBUMIN II	BCG - Colorimetric II	Mindray	124	3.21	8.83	0.28	2.2 g/dl	-3.56	0.05
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Mindray	122	8.39	12.38	1.04	9.9 mg/dl	1.45	0.19
8	CHOLESTROL II	CHOD-PAP II	Mindray	138	113.82	10.26	11.68	120 mg/dl	0.53	1.99
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Mindray	126	230.68	9.14	21.09	220 mg/dl	-0.51	3.76

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

SUMMARY



Lab Name DR DHAM S DIAGNOSTIC CENTRE Lab No 11089
 Month October Year 2021
 Constituent Group Chemistry II



Date of Result Entered : 21/10/2021

Date of Report Published : 01/11/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Mindray	154	109.15	10.14	11.07	116.6 mg/dl	0.67	1.78
2	UREA II	Urease UV / GLDH II	Mindray	124	26.65	11.74	3.13	13.3 mg/dl	-4.27	0.56
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Mindray	95	1.33	11.04	0.15	1.33 mg/dl	0.00	0.03
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Mindray	131	7.14	11.17	0.80	6 mg/dl	-1.43	0.14
5	T-PROTEIN II	Biuret - Colorimetric II	Mindray	138	5.16	8.55	0.44	4.7 g/dl	-1.04	0.08
6	ALBUMIN II	BCG - Colorimetric II	Mindray	137	3.16	8.45	0.27	2.2 g/dl	-3.60	0.05
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Mindray	130	5.84	13.84	0.81	6.3 mg/dl	0.57	0.14
8	CHOLESTROL II	CHOD-PAP II	Mindray	143	104.86	9.87	10.35	132 mg/dl	2.62	1.73
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Mindray	116	99.58	12.21	12.15	84.6 mg/dl	-1.23	2.26

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss

CMC EQAS Co-Ordinator

Christian Medical College, Vellore

***** End of Report *****