# Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No: 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone: 0452 - 2561691 Mobile: +91 94434 74610



GSTIN. 33AAHCA8053H1ZR

DL No. MDU/4205/20B,MDU/3971/21B

CIN: U33110TN2009PTC072140

## **CALIBRATION CERTIFICATE**

-			
			-
			200

Customer	M/S, Indian Medical Diagnostic Centre,	Certificate No:	AMS/LAB/2889586
	593/A1, KJS Complex, Near Canara Bank Main	Issuing Date:	19/07/2021
			1 of 1

Details of Device under Calibration (DUC):

DUC:	Microscope	Id. No:	IMDC/LAB/MS/001	
Make:	Almicro	Department:	Laboratory	
Model:	BM – 6bi Ultra	Calibrated On:	17/07/2021	
Sl. No:		Calibration		
Type: Electric Binocular		Due Suggested:	16/07/2022	
Range &		DUC Condition		
Least Count:		on Receipt:	OK	

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electrical Safety Analyser	RigelSafeTest60	PSG/CIRD/BMCC/5507/2021	18/01/2022
2	Digital Lux Meter	HTC/LX - 103	PC/01/21/ATL/0194-02	22/01/2022
3	Thermo Hygrometer	Beetech/088	PC/01/21/ATL/0194-01	22/01/2022

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

#### **ESA Test Results:**

SL.No	Parameter	ESA Reading	Unit	Normal Range	Condition
1	Current	68.0	mA	100	Normal
2	Frequency	50.1	Hz	49 -50	Normal
3	Mains Voltage Line - Neutral	235.3	Volts	220 - 240	Normal
4	Mains Voltage Line - Earth	233.7	Volts	220 - 240	Normal
5	Mains Voltage Neutral - Earth	4.5	Volts	< 7	Normal

<sup>\*</sup>Electrical Safety: Please check the attached log sheet

SL.No.	Knob Position	DUC Reading (LUX)	Distance (Cm)
1	Low	2600	5.0
2	Medium	9120	5.0
3	High	18300	5.0

## Note:

- 1. Calibration Points were selected as per customer request.
- 2. Each reading is taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 Badings.
- 3. Results Reported are valid at the time of and under the stated conditions of measurements.

## Calibrated By:

(Calibration Engineer)

b. 894



Approved by: (Technical Manager)

6.894