

VIEW LAB MONTHLY SUMMARY

Lab Name	COSMOS CLINICS AND DIAGNOSTICS		Details About Robust Analysis							
Lab No	10077		Detail About Monthly Summary							
Month	July		Detail about SDI							
Year	2021									
Constituent Group	Chemistry I									
Click on the analyte to view Graphical Data			All Analyser Result		Print					
			Date of Result Entered : 12/07/2021							
			Date of Report Published : 02/08/2021							
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE	GOD-POD	Any Analyser (Automation / Semi Automation)	704	100.34	9.43	9.46	121 mg/dl	2.18	0.71
2	UREA	Urease UV / GLDH	Any Analyser (Automation / Semi Automation)	573	137.38	10.95	15.04	4.2 mg/dl	-8.86	1.26
3	CREATININE	Jaffes Kinetic - Alkaline picrate	Any Analyser (Automation / Semi Automation)	482	7.60	12.25	0.93	8.7 mg/dl	1.18	0.08
4	T.BILIRUBIN	Diazonium salt (Colorimetric) / Jendrassik	Any Analyser (Automation / Semi Automation)	568	5.80	13.03	0.76	6.8 mg/dl	1.32	0.06
5	T-PROTEIN	Biuret - Colorimetric	Any Analyser (Automation / Semi Automation)	691	5.29	11.14	0.59	5.5 g/dl	0.36	0.04
6	CALCIUM	Arsenazo III	Any Analyser (Automation / Semi Automation)	476	10.29	7.90	0.81	13.4 mg/dl	3.84	0.07
7	PHOSPHORUS	Molybdate UV / Phosphomolybdate complex	Any Analyser (Automation / Semi Automation)	319	6.39	12.37	0.79	6.7 mg/dl	0.39	0.09
8	URIC ACID	Enzymatic / Uricase Colorimetric	Any Analyser (Automation / Semi Automation)	672	8.84	13.24	1.17	14.9 mg/dl	5.18	0.09
9	CHOLESTEROL	CHOD-PAP	Any Analyser (Automation / Semi Automation)	678	107.24	11.24	12.06	131 mg/dl	1.97	0.93
10	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Any Analyser (Automation / Semi Automation)	635	220.96	11.08	24.48	255 mg/dl	1.39	1.94
11	HDL CHO	Direct method / Enzymatic colorimetric	Any Analyser (Automation / Semi Automation)	456	27.11	18.80	5.10	53.6 mg/dl	5.19	0.48
12	SODIUM	ISE - Direct	Any Analyser (Automation / Semi Automation)	579	142.97	3.40	4.87	148 mmol/L	1.03	0.40
13	POTASSIUM	ISE - Direct	Any Analyser (Automation / Semi Automation)	580	3.56	5.79	0.21	3.7 mmol/L	0.68	0.02
14	CHLORIDE	ISE - Direct	Any Analyser (Automation / Semi Automation)	455	102.25	4.39	4.49	97 mmol/L	-1.17	0.42

			Automation)							
15	AST	UV kinetic without PLP (P-5-P)	Any Analyser (Automation / Semi Automation)	379	110.19	15.80	17.40	90 U/L	-1.16	1.79
16	ALT	UV kinetic without PLP (P-5-P)	Any Analyser (Automation / Semi Automation)	276	36.42	19.52	7.11	41 U/L	0.64	0.86
17	ALP	PNP AMP kinetic	Any Analyser (Automation / Semi Automation)	427	146.72	16.69	24.49	190 U/L	1.77	2.37
18	AMYLASE	CNPG3	Any Analyser (Automation / Semi Automation)	393	111.19	17.02	18.93	118.7 U/L	0.40	1.91
19	IRON	Ferro Zine (No Protein Removal)	Any Analyser (Automation / Semi Automation)	49	182.56	9.98	18.22	705.2 ug/dl	28.68	5.21
20	MAGNESIUM	Xylidyl Blue	Any Analyser (Automation / Semi Automation)	121	3.31	16.81	0.56	4.9 mg/dl	2.84	0.10

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out



COSMOS CLINICS AND DIAGNOSTICS
#503, SMV Arcade, Andrahalli Main Road, Syndicate
Bank Layout, Bengaluru- 560091

Title	CCD.LAB.FR.002		
Document Type	Format	Document Number	
Prepared By	Mamatha	Checked by/Approved	Tanveer /Dr.Nirupama.S
Revision No.	0	Issue Date	

Laboratory: ONCOSTEM DIAGNOSTICS PVT. LTD.

Date:

5 Aug 2021

Area of NC occurrence:

PT Outlier.

NC No. PT-N:001

Major/minor

Non-Conformity (NC) Raised:

July EQAI sample seen, result obtained a
2nd Aug show here, calcium, use Add end HDL (outlier)

Immediate corrective action :

- calibration carried out , QC seen.

Root cause analysis:

Reagent stability due to heating or prolonged exposure
Corrective Action - fresh reagent to be taken and sample
to be seen in batch.

CORRECTIVE ACTION PROPOSED:

Reagent maintenance and lamp replacement
carried out.

EVIDENCE OF IMPLEMENTATION FOR CONTINUAL IMPROVEMENT:

- QC seen on subsequent days are within range.

VERIFIED BY:


Dr Nirupama.S

PHOENIX DIAGNOSTICS Calibration report

Item: CALC

Calibration method: Two-point
Curve

Calibration time: 2021-08-17 17:03:16




Test Date
2021-08-17 16:52:24
2021-08-17 16:53:00

Response
35.6878
5860.1311

Conc.
0
10.7

Put
↔

Reaction Curve




2
Peak

PHOENIX DIAGNOSTICS Calibration report

Calibration method: Two-point
Curve

Calibration time: 2021-08-17 17:09:01




2021-11-08 09:51:2

Test Date	Response	Conc.
2021-08-17 16:58:30	5.2929	0
2021-08-17 16:59:06	1970.7051	5.2

FormCalibReq

Reaction Curve



Red
↗


PHOENIX DIAGNOSTICSCalibration report

Item: UREA

Calibration method: Two-point

Calibration time: 2021-08-17 17:04:44

Curve



2021-11-08 09:50:2


Test Date	Response	Conc.
2021-08-17 16:53:36	102.829	0
2021-08-17 16:54:14	1426.836	103

Dec

FormCalibReq

Reaction Curve

URECA



Sheel
↙

QC information

Test time	QC Liquid	Item	Ave.	SD	Result	SD Range	In Control
2021-08-17 17:42	BIO-RAD L2	DBIL	0.9	0.09	1.377	0.72-1.08	N
2021-08-17 17:41	BIO-RAD L2	TBIL	4.5	0.45	14.81	3.6-5.4	N
2021-08-17 17:41	BIO-RAD L2	<u>URIC</u>	10.3	1	<u>8.28</u>	<u>8.3-12.3</u>	N
2021-08-17 17:40	BIO-RAD L2	HDL	27	3	39.2	21-33	N
2021-08-17 17:40	BIO-RAD L2	GLU	301	30	291	241-361	Y
2021-08-17 17:39	BIO-RAD L2	CREA	5.35	0.75	5.52	3.85-6.85	Y
2021-08-17 17:38	BIO-RAD L2	GGT	158	20	152.9	118-198	Y
2021-08-17 17:38	BIO-RAD L2	UREA	100	10	101.41	80-120	Y
2021-08-17 17:37	BIO-RAD L2	<u>CALC</u>	11.8	0.8	<u>12</u>	<u>10.2-13.4</u>	Y
2021-08-17 17:37	BIO-RAD L2	TPR	4.5	0.4	4.67	3.7-5.3	Y
2021-08-17 17:36	BIO-RAD L2	ALP	433	40	333	353-513	N
2021-08-17 17:35	BIO-RAD L2	SGOT	157	25	194	107-207	Y
2021-08-17 17:35	BIO-RAD L2	TGL	115	11.4	112	92.2-137.8	Y
2021-08-17 17:34	BIO-RAD L2	CHOL	106	19	95	68-144	Y
2021-08-17 17:34	BIO-RAD L2	SGPT	107	23	94	61-153	Y
2021-08-17 08:54	BIO-RAD L2	D BILU	0	0	0.187	0-0	N
2021-08-17 08:53	BIO-RAD L2	HDL	27	3	37.4	21-33	N
2021-08-17 08:53	BIO-RAD L2	CALC	11.8	0.8	9.6	10.2-13.4	N
2021-08-17 08:52	BIO-RAD L2	URIC	10.3	1	11.39	8.3-12.3	Y
2021-08-17 08:51	BIO-RAD L2	GGT	158	20	164.6	118-198	Y
2021-08-17 08:51	BIO-RAD L2	ALP	433	40	335	353-513	N
2021-08-17 08:50	BIO-RAD L2	SGPT	107	23	109	61-153	Y
2021-08-17 08:50	BIO-RAD L2	SGOT	157	25	181	107-207	Y
2021-08-17 08:50	BIO-RAD L2	TPR	4.5	0.4	1.44	3.7-5.3	N
2021-08-17 08:49	BIO-RAD L2	TBIL	4.5	0.45	0	3.6-5.4	N
2021-08-17 08:48	BIO-RAD L2	TGL	115	11.4	80	92.2-137.8	N
2021-08-17 08:48	BIO-RAD L2	CHOL	106	19	95	68-144	Y
2021-08-17 08:47	BIO-RAD L2	CREA	5.35	0.75	5.36	3.85-6.85	Y
2021-08-17 08:47	BIO-RAD L2	<u>UREA</u>	100	10	<u>115.56</u>	<u>80-120</u>	Y
2021-08-17 08:46	BIO-RAD L2	GLU	301	30	272	241-361	Y
2021-08-17 08:45	BIO-RAD L2						

Technical Service Report

CUSTOMER NAME: Cosmos Diagnostics
ADDRESS: #503 SMV Arcade Andrahalli Main road, Syndicate Bank layout Ban-91.
TEL. NO:

DATE: 20/06/2021
 India service net AMC
 CHARGEABLE CALL P.M. VISIT
 COURTESY CALL APPLICATION

VOLTAGE: Line/Neutral: 230...V, Line /Earth: 230...V, Earth/Neutral: 0...V (should not >005)
POWER SOURCE: DIRECT / UPS / CVT

INSTRUMENT: Fully automatic analyser
MODEL: Emagic **SERIAL NO:**

WORK CARRIED OUT AT
 SITE SERVICE CENTRE

PROBLEM REPORTED: Result Varied - (QC outliers)

PARAMETERS (Synergy Bio 1904C)
T: ok **A:** ok **Pump Speed :** ok
Pump: ok **Front Board:** ok **Gain Cal:** ok **Pump Cal:** ok

PARAMETERS (AutoLab Star Plus/ Versa)
Pump Calibration:
 500 ul (800-1200 ul)

FILTER STATUS:

Filters in nm	340	405					
Gain							
A/D Value (for Star Plus & Versa)	48572	46420					
Abs of D/Water							

Instrument Status:
 Outer Body: ok Display: ok Key Pad: ok Pump Motor: ok Printer: ok
 Aspiration Button: ok Tubing Set: chopped Flow Cell: ok Internal Connectors: ok Software: ok
 Remarks on the defective part / parts if any: Cuvette Holder/segment changed completely.

SERVICE DETAILS

FOLLOWING PARTS HAVE BEEN REPLACED (if changed)

PART NO.	DESCRIPTION	QTY	PRICE
1.	Cuvettes Holder	1	-

TO BE FILLED IN BY CUSTOMER

PREVENTIVE MAINTENANCE/ SERVICE WORK CARRIED OUT SATISFACTORILY

USED FROM COMPANY'S STOCK CUSTOMER'S STOCK

FAULT RECIFIED & INSTRUMENT IS WORKING SATISFACTORILY

SERVICE CHARGE / SPARE PARTS' CHARGE DETAILS:
 Received with thanks the sum of Rs. (Rupees.....) dated vide cash/Cheque No..... as the service charge/ spare parts' charge for servicing the above mentioned instrument.

COSMOS CLINICS & DIAGNOSTICS
 (A Unit of Phoenix Diagnostics)
 No. 503, SMV Arcade, Andrahalli Main Road, Syndicate Bank Layout, Bengaluru - 560091

NOTE:
 Parts replaced are chargeable once warranty Period is over.
 Consumables like Printer Head, Lamp, tubings, paper rolls, glass parts etc. are not covered by warranty and hence are chargeable.
 If the customer is not under AMC, customer needs to pay service charge to the engineer for any sort of service.

20.06.2021 *[Signature]*
DATE: **CUSTOMER SIGNATURE WITH**

ENGINEER' REMARKS Cuvette Rack/ Holder Replaced and working found ok.

OFFICE USE
RECEIVED ON:
CHECKED BY:

ENGINEERS' SIGNATURE & NAME: *[Signature]*



Technical Service Report

(Biom Auto Analyser)

CUSTOMER NAME: Chaitanya Lippa

Address: _____

Contact No.: _____

Voltage: Line/Neutral/Earth: 230V Line/Earth: 230V Earth/Neutral: 0 V (Should not - 000)

Power Source: Direct/UPS/EVT

INSTRUMENT: Fully Automated Analyser

MODEL: 1000 **SERIAL NO.:** _____

PROBLEM REPORTED: Result Variation

Date: 22/07/2021

Warranty I AMC

Chargeable Call RPM Visit

Courtesy Call Application

WORK CARRIED OUT AT SITE SERVICE CENTRE

PARAMETERS (Synergy Bio 1004C)

T: ok A: ok Pump Speed: ok

Pump: ok Front Board: ok Gain Cal: ok Pump Cal: ok

PARAMETERS (Autolab / Versa)

Pump Calibration: 500 ul

FILTER STATUS:

Filters in mm	340	405					
Gain							
A/D Value							
Abs of D/Water							

Instrument Status:

Outer Body: ok Display / I Screen: ok Key Pad: ok Pump Motor: ok Printer: ok

Aspiration Button: ok Tubing Set: ok Flow Cell: ok Internal Connectors: ok Software: ok

Remarks on the defective part / parts if any: No

SERVICE DETAILS

FOLLOWING PARTS HAVE BEEN REPLACED (if changed)

PART NO	DESCRIPTION	QTY	PRICE

TO BE FILLED IN BY CUSTOMER

PREVENTIVE MAINTENANCE / SERVICE WORK CARRIED OUT SATISFACTORILY

FAULT RECTIFIED & INSTRUMENT IS WORKING SATISFACTORILY

USED FROM COMPANY'S STOCK CUSTOMER'S STOCK

SERVICE CHARGE / SPARE PART'S CHARGE DETAILS:

Received with thanks the sum of Rs. (Rupees,) vide cash/Cheque No. dated as the service charge/spare part's charge for servicing the above mentioned instrument.

NOTE:

- Parts replaced are chargeable once warranty Period is over.
- Consumables like printer Head, Lamp, Tubings, paper rolls, glass parts, Flowcell, I-Screen etc are not covered under warranty and hence are chargeable.
- If the customer is not under AMC, customer needs to pay service charge to the engineer for any sort of service.

ENGINEER'S REMARKS: checked the machine properly and found that lamp was damaged. So adjust the lamp now the machine is working fine.

ENGINEER'S SIGNATURE & NAME: [Signature]

Seal & Signature of the Customer

[Signature]

Name: _____

(with seal)

DATE: _____

QC information

Test time	QC Liquid	Item	Ave	SD	Result	SD Range	In Control
2021-09-11 08:55	BIO-RAD L2	HEX	27	3	31.4	21-33	Y
2021-09-11 08:54	BIO-RAD L2	CALC	11.8	0.8	12.2	10.2-13.4	Y
2021-09-11 08:54	BIO-RAD L2	URIC	10.3	1	8.28	8.3-12.3	N
2021-09-11 08:53	BIO-RAD L2	GGT	158	20	165.4	118-198	Y
2021-09-11 08:53	BIO-RAD L2	ALP	433	40	393	353-513	Y
2021-09-11 08:52	BIO-RAD L2	SGPT	107	23	90	61-153	Y
2021-09-11 08:51	BIO-RAD L2	SGOT	157	25	165	107-207	Y
2021-09-11 08:51	BIO-RAD L2	TFR	4.5	0.4	4.12	3.7-5.3	Y
2021-09-11 08:50	BIO-RAD L2	DBIL	0.9	0.09	1.571	0.72-1.08	N
2021-09-11 08:50	BIO-RAD L2	TBIL	4.5	0.45	4.984	3.6-5.4	Y
2021-09-11 08:49	BIO-RAD L2	TGL	115	11.4	93	92.2-137.8	Y
2021-09-11 08:48	BIO-RAD L2	CHOL	106	19	97	68-144	Y
2021-09-11 08:48	BIO-RAD L2	CREA	5.35	0.75	5.85	3.85-6.85	Y
2021-09-11 08:47	BIO-RAD L2	UREA	100	10	166.1	80-120	N
2021-09-11 08:47	BIO-RAD L2	GLU	301	30	276	241-361	Y


Red

QC information

Test time	QC Liquid	Item	Ave.	SD	Result	SD Range	In Control
2021-09-16 10:59	BIO-RAD L2	CALC	11.8	0.8	12.4	10.2-13.4	Y
2021-09-16 10:55	BIO-RAD L2	URIC	10.3	1	8.83	8.3-12.3	Y
2021-09-16 10:55	BIO-RAD L2	GGT	158	20	113.3	118-198	N
2021-09-16 10:03	BIO-RAD L2	HDL	27	3	31.9	21-33	Y
2021-09-16 10:03	BIO-RAD L2	CALC	11.8	0.8	0	10.2-13.4	N
2021-09-16 10:02	BIO-RAD L2	URIC	10.3	1	0	8.3-12.3	N
2021-09-16 10:01	BIO-RAD L2	GGT	158	20	0	118-198	N
2021-09-16 10:01	BIO-RAD L2	ALP	433	40	453	353-513	Y
2021-09-16 10:00	BIO-RAD L2	SGPT	107	23	101	61-153	Y
2021-09-16 10:00	BIO-RAD L2	SGOT	157	25	225	107-207	N
2021-09-16 09:59	BIO-RAD L2	TPP	4.5	0.4	4.27	3.7-5.3	Y
2021-09-16 09:58	BIO-RAD L2	DBIL	0.9	0.09	1.673	0.72-1.08	N
2021-09-16 09:58	BIO-RAD L2	TBIL	4.5	0.45	5.114	3.6-5.4	Y
2021-09-16 09:57	BIO-RAD L2	TGL	115	11.4	93	92.2-137.8	Y
2021-09-16 09:57	BIO-RAD L2	CHOL	106	19	102	68-144	Y
2021-09-16 09:56	BIO-RAD L2	CREA	5.35	0.75	5.28	3.85-6.85	Y
2021-09-16 09:55	BIO-RAD L2	UREA	100	10	103.6	80-120	Y
2021-09-16 09:55	BIO-RAD L2	GLU	301	30	317	241-361	Y

Paul

LAB MONTHLY SUMMARY

Lab Name **COSMOS CLINICS AND DIAGNOSTICS**Lab No **10077**Month **August**Year **2021**Constituent Group **Chemistry I**Date of Result Entered : **19/08/2021**Date of Report Published : **01/09/2021**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE	GOD-POD	Any Analyser (Automation / Semi Automation)	661	208.21	8.82	18.37	228 mg/dl	1.08	1.43
2	UREA	Urease UV / GLDH	Any Analyser (Automation / Semi Automation)	552	76.23	10.49	7.99	80.8 mg/dl	0.57	0.68
3	CREATININE	Jaffes Kinetic - Alkaline picrate	Any Analyser (Automation / Semi Automation)	465	3.19	11.60	0.37	3.45 mg/dl	0.70	0.03
4	T.BILIRUBIN	Diazonium salt (Colorimetric) / Jendrassik	Any Analyser (Automation / Semi Automation)	566	2.65	17.54	0.46	5.23 mg/dl	5.61	0.04
5	T-PROTEIN	Biuret - Colorimetric	Any Analyser (Automation / Semi Automation)	667	5.28	10.68	0.56	5.7 g/dl	0.74	0.04
6	CALCIUM	Arsenazo III	Any Analyser (Automation / Semi Automation)	455	9.86	8.53	0.84	10.8 mg/dl	1.12	0.08
7	PHOSPHORUS	Molybdate UV / Phosphomolybdate complex	Any Analyser (Automation / Semi Automation)	313	3.62	14.76	0.54	4.39 mg/dl	1.44	0.06
8	URIC ACID	Enzymatic / Uricase Colorimetric	Any Analyser (Automation / Semi Automation)	541	4.78	16.55	0.79	5.5 mg/dl	0.91	0.07
9	CHOLESTEROL	CHOD-PAP	Any Analyser (Automation / Semi Automation)	632	108.97	10.30	11.23	110 mg/dl	0.09	0.89
10	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Any Analyser (Automation / Semi Automation)	623	150.12	11.79	17.70	184 mg/dl	1.91	1.42
11	HDL CHO	Direct method / Enzymatic colorimetric	Any Analyser (Automation / Semi Automation)	442	27.12	18.97	5.14	50.1 mg/dl	4.47	0.49
12	SODIUM	ISE - Direct	Any Analyser (Automation / Semi Automation)	561	138.42	3.00	4.16	143 mmol/L	1.10	0.35
13	POTASSIUM	ISE - Direct	Any Analyser (Automation / Semi Automation)	560	4.02	5.08	0.20	4.2 mmol/L	0.88	0.02
14	CHLORIDE	ISE - Direct	Any Analyser (Automation / Semi Automation)	435	104.23	4.22	4.40	99 mmol/L	-1.19	0.42
15	AST	UV kinetic without PLP (P-5-P)	Any Analyser (Automation / Semi Automation)	366	60.40	14.27	8.62	62 U/L	0.19	0.90
16	ALT	UV kinetic without PLP (P-5-P)	Any Analyser (Automation / Semi Automation)	330	88.23	18.94	16.71	103 U/L	0.88	1.84
17	ALP	PNP AMP kinetic	Any Analyser (Automation / Semi Automation)	367	495.15	16.84	83.38	356 U/L	-1.67	8.70
18	AMYLASE	CNPG3	Any Analyser (Automation / Semi Automation)	371	154.87	18.64	28.87	164.2 U/L	0.32	3.00
19	IRON	Ferro Zine (No Protein Removal)	Any Analyser (Automation / Semi Automation)	51	177.94	10.06	17.90	619.6 ug/dl	24.67	5.01

			Automation)							
20	MAGNESIUM	Xylidyl Blue	Any Analyser (Automation / Semi Automation)	139	2.76	16.44	0.45	3.5 mg/dl	1.63	0.08

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss

CMC EQAS Co-Ordinator

Christian Medical College, Vellore

***** End of Report *****



COSMOS CLINICS AND DIAGNOSTICS
#503, SMV Arcade, Andrahalli Main Road, Syndicate
Bank Layout, Bengaluru- 560091

Title			
Document Type	Format	Document Number	CCD.LAB.FR.002
Prepared By	Mamatha	Checked by/Approved	Tanveer /Dr.Nirupama.S
Revision No.	0	Issue Date	

Laboratory: ONCOSTEM DIAGNOSTICS PVT. LTD.

Date:

Sept '2021

Area of NC occurrence:

PT EQAS.

NC No.

PT EQAS: 003.

Major/minor

Non-Conformity (NC) Raised:

- August EQAS results showed T.B to be within HDL as an outlier.

Immediate corrective action:

- Calibrated and QC run, Values found within range.

Root cause analysis:

- Random Error, However repeat susceptibility due to lack of cooling noticed.

CORRECTIVE ACTION PROPOSED:

- To go ahead for fully automated Analyze with inbuilt freezer.

EVIDENCE OF IMPLEMENTATION FOR CONTINUAL IMPROVEMENT

Attached.

VERIFIED BY:


Purp one!

PHOENIX DIAGNOSTICSCalibration report

Item: HDL

Calibration method: Two-point

Calibration time: 2021-10-27 12:13:34




Test Date	Response	Conc.
2021-10-27 12:01:12	122.5411	0
2021-10-27 12:01:48	1354.0338	48.9

Test person:

Reviewer:

Print Time: 2021-11-11 13:35:37

Reaction Curve




PHOENIX DIAGNOSTICS Calibration report

Item: HDL

Calibration method: Two-point

Calibration time: 2021-09-09 15:52:08



Test Date	Response	Conc.
2021-09-09 15:39:46	80.388	0
2021-09-09 15:40:22	-1054.5851	48.9

2021-11-11 13:31:1

Test person:


Review er:

Print Time: 2021-11-11 13:31:16

1102

FormCalibReq

Reaction Curve




PHOENIX DIAGNOSTICS Calibration report

Item: HDL

Calibration method: Two-point

Calibration time: 2021-09-14 18 08 36



2021-11-11 13:32:4

Test Date	Response	Conc.
2021-09-14 17:56:00	90.9101	0
2021-09-14 17:56:36	986.0464	48.9


Test person:

Reviewer:

Print Time: 2021-11-11 13:32:43

FormCallbReq

Reaction Curve




PHOENIX DIAGNOSTICS Calibration report

Item: HDL

Calibration method: Two-point

Calibration time: 2021-09-15 19:13:03



Test Date	Response	Conc.
2021-09-15 19:00:41	87.0722	0
2021-09-15 19:01:17	902.2998	48.9

2021-11-11 13:34:3


Test person:

Review er:

Print Time: 2021-11-11 13:34:37

FormCallbReq

Reaction Curve



QC information


Test time	QC Liquid	Item	Ave.	SD	Result	SD Range	In Control
2021-08-29 14:29	BIO-RAD L2	<u>HDL</u>	27	3	<u>32.2</u>	21-33	Y
2021-08-29 14:29	BIO-RAD L2	CALC	11.8	0.8	2.9	10.2-13.4	N
2021-08-29 14:28	BIO-RAD L2	URIC	10.3	1	0	8.3-12.3	N
2021-08-29 14:27	BIO-RAD L2	GGT	158	20	188.8	118-198	Y
2021-08-29 14:27	BIO-RAD L2	ALP	433	40	415	353-513	Y
2021-08-29 12:10	BIO-RAD L2	HDL	27	3	0	21-33	N
2021-08-29 12:10	BIO-RAD L2	CALC	11.8	0.8	0	10.2-13.4	N
2021-08-29 12:09	BIO-RAD L2	URIC	10.3	1	0	8.3-12.3	N
2021-08-29 12:08	BIO-RAD L2	GGT	158	20	0	118-198	N
2021-08-29 12:08	BIO-RAD L2	ALP	433	40	0	353-513	N
2021-08-29 12:07	BIO-RAD L2	SGPT	107	23	29	61-153	N
2021-08-29 12:07	BIO-RAD L2	SGOT	157	25	170	107-207	Y
2021-08-29 12:06	BIO-RAD L2	TPR	4.5	0.4	1.63	3.7-5.3	N
2021-08-29 12:05	BIO-RAD L2	DBIL	0.9	0.09	1.798	0.72-1.08	N
2021-08-29 12:05	BIO-RAD L2	TBIL	4.5	0.45	5.467	3.6-5.4	N
2021-08-29 12:04	BIO-RAD L2	TGL	115	11.4	100	92.2-137.8	Y
2021-08-29 12:04	BIO-RAD L2	CHOL	106	19	109	68-144	Y
2021-08-29 12:03	BIO-RAD L2	CREA	5.35	0.75	5.25	3.85-6.85	Y
2021-08-29 12:02	BIO-RAD L2	UREA	100	10	130.15	80-120	N
2021-08-29 12:02	BIO-RAD L2	GLU	301	30	316	241-361	Y

PHOENIX DIAGNOSTICS Calibration report

Item: TBIL

Calibration method: Two-point

Calibration time: 2021-09-14 18:09:12



Test Date	Response	Conc.
2021-09-14 17:58:29	13.9723	0
2021-09-14 17:59:08	2142.5462	4.49

Deed

Test person:


Review er:

Print Time: 2021-11-12 12:58:00

FormCalibReq

ABS

Reaction Curve



Done

QC information

Test time	QC Liquid	Item	Ave.	SD	Result	SD Range	In Control
2021-09-14 18:19	BIO-RAD L2	HDL	27	3	0.8	21-33	N
2021-09-14 18:18	BIO-RAD L2	CALC	11.8	0.8	22	10.2-13.4	N
2021-09-14 18:17	BIO-RAD L2	URIC	10.3	1	9.62	8.3-12.3	Y
2021-09-14 18:17	BIO-RAD L2	GGT	158	20	127.9	118-198	Y
2021-09-14 18:16	BIO-RAD L2	ALP	433	40	0	353-513	N
2021-09-14 18:16	BIO-RAD L2	SGPT	107	23	1608	61-153	N
2021-09-14 18:15	BIO-RAD L2	SGOT	157	25	17956	107-207	N
2021-09-14 18:14	BIO-RAD L2	TPR	4.5	0.4	156.1	3.7-5.3	N
2021-09-14 18:14	BIO-RAD L2	DBIL	0.9	0.09	1.533	0.72-1.08	N
2021-09-14 18:13	BIO-RAD L2	TBIL	4.5	0.45	2.856	3.6-5.4	N
2021-09-14 18:13	BIO-RAD L2	TGL	115	11.4	91	92.2-137.8	N
2021-09-14 18:12	BIO-RAD L2	CHOL	106	19	0	68-144	N
2021-09-14 18:11	BIO-RAD L2	CREA	5.35	0.75	6.07	3.85-6.85	Y
2021-09-14 18:11	BIO-RAD L2	UREA	100	10	111.51	80-120	Y
2021-09-14 18:10	BIO-RAD L2	GLU	301	30	289	241-361	Y
2021-09-14 09:38	BIO-RAD L2	HDL	27	3	28.4	21-33	Y
2021-09-14 09:37	BIO-RAD L2	CALC	11.8	0.8	12	10.2-13.4	Y
2021-09-14 09:36	BIO-RAD L2	URIC	10.3	1	7.62	8.3-12.3	N
2021-09-14 09:36	BIO-RAD L2	GGT	158	20	156.8	118-198	Y
2021-09-14 09:35	BIO-RAD L2	ALP	433	40	389	353-513	Y
2021-09-14 09:34	BIO-RAD L2	SGPT	107	23	83	61-153	Y
2021-09-14 09:34	BIO-RAD L2	SGOT	157	25	157	107-207	Y
2021-09-14 09:33	BIO-RAD L2	TPR	4.5	0.4	1.49	3.7-5.3	N
2021-09-14 09:33	BIO-RAD L2	DBIL	0.9	0.09	1.48	0.72-1.08	N
2021-09-14 09:32	BIO-RAD L2	TBIL	4.5	0.45	<u>5.464</u>	3.6-5.4	N
2021-09-14 09:31	BIO-RAD L2	TGL	115	11.4	91	92.2-137.8	N
2021-09-14 09:31	BIO-RAD L2	CHOL	106	19	100	68-144	Y
2021-09-14 09:30	BIO-RAD L2	CREA	5.35	0.75	5.43	3.85-6.85	Y
2021-09-14 09:30	BIO-RAD L2	UREA	100	10	260.21	80-120	N
2021-09-14 09:29	BIO-RAD L2	GLU	301	30	281	241-361	Y

QC information

Test time	QC Liquid	Item	Ave.	SD	Result	SD Range	In Control
2021-08-29 14:29	BIO-RAD L2	<u>HDL</u>	27	3	<u>32.2</u>	21-33	Y
2021-08-29 14:29	BIO-RAD L2	CALC	11.8	0.8	2.9	10.2-13.4	N
2021-08-29 14:28	BIO-RAD L2	URIC	10.3	1	0	8.3-12.3	N
2021-08-29 14:27	BIO-RAD L2	GGT	158	20	188.8	118-198	Y
2021-08-29 14:27	BIO-RAD L2	ALP	433	40	415	353-513	Y
2021-08-29 12:10	BIO-RAD L2	HDL	27	3	0	21-33	N
2021-08-29 12:10	BIO-RAD L2	CALC	11.8	0.8	0	10.2-13.4	N
2021-08-29 12:09	BIO-RAD L2	URIC	10.3	1	0	8.3-12.3	N
2021-08-29 12:08	BIO-RAD L2	GGT	158	20	0	118-198	N
2021-08-29 12:08	BIO-RAD L2	ALP	433	40	0	353-513	N
2021-08-29 12:07	BIO-RAD L2	SGPT	107	23	29	61-153	N
2021-08-29 12:07	BIO-RAD L2	SGOT	157	25	170	107-207	Y
2021-08-29 12:06	BIO-RAD L2	TPR	4.5	0.4	1.63	3.7-5.3	N
2021-08-29 12:05	BIO-RAD L2	DBIL	0.9	0.09	1.798	0.72-1.08	N
2021-08-29 12:05	BIO-RAD L2	TBIL	4.5	0.45	5.467	3.6-5.4	N
2021-08-29 12:04	BIO-RAD L2	TGL	115	11.4	100	92.2-137.8	Y
2021-08-29 12:04	BIO-RAD L2	CHOL	106	19	109	68-144	Y
2021-08-29 12:03	BIO-RAD L2	CREA	5.35	0.75	5.25	3.85-6.85	Y
2021-08-29 12:02	BIO-RAD L2	UREA	100	10	130.15	80-120	N
2021-08-29 12:02	BIO-RAD L2	GLU	301	30	316	241-361	Y

VIEW LAB MONTHLY SUMMARY

Lab Name	COSMOS CLINICS AND DIAGNOSTICS		Details About Robust Analysis							
Lab No	10077		Detail About Monthly Summary							
Month	September		Detail about SDI							
Year	2021									
Constituent Group	Chemistry I									
Click on the analyte to view Graphical Data			All Analyser Result		Print					
			Date of Result Entered : 16/09/2021							
			Date of Report Published : 05/10/2021							
SI.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	GLUCOSE	GOD-POD	Any Analyser (Automation / Semi Automation)	709	81.33	9.78	7.96	78 mg/dl	-0.42	0.60
2	UREA	Urease UV / GLDH	Any Analyser (Automation / Semi Automation)	578	93.58	10.00	9.36	131 mg/dl	4.00	0.78
3	CREATININE	Jaffes Kinetic - Alkaline picrate	Any Analyser (Automation / Semi Automation)	479	7.01	11.04	0.77	7.9 mg/dl	1.15	0.07
4	T.BILIRUBIN	Diazonium salt (Colorimetric) / Jendrassik	Any Analyser (Automation / Semi Automation)	541	1.75	16.28	0.28	4 mg/dl	8.04	0.02
5	T-PROTEIN	Biuret - Colorimetric	Any Analyser (Automation / Semi Automation)	666	5.37	9.63	0.52	4.1 g/dl	-2.46	0.04
6	CALCIUM	Arsenazo III	Any Analyser (Automation / Semi Automation)	504	8.67	9.82	0.85	9.4 mg/dl	0.86	0.08
7	PHOSPHORUS	Molybdate UV / Phosphomolybdate complex	Any Analyser (Automation / Semi Automation)	351	3.35	17.69	0.59	4 mg/dl	1.10	0.06
8	URIC ACID	Enzymatic / Uricase Colorimetric	Any Analyser (Automation / Semi Automation)	633	7.81	11.90	0.93	9.68 mg/dl	2.01	0.07
9	CHOLESTEROL	CHOD-PAP	Any Analyser (Automation / Semi Automation)	667	111.25	8.59	9.55	117 mg/dl	0.60	0.74
10	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Any Analyser (Automation / Semi Automation)	615	245.32	10.30	25.26	350 mg/dl	4.14	2.04
11	HDL CHO	Direct method / Enzymatic colorimetric	Any Analyser (Automation / Semi Automation)	443	27.77	16.06	4.46	24.1 mg/dl	-0.82	0.42
12	SODIUM	ISE - Direct	Any Analyser (Automation / Semi Automation)	605	127.77	3.62	4.62	136 mmol/L	1.78	0.38
13	POTASSIUM	ISE - Direct	Any Analyser (Automation / Semi Automation)	608	5.30	5.64	0.30	5.6 mmol/L	1.00	0.02
14	CHLORIDE	ISE - Direct	Any Analyser (Automation / Semi Automation)	472	103.50	4.05	4.19	101 mmol/L	-0.60	0.39

			Automation)							
15	AST	UV kinetic without PLP (P-5-P)	Any Analyser (Automation / Semi Automation)	394	86.33	11.92	10.29	105 U/L	1.81	1.04
16	ALT	UV kinetic without PLP (P-5-P)	Any Analyser (Automation / Semi Automation)	355	257.48	16.34	42.08	296 U/L	0.92	4.47
17	ALP	PNP AMP kinetic	Any Analyser (Automation / Semi Automation)	452	183.43	18.11	33.22	240 U/L	1.70	3.13
18	AMYLASE	CNPG3	Any Analyser (Automation / Semi Automation)	390	73.03	16.72	12.21	52.3 U/L	-1.70	1.24
19	MAGNESIUM	Xylidyl Blue	Any Analyser (Automation / Semi Automation)	155	2.40	15.43	0.37	3 mg/dl	1.62	0.06

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out



COSMOS CLINICS AND DIAGNOSTICS
#503, SMV Arcade, Andrahalli Main Road, Syndicate
Bank Layout, Bengaluru- 560091

Title			
Document Type	Format	Document Number	CCD.LAB.FR.002
Prepared By	Mamatha	Checked by/Approved	Tanveer /Dr.Nirupama.S
Revision No.	0	Issue Date	

Laboratory: ONCOSTEM DIAGNOSTICS PVT. LTD.

Date: 03 Sept '2021
Area of NC occurrence: PT Septem Octlier.

NC No. PT.NC:002.
Major/minor


Non-Conformity (NC) Raised:
TGL, ma, HDL, B. outlier. —

Immediate corrective action :
Changed the ket from printer to Euro ~~printer~~ compatible with machine.

Root cause analysis:
- Lamp issues & reagent instability
Corrected

CORRECTIVE ACTION PROPOSED:
- Proposal for in built freezer automated Biochemistry Analyser as recommended


EVIDENCE OF IMPLEMENTATION FOR CONTINUAL IMPROVEMENT:
Attached

VERIFIED BY:

Dr. Nirupama S.

GENIX DIAGNOSTICS Calibration report

Response

Curve



2021-11-05 11:31:2

Test Date	Response	Conc
2021-09-15 17:16:40	15 0243	0
2021-09-15 17:17:16	1958.9194	140


1-09-15 18:15	BIO-RAD L2	GGT	158	40	0	61-153
1-09-15 18:15	BIO-RAD L2	ALP	433	23	0	107-207
1-09-15 18:14	BIO-RAD L2	SGPT	107	25	29	3.7-5.3
1-09-15 18:14	BIO-RAD L2	SGOT	157	0.4	0	0.72-1.08
1-09-15 18:13	BIO-RAD L2	TPR	4.5	0.09	1.661	3.6-5.4
1-09-15 18:12	BIO-RAD L2	DBIL	0.9	0.45	5.206	92.2-137.8
1-09-15 18:12	BIO-RAD L2	TBIL	4.5	11.4	97	68-144
1-09-15 18:11	BIO-RAD L2	<u>TGL</u>	115	19	109	3.85-6.85
1-09-15 18:10	BIO-RAD L2	CHOL	106	0.75	0	80-120
1-09-15 18:10	BIO-RAD L2	CREA	5.35	10	23009.55	241-361
1-09-15 18:09	BIO-RAD L2	UREA	100	30	283	68-144
1-09-15 18:09	BIO-RAD L2	GLU	301	19	0	3.85-6.85
1-09-15 18:08	BIO-RAD L2	CHOL	106	0.75	0	80-120
1-09-15 10:12	BIO-RAD L2	CREA	5.35	10	0	241-361
1-09-15 10:12	BIO-RAD L2	UREA	100	30	0	107-207
1-09-15 10:11	BIO-RAD L2	GLU	301	25	0	3.7-5.3
1-09-15 10:11	BIO-RAD L2	SGOT	157	0.4	0	0.72-1.08
1-09-15 09:11	BIO-RAD L2	TPR	4.5	0.09	0	3.6-5.4
1-09-15 09:11	BIO-RAD L2	DBIL	0.9	0.45	0	92.2-137.8
1-09-15 09:10	BIO-RAD L2	TBIL	4.5	11.4	0	68-144
1-09-15 09:09	BIO-RAD L2	TGL	115	19	30	3.85-6.85
1-09-15 09:09	BIO-RAD L2	CHOL	106	0.75	0	80-120
1-09-15 09:08	BIO-RAD L2	CREA	5.35	10	0	241-361
1-09-15 09:08	BIO-RAD L2	UREA	100	30	258	
1-09-15 09:07	BIO-RAD L2	GLU	301			
1-09-15 09:06	BIO-RAD L2					

PHOENIX DIAGNOSTICS Calibration report

Item: UREA

Calibration time: 2021-09-07 14:38:06

Calibration method: Two-point
Curve




2021-11-05 11:38:2

Test Date	Response	Conc
2021-09-07 14:27:00	94.6291	0
2021-09-07 14:27:36	1435.6153	103

FormCalibReq


Reaction Curve



Test time
2021-09-07 14:38
2021-09-07 14:38

QC information

Test time	QC Liquid	Item	Ave	SD	Result	SD Range	In Control
2021-09-07 14:38	BIO-RAD L2	HDL	27	3	30.8	21-33	Y
2021-09-07 14:37	BIO-RAD L2	CALC	11.8	0.8	12.8	10.2-13.4	Y
2021-09-07 14:36	BIO-RAD L2	URIC	10.3	1	7.52	8.3-12.3	N
2021-09-07 14:36	BIO-RAD L2	GGT	158	20	169.4	118-198	Y
2021-09-07 14:35	BIO-RAD L2	ALP	433	40	387	353-513	Y
2021-09-07 14:34	BIO-RAD L2	SGPT	107	23	24	61-153	N
2021-09-07 14:34	BIO-RAD L2	SGOT	157	25	161	107-207	Y
2021-09-07 14:33	BIO-RAD L2	TPR	4.5	0.4	4.26	3.7-5.3	Y
2021-09-07 14:33	BIO-RAD L2	DBIL	0.9	0.09	1.733	0.72-1.08	N
2021-09-07 14:32	BIO-RAD L2	TBIL	4.5	0.45	4.795	3.6-5.4	Y
2021-09-07 14:31	BIO-RAD L2	TGL	115	11.4	123	92.2-137.8	Y
2021-09-07 14:31	BIO-RAD L2	CHOL	106	19	107	68-144	Y
2021-09-07 14:30	BIO-RAD L2	CREA	5.35	0.75	6.05	3.85-6.85	Y
2021-09-07 14:30	BIO-RAD L2	<u>UREA</u>	100	10	<u>110.78</u>	80-120	Y
2021-09-07 14:29	BIO-RAD L2	<u>GLU</u>	301	30	<u>293</u>	241-361	Y



Test Date	Response	Conc.
2021-09-14 17:58:29	13.9723	0
2021-09-14 17:59:08	2142.5462	4.49

2021-11-05 11:39:23


Reviewer:

Print Time: 2021-11-05 11:39:23

Test person:

FormCalibReq

Reaction Curve



est time
2021-09-14 18:19
QC Liquid
BIO-5
2021-09-14 18:18
2021-09-14 18:18
2021-09-14 18:18

QC information

Test time	QC Liquid	Item	Ave.	SD	Result	SD Range	In Control
2021-09-14 18:19	BIO-RAD L2	HDL	27	3	0.8	21-33	N
2021-09-14 18:18	BIO-RAD L2	CALC	11.8	0.8	22	10.2-13.4	N
2021-09-14 18:17	BIO-RAD L2	URIC	10.3	1	9.62	8.3-12.3	Y
2021-09-14 18:17	BIO-RAD L2	GGT	158	20	127.9	118-198	Y
2021-09-14 18:16	BIO-RAD L2	ALP	433	40	0	353-513	N
2021-09-14 18:16	BIO-RAD L2	SGPT	107	23	1608	61-153	N
2021-09-14 18:15	BIO-RAD L2	SGOT	157	25	17956	107-207	N
2021-09-14 18:14	BIO-RAD L2	TPR	4.5	0.4	156.1	3.7-5.3	N
2021-09-14 18:14	BIO-RAD L2	DBIL	0.9	0.09	1.533	0.72-1.08	N
2021-09-14 18:14	BIO-RAD L2	TBIL	4.5	0.45	2.856	3.6-5.4	N
2021-09-14 18:13	BIO-RAD L2	TBIL	115	11.4	91	92.2-137.8	N
2021-09-14 18:13	BIO-RAD L2	TGL	106	19	0	68-144	Y
2021-09-14 18:12	BIO-RAD L2	CHOL	106	19	0	68-144	Y
2021-09-14 18:11	BIO-RAD L2	CHOL	106	19	6.07	3.85-6.85	Y
2021-09-14 18:11	BIO-RAD L2	CREA	5.35	0.75	6.07	3.85-6.85	Y
2021-09-14 18:11	BIO-RAD L2	UREA	100	10	111.51	80-120	Y
2021-09-14 18:10	BIO-RAD L2	UREA	100	10	289	241-361	Y
2021-09-14 18:10	BIO-RAD L2	GLU	301	30	28.4	21-33	Y
2021-09-14 09:38	BIO-RAD L2	HDL	27	3	12	10.2-13.4	Y
2021-09-14 09:37	BIO-RAD L2	HDL	27	3	12	10.2-13.4	N
2021-09-14 09:37	BIO-RAD L2	CALC	11.8	0.8	7.62	8.3-12.3	Y
2021-09-14 09:36	BIO-RAD L2	URIC	10.3	1	156.8	118-198	Y
2021-09-14 09:36	BIO-RAD L2	GGT	158	20	389	353-513	Y
2021-09-14 09:35	BIO-RAD L2	GGT	158	20	389	353-513	Y
2021-09-14 09:35	BIO-RAD L2	ALP	433	40	83	61-153	Y
2021-09-14 09:34	BIO-RAD L2	SGPT	107	23	157	107-207	Y
2021-09-14 09:34	BIO-RAD L2	SGOT	157	25	157	107-207	N
2021-09-14 09:33	BIO-RAD L2	TPR	4.5	0.4	1.49	3.7-5.3	N
2021-09-14 09:33	BIO-RAD L2	TPR	4.5	0.4	1.48	3.7-5.3	N
2021-09-14 09:33	BIO-RAD L2	DBIL	0.9	0.09	1.48	0.72-1.08	N
2021-09-14 09:32	BIO-RAD L2	DBIL	0.9	0.09	1.48	0.72-1.08	N
2021-09-14 09:32	BIO-RAD L2	TBIL	4.5	0.45	5.464	3.6-5.4	N
2021-09-14 09:32	BIO-RAD L2	TBIL	4.5	0.45	91	92.2-137.8	N
2021-09-14 09:31	BIO-RAD L2	TGL	115	11.4	91	92.2-137.8	N
2021-09-14 09:31	BIO-RAD L2	TGL	115	11.4	100	68-144	N
2021-09-14 09:31	BIO-RAD L2	CHOL	106	19	100	68-144	N
2021-09-14 09:30	BIO-RAD L2	CHOL	106	19	5.43	3.85-6.85	N
2021-09-14 09:30	BIO-RAD L2	CHOL	106	19	5.43	3.85-6.85	N
2021-09-14 09:30	BIO-RAD L2	CREA	5.35	0.75	260.21	80-120	N
2021-09-14 09:30	BIO-RAD L2	UREA	100	10	281	241-361	N
2021-09-14 09:29	BIO-RAD L2	UREA	100	10	281	241-361	N
2021-09-14 09:29	BIO-RAD L2	GLU	301	30	281	241-361	N



Preventive Maintenance Report (PM Visit)

CUSTOMER NAME : COSMOS DIAGNOSTIC **Date :** 08 08 2021

Address :
 Door No #503 Street Anandhalli Main Road
 Area SMV Arcade District Bengaluru Pincode 560091
 State Karnataka Landmark Syndicate Bank Layout
 Contact No 9945670868 Email lab@cosmosdiagnostics.in

PM Visit Number (Ex : 1st Visit or 2nd Visit) 2nd

WARRANTY
 AMC

Voltage : Line / Neutral 230 V, Line/Earth 230V, Earth/Neutral : 0 V (Should not >005)
Power Source : Direct / UPS / CVT

Instrument : J Magic M7 **WORK CARRIED OUT AT**
 SITE SERVICE CENTRE

SERIAL NO : 3 1 1 8 7 4 1 7

NEXT CALIBRATION 09/08/2022

PARAMETERS

T: OK A: OK Pump Speed: OK
 Pump: OK Front Board: OK Gain Cal: OK Pump Cal: OK

PARAMETERS : 950ul
 Pump Calibration : 500 ul (800-1200 ul)

FILTER STATUS :

Filters in nm	340	405	505	546	578	600	630	
Gain	15	OK	OK	OK	OK	OK	OK	
A/D Value	20	OK	OK	OK	OK	OK	OK	
Abs of D/Water	10	OK	OK	OK	OK	OK	OK	

Instrument Status :

Outer Body : OK Display : OK Key Pad : OK Pump Motor : OK Printer : OK
 Aspiration Button : OK Tubing Set : OK Flow Cell : OK Internal Connectors : OK Software : OK
 Remarks on the defective part / parts if any : OK

ENGINEER'S REMARKS Calibration Completed
 Calibration Period 08/08/21 to 08/08/22

CUSTOMER'S FEEDBACK

ENGINEER'S SIGNATURE & NAME :
 Nikhil K

Note : Very Important

 Seal & Signature of the Customer

VIEW LAB MONTHLY SUMMARY

Lab Name	COSMOS CLINICS AND DIAGNOSTICS		Details About Robust Analysis							
Lab No	10077		Detail About Monthly Summary							
Month	October		Detail about SDI							
Year	2021									
Constituent Group	Chemistry I									
Click on the analyte to view Graphical Data			All Analyser Result		Print					
			Date of Result Entered : 19/10/2021							
			Date of Report Published : 01/11/2021							
Sl.No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	GLUCOSE	GOD-POD	Any Analyser (Automation / Semi Automation)	702	315.96	8.56	27.03	311 mg/dl	-0.18	2.04
2	UREA	Urease UV / GLDH	Any Analyser (Automation / Semi Automation)	594	43.89	10.67	4.69	43.9 mg/dl	0.00	0.38
3	CREATININE	Jaffes Kinetic - Alkaline picrate	Any Analyser (Automation / Semi Automation)	486	1.13	13.39	0.15	1.31 mg/dl	1.19	0.01
4	T.BILIRUBIN	Diazonium salt (Colorimetric) / Jendrassik	Any Analyser (Automation / Semi Automation)	557	6.61	11.91	0.79	8.6 mg/dl	2.52	0.07
5	T-PROTEIN	Biuret - Colorimetric	Any Analyser (Automation / Semi Automation)	677	5.29	9.77	0.52	4 g/dl	-2.50	0.04
6	CALCIUM	Arsenazo III	Any Analyser (Automation / Semi Automation)	506	9.53	8.10	0.77	10.6 mg/dl	1.39	0.07
7	PHOSPHORUS	Molybdate UV / Phosphomolybdate complex	Any Analyser (Automation / Semi Automation)	319	5.40	11.35	0.61	5.53 mg/dl	0.21	0.07
8	URIC ACID	Enzymatic / Uricase Colorimetric	Any Analyser (Automation / Semi Automation)	681	5.40	16.16	0.87	6.25 mg/dl	0.97	0.07
9	CHOLESTEROL	CHOD-PAP	Any Analyser (Automation / Semi Automation)	677	105.70	11.01	11.63	108 mg/dl	0.20	0.89
10	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Any Analyser (Automation / Semi Automation)	597	97.58	14.69	14.34	96 mg/dl	-0.11	1.17
11	HDL CHO	Direct method / Enzymatic colorimetric	Any Analyser (Automation / Semi Automation)	438	27.23	16.03	4.37	22.1 mg/dl	-1.18	0.42
12	SODIUM	ISE - Direct	Any Analyser (Automation / Semi Automation)	618	135.54	3.02	4.09	145 mmol/L	2.31	0.33
13	POTASSIUM	ISE - Direct	Any Analyser (Automation / Semi Automation)	627	3.25	6.34	0.21	3.5 mmol/L	1.21	0.02
14	CHLORIDE	ISE - Direct	Any Analyser (Automation / Semi Automation)	486	107.16	4.46	4.78	105 mmol/L	-0.45	0.43
15	AST	UV kinetic without PLP (P-5-P)	Any Analyser (Automation / Semi Automation)	383	208.31	10.67	22.23	199 U/L	-0.42	2.27
16	ALT	UV kinetic without PLP	Any Analyser	374	76.11	17.16	13.06	93 U/L	1.29	1.35

		(P-5-P)	(Automation / Semi Automation)							
17	ALP	PNP AMP kinetic	Any Analyser (Automation / Semi Automation)	426	93.76	18.94	17.76	242 U/L	8.35	1.72
18	AMYLASE	CNPG3	Any Analyser (Automation / Semi Automation)	407	125.52	17.70	22.21	176.7 U/L	2.3	2.20
19	IRON	Ferro Zine (No Protein Removal)	Any Analyser (Automation / Semi Automation)	51	186.23	8.19	15.25	550.9 ug/dl	23.91	4.27
20	MAGNESIUM	Xylidyl Blue	Any Analyser (Automation / Semi Automation)	157	1.84	17.49	0.32	2.41 mg/dl	1.77	0.05

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.


Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out



COSMOS CLINICS AND DIAGNOSTICS
#503, SMV Arcade, Andrahalli Main Road, Syndicate
Bank Layout, Bengaluru- 560091

Title			
Document Type	Format	Document Number	CCD.LAB.FR.002
Prepared By	Mamatha	Checked by/Approved	Tanveer /Dr.Nirupama.S
Revision No.	0	Issue Date	

Laboratory: ONCOSTEM DIAGNOSTICS PVT. LTD.	
Date: 12/11/2021	Area of NC occurrence: PT EQAS : OF October.2021
NC No.	Major/minor
Non-Conformity (NC) Raised: Alkaline Phosphatase out of spec.	
Immediate corrective action : - Calibration of Alkaline phosphatase and QC to be carried out.	
Root cause analysis: - Random Error.	
CORRECTIVE ACTION PROPOSED: Calibration and QC to be carried photo out once in 3 days since Alkaline Phosphatase is a sensitive parameter.	
EVIDENCE OF IMPLEMENTATION FOR CONTINUAL IMPROVEMENT: Will be Attached upon carrying out Calibration and QC	
VERIFIED BY:  12/11/2021.	