

LAB MONTHLY SUMMARY



Lab Name **GOWTHAM DIAGNOSTIC CENTRE** Lab No **13669**
 Month **April** Year **2021**
 Constituent Group **Chemistry II**



Date of Result Entered : 17/04/2021

Date of Report Published : 01/05/2021

Sl.No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	879	310.68	9.27	28.80	307.6 mg/dl	-0.11	1.94
2	UREA II	Others (Urease Berthelot / Ned Dye) II	Any Analyser (Automation / Semi Automation)	847	30.12	15.52	4.68	30 mg/dl	-0.03	0.32
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	718	6.46	12.06	0.78	6.8 mg/dl	0.44	0.06
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	791	3.42	17.30	0.59	3.9 mg/dl	0.81	0.04
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	680	5.50	8.70	0.48	6 g/dl	1.04	0.04
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	685	3.30	9.80	0.32	3.2 g/dl	-0.31	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	690	4.59	19.39	0.89	4.8 mg/dl	0.24	0.07
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	944	117.93	11.98	14.13	116.5 mg/dl	-0.10	0.92
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	790	224.07	12.54	28.10	231.9 mg/dl	0.28	2.00

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator

Christian Medical College, Vellore

***** End of Report *****