

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN. 33AAHCA8053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140

CALIBRATION CERTIFICATE



Customer Name & Address:	M/S. Gowtham Diagnostic Centre, 744-B, Uthukuli Main Road, OPP. GH, Kunnathur, Tirupur Dt. - 638 103.	Certificate No:	AMS/LAB/2889728
		Issuing Date:	14/08/2021
		Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Refrigerator -Reagent	Id. No:	GDC/LAB/REFR/001
Make:	Electrolux	Department:	Laboratory
Model:	EJ163/2016	Calibrated On:	13/08/2021
Sl. No:	200387	Calibration Due Suggested:	12/08/2022
Type:	Drug / Reagent	DUC Condition on Receipt:	OK
Range & Least Count:	0 - 6°C / LC 1°C		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	RTD Thermometer	Center / RTD 375	PC/01/21/ATL/0107-01	22/01/2022
2	Thermo Hygrometer	Beetech/088	PC/01/21/ATL/0194-01	22/01/2022

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

Test Results:

SL.No.	Place	DUC Reading (°C)	Standard Reading (°C)	Correction (°C)
1	Freezer	FRZ 1	0.85	0
2		FRZ 2	1.07	0
3	Chiller	CHI 1	2.37	0
4		CHI 2	2.48	0
5	Chamber I	CHA 1	3.51	0
6		CHA 2	3.89	0
7	Chamber II	CHA 1	4.26	0
8		CHA 2	4.38	0
9	Chamber III	CHA 1	5.21	0
10		CHA 2	5.17	0

Allowable Tolerance: User Defined.

Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:
(Calibration Engineer)

B. P. G.



Approved by:
(Technical Manager)

B. P. G.