









Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments
 304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092
 Phone / Fax : (011) 22453259, 22434362, 9599482307
 E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF DIGITAL TEMPERATURE INDICATOR WITH SENSOR

Certificate No:	21100010234	Issued On	11/08/2021
ULR No:	CC214621100010234F	Calibration Date	10/08/2021
Job Identification No:	ICS/C/T/08/10234	Next Due Date	09/08/2022
Ref. No:	SRF- Dated: 07/08/2021		

CALIBRATED FOR: Dr. Dham's Diagnostic Centre
 3856/23, Reghar Pura,
 Karol Bagh, New Delhi-110005.

EQUIPMENT DESCRIPTION

Name	Digital Temperature Indicator with Sensor		
Make/Model	Aceteq / ST-9283B	Visual Inspection	Ok
Serial No.	---	I.D. No.	---
Range	-50°C to 300°C	Least Count	0.1°C
Location	---	Calibration Site	In-Lab

ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 3.0° C	Humidity	50 ± 10 % RH
-------------	------------------	----------	--------------

STANDARD REFERENCE DETAILS

Traceable to National Standard

Name of the Standard	Make	Certificate No.	Calibration Agency/Traceability	Date of Calibration	Valid up to
Digital Temp. Indicator with Sensor	"Lutron"	21100000001	Indian Calibration Services, Delhi	01/01/2021	01/01/2022

USED EQUIPMENT DETAILS Methanol Bath, Water Bath and Dry Block Calibrator
PRINCIPLE/METHODOLOGY OF CALIBRATION: As per Calibration Procedure No.: ICS-CAL/SOP-TH02, ITS-90 (Comparison Method)


RESULTS: Thermal Calibration

On Working Range:	U.U.C. READING (In °C)	CORRECTED VALUE OF STANDARD (in °C) (Avg.)	Uncertainty at approx 95% C.L and coverage factor k=2 (in °C)
	-030.0	-029.88	± 0.20
	050.0	050.14	± 0.30
	100.0	100.21	± 0.49
	200.0	200.4	± 0.49
	250.0	250.6	± 1.52

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.
 *** END OF REPORT ***

- Note:** 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
 2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
 3. Total liability of our Organisation is limited to the invoiced amount.
 4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
 5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
 6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of
Indian Calibration Services

Checked by 
 Authorised Signatory
RAJESH KAPOOR
 Operation Head

PDF Created Using



Camera Scanner

Easily Scan documents & Generate PDF



<https://play.google.com/store/apps/details?id=photo.pdf maker>