

Issued from : Choksi Laboratories Limited  
 Survey No. 9/1, Balaji Tulsiyana Industrial Estate,  
 Kumed, Off. M.R. 10 Toll Naka, Indore -452010 (M.P.), INDIA

# CERTIFICATE OF CALIBRATION



CALIBRATION CERTIFICATE NUMBER :03900003/01/21-22

EQUIPMENT NAME :- MICROPIPETTE

RECEIVED ON	CALIBRATED ON	CALIBRATION DUE DATE RECOMMENDED	PAGE NO.
28.05.2021	01.06.2021	01.06.2022	1 of 1

I. Calibrated for :  
 PRESTIGE MEDICAL INVESTIGATION, INDORE  
 12/2, NEW AGRAWAL NAGAR, AGRASEN SQUARE  
 Indore, Madhya Pradesh  
 India - 452001

II. Description and Identification of Instrument :

Sample Id :	03900003/01
Make :	Thermo LabSystems
Model :	Not mentioned
Type :	Not mentioned
Serial No. :	Not mentioned

Capacity :	100 µl
Calibrated Range :	100 µl
Least Count :	Not applicable
Condition On Receipt :	Satisfactory
Identification No.:	Not Mentioned

III. Environment Condition :-

Temperature :	20 ± 2 °C
Relative Humidity :	50 ± 10 % RH

IV. Standard Calibrator and Measuring Equipment Used :

Nomenclature	Range	Uncertainty	Certificate No.	Traceability	Validity
Weight Box(E1)	1 mg to 200 g	± 0.001 to 0.03 mg	SCW-1120-268	CC-2813	05.11.2023
With Precision Balance, Distilled Water & Digital Thermometer					

V. Results: (a) Volume

Serial No.	Scale	UUC #	Standard Average	Correction	Uncertainty (For k=2, 95.45 % C.L.) : ±
1	µl	100.0	100.3927113	0.392711	0.054

Note : 01. UUC # : Unit Under Calibration

: 02. Certificate shall not be reproduced in full without the written approval of the laboratory

Other Details : 01. Certificate issue date : 02/06/2021  
 02. Location of Calibration : Lab  
 03. Temperature of Distilled water used for calibration is 20.1 °C.  
 04. The results relate only to the item calibrated

VI. Inference:

01. The correction is positive except when indicated other wise it has to be algebraically added to the UUC # reading in order to get the actual reading .

VII. Reference:

01. The Calibration is carried out as per the CLL Calibration Procedure  
 No. CLL-IDR-SWP-CAL-000045.  
 02. All the values have been rounded off as per IS:2-1960.  
 03. The laboratory is NABL Accredited in the field of Mechanical as per Certificate No. , CC2089. ULR NO.:CC208921000001231F

VIII. Details of Adjustment made ( If any)

Not Done

CALIBRATED BY IDR01183

AUTHORISED SIGNATORY

\*\*END OF CALIBRATION CERTIFICATE\*\*

Signature Not Verified  
 Mitesh J. Doshi  
 02.06.2021 15:44



Calibration Tag  
Valid Effective Date: 2019

5-01-Annex-02

Last Ver 4.0 Dated 01-04-2017

**CHOKSI**  
LABORATORIES LIMITED  
ANALYTICAL, TESTING & CALIBRATION LABORATORY GROUP

Next Review Due On: 01-04-2022

Equipment ID:

COC ID:

Equipment No.:

Calibrated On:  
(Date of Initiation)

Calibration Completed on:  
(Date of completion)

Recommended Due Date :

Valid Calibration Range :

Equipment Serial No.:

Signature:

Not mentioned

Not mentioned

03/03/2019

01/06/2021

01/06/2021

100 µg

Not mentioned

*[Signature]*

NOTE: Calibration Tag (to be displayed on the instrument after calibration, along with Equipment Tag, CL-MSP-5-5-01-05).  
Replace after every calibration. Attach the old calibration tags with its corresponding CoC (Certificate of Calibration).  
Give CoC as follows:  
CL - 3 Character Branch Code - 'Coc' - 3 Character Equipment ID / ddmmyy  
(For E.g. IDR-Coc-001/010415 represents calibration tag or Equipment ID '001' at CL's Indore Branch, on 01 APRIL 2015).