

CERTIFICATE OF CALIBRATION

Issued from : Choksi Laboratories Limited
 Survey No. 9/1, Balaji Tulsiyana Industrial Estate,
 Kumedl, Off. M.R. 10 Toll Naka, Indore - 452010 (M.P.), INDIA



CHOKSI
 LABORATORIES LIMITED

CALIBRATION CERTIFICATE NUMBER : 03900003/02/21-22

EQUIPMENT NAME : -		MICROPIPETTE	PAGE NO.
RECEIVED ON	CALIBRATED ON	CALIBRATION DUE DATE RECOMMENDED	1 of 1
28.05.2021	01.06.2021	01.06.2022	

I. Calibrated for : **PRESTIGE MEDICAL INVESTIGATION, INDORE**
 12/2, NEW AGRAWAL NAGAR, AGRASEN SQUARE
 Indore, Madhya Pradesh
 India - 452001

II. Description and Identification of Instrument :

Sample Id :	03900003/02
Make :	Thermo LabSystems
Model :	Not mentioned
Type :	Not mentioned
Serial No. :	Not mentioned

Capacity :	1000 µl
Calibrated Range :	1000 µl
Least Count :	Not applicable
Condition On Receipt :	Satisfactory
Identification No.:	Not Mentioned

III. Environment Condition :-

Temperature :	20 ± 2 °C
Relative Humidity :	50 ± 10 % RH

IV. Standard/Calibrator and Measuring Equipment Used :

Nomenclature	Range	Uncertainty	Certificate No.	Traceability	Validity
Weight Box(E1)	1 mg to 200 g	± 0.001 to 0.03 mg	SCW-1120-258	CC-2813	05.11.2023
With Precession Balance, Distilled Water & Digital Thermometer					

V. Results: (a) Volume

Serial No.	Scale	UUC #	Standard Average	Correction	Uncertainty (For k=2, 95.45 % C.L.) : ±
1	µl	1000.0	1002.9364274	2.936427	0.054

- Note : 01. UUC # : Unit Under Calibration
 : 02. Certificate shall not be reproduced in full without the written approval of the laboratory
- Other Details : 01. Certificate issue date : 02/06/2021
 02. Location of Calibration : Lab
 03. Temperature of Distilled water used for calibration is 20.2 °C.
 04. The results relate only to the item calibrated

VI. Inference:

01. The correction is positive except when indicated other wise It has to be algebraically ' added to the UUC # reading in order to get the actual reading .

VII. Reference:

01. The Calibration is carried out as per the CLL Calibration Procedure
 No. CLL-IDR-SWP-CAL-000045.
 02. All the values have been rounded off as per IS:2-1960.
 03. The laboratory is NABL Accredited in the field of Mechanical as per Certificate No. , CC-2089. ULR NO.:CC208921000001232F

VIII. Details of Adjustment made (If any) Not Done

CALIBRATED BY IDR01183

AUTHORISED SIGNATORY

Signature Not Verified
 Mitlesh J. Doshi IDR01151
 02.06.2021 14:46

END OF CALIBRATION CERTIFICATE



Ver 5.0 Effective Date: 01-05-2019
01-Annex-02

Last Ver 4.0 Dated 01-04-2017



Next Review Due On: 01-04-2022

Equipment ID: **Not mentioned**
COC ID: **03900003/02/25-22**
Equipment Name: **micropipette**
Calibrator Dn: **01/06/2021**
(Date of Initiation)
Calibration Completed on: **01/06/2021**
(Date of completion)
Recommended Due Date: **1000 Jul**
Valid Calibration Range: **Not mentioned**
Equipment Serial No.: **Not mentioned**
Signature: **[Signature]**

NOTE: Calibration Tag (to be displayed on the instrument after calibration, along with Equipment Tag, CLL-MSP-S-01-05),
Replace after every calibration. Attach the old calibration tags with its corresponding CoC (Certificate of Calibration).
Give CoC as follows:
CLL - 3 Character Branch Code - 'Coc' - 3 Character Equipment ID / ddnmYYYY
(For E.g. IDR-Coc-001/010415 represents calibration tag or Equipment ID '001' at CLL's Indore Branch, on 01 APRIL 2015).