

CERTIFICATE OF CALIBRATION

Issued from : Choksi Laboratories Limited
Survey No. 9/1, Balaji Tulsiyana Industrial Estate,
Kumedi, Off. M.J.R. 10 Toll Naka,
Indore -452010 (M.P.), INDIA



CALIBRATION CERTIFICATE NUMBER : 03900003/21-22

EQUIPMENT NAME : - CENTRIFUGE MACHINE

RECEIVED ON	CALIBRATED ON	CALIBRATION DUE DATE RECOMMENDED	PAGE NO.
28.05.2021	31.05.2021	31.05.2022	1 of 1

I. Calibrated for : **PRESTIGE MEDICAL INVESTIGATION, INDORE**
12/2, NEW AGRAWAL NAGAR, AGRASEN SQUARE
Indore, Madhya Pradesh
India - 452001

II. Description and Identification of Instrument :

Sample Id :	03900003
Make:	Remi
Model:	C-854/6
Type:	Regulator control
Serial No.	ZGBN-06244

Range:	Maximum 4000 rpm
Calibrated Range:	140 to 3500 rpm
Least Count:	Not Applicable
Condition on receipt:	Satisfactory
Identification No. :	Not Mentioned

III. Environment Condition-

Temperature :	25 ± 2 °C
Relative Humidity :	50±10 % RH

IV. Standard/Calibrator and Measuring Equipment Used :

Nomenclature	Range	Uncertainty	Certificate No.	Traceability	Validity
Digital Tachometer	10 to 25007 rpm	± 0.7 to 4 rpm	M-200629-21-6	CC-2634	18.07.2021

V. Results:

Serial No.	Scale	UUC [®] Regulator points	Standard Average	Correction	Uncertainty (For k=2, 95.45 % C.L.) : ±
1	rpm	1	141	-	0.42 % FS
2	rpm	2	802	-	0.42 % FS
3	rpm	3	2087	-	0.42 % FS
4	rpm	4	2523	-	0.42 % FS
5	rpm	5	3510	-	0.42 % FS

Note : 01. UUC[®] : Unit Under Calibration
: 02. Certificate shall not be reproduced in full without the written approval of the laboratory.
: 03. The results relate only to the item calibrated.

Other Details: 01. Certificate Issue Date : 02/06/2021
02. Location of calibration : Site

VI. Inference:

01. The correction is positive except when indicated, other wise it has to be algebraically added to the UUC[®] reading in order to get the actual reading.

VII. Reference

01. The Calibration is carried out as per the CLL Calibration Procedure No. CLL-IDR-SWP-CAL-00035.
02. All the values have been rounded off as per IS:2-1960.
03. The laboratory is NABL Accredited having certificate No. CC-2089. ULR NO. CC208921000001230F

VIII. Details of Adjustment Made (If any) Not Done

CALIBRATED BY IDR00910

AUTHORISED SIGNATORY

Signature Not Verified
Mitesh J. Gaur | IDR01151
02.06.2021 13:44

END OF CALIBRATION CERTIFICATE

CLL-MSP-5.5-01-Annex-02
Calibration Tag



Equipment ID: _____
CoC ID: Not Mentioned
Equipm Name: 03900003/21-22
Cal/brated On: Centrifuge
(Date of initiation) 31/05/2021
Calibration Completed on: 31/05/2021
(Date of completion) 31/05/2021
Recommended Due Date: 31/05/2022
Valid Calibration Range: _____
Equipment Serial No.: Z6BN-06244
Signature: [Signature]

NOTE: Calibration tags are to be maintained as per instructions and standards along with equipment. The tag and its contents should be used only for the purpose of calibration and not for any other purpose. The tag should be destroyed after the calibration is completed. The tag should be destroyed after the calibration is completed. The tag should be destroyed after the calibration is completed.

