

# CERTIFICATE OF CALIBRATION

Issued from : Choksi Laboratories Limited  
 Survey No. 8/1, Balaji Tulsiyana Industrial Estate,  
 Kumed, Off. M.R. 10 Toll Naka,  
 Indore - 452010 (M.P.), INDIA



**CHOKSI**  
 LABORATORIES LIMITED

CALIBRATION CERTIFICATE NUMBER 03808003-03/21-22

<b>EQUIPMENT NAME :</b> GLASS THERMOMETER		<b>PAGE NO.</b>
<b>RECEIVED ON</b>	<b>CALIBRATED ON</b>	<b>CALIBRATION DUE DATE RECOMMENDED</b>
28.05.2021	01.06.2021	01.06.2022
		1 of 1

**I. Calibrated for :** PRESTIGE MEDICAL INVESTIGATION, INDORE  
 12/2, NEW AGRAWAL NAGAR, AGRASEN SQUARE  
 Indore, Madhya Pradesh  
 India - 452001

**II. Description and Identification of Instrument :**

Sample Id:	03900003-03
Make:	Merck
Model:	Not mentioned
Type:	Not mentioned
Serial No.	Not mentioned

Range:	0 to 50 °C
Calibrated Range:	0 to 50 °C
Least Count:	1 °C
Condition on receipt:	Satisfactory
Identification No. :	Not mentioned

**III. Environment Condition-**

Temperature :	25 ± 4 °C
Relative Humidity :	30 to 75 % RH

**IV. Standard Calibrator and Measuring Equipment Used :**

Nomenclature	Range	Uncertainty	Certificate No.	Traceability	Validity
PRT Sensor With Indicator with Temperature sources.	-196 to 650 °C	± 10.34 to 12.53 m °C	M-201130-11-1	CC-2634	05.12.2021

**V. Results: Thermal Calibration (Temperature)**

Serial No.	Unit	UUC <sup>®</sup> Average	Standard Average	Correction	Uncertainty (For k=2, 95.45 % C.L.) : ±
1	°C	0	0.0	0.0	0.58
2	°C	10	10.1	0.1	0.58
3	°C	20	20.1	0.1	0.58
4	°C	30	30.3	0.3	0.58
5	°C	50	50.4	0.4	0.58

**Note :** (1) UUC<sup>®</sup> - Unit Under Calibration

(2) Certificate shall not be reproduced in full without the written approval of the laboratory

**Other Details**

- 01. Certificate issue date : 02/06/2021
- 02. Location of calibration Lab
- 03. The results relate only to the item calibrated

**VI. Inference:**

01. The correction is positive except when indicated other wise it has to be algebraically added to the UUC<sup>®</sup> reading in order to get the actual reading.

**VII. Reference:**

- 01. The Calibration is carried out as per the CLL Calibration Procedure No. CLL-IDR-SWP-CAL-01B as per IS 6274-1971.
- 02. All the values have been rounded off as per IS 2-1960
- 03. The laboratory is NABL Accredited in the field of Thermal Calibration as per Certificate No. CG-2089, ULR No. CG208921000001231F

**VIII. Details of Adjustment Made (If any)**

Not Done

**CALIBRATED BY** IDR00910

**AUTHORISED SIGNATORY**

Signature Not Verified  
 Nitish J. Datta IDR01151  
 02.06.2021 11:44

\*\*\*\*End of Calibration Certificate\*\*\*\*



Emflam - Plus  
Neurobion Merck<sup>®</sup>

CLL-MSP-5.5-01-Annex.-02

Calibration Tag



Ver 5.0 Effective Date: 01-04-2012

Last Ver 4.0 Dated 01-04-2017

Next Review Due On: 01-04-2022

Equipment ID:

Not mentioned

Coc ID:

03900003-03121-22

Equipment Name:

Glass Thermometer

Calibrated On:

01/06/2021

(Date of Initiation)

Calibration Completed on:

01/06/2021

(Date of completion)

Recommended Due Date:

01/06/2022

Valid Calibration Range:

Refer Coc

Equipment Serial No:

Not mentioned

Signature:

*[Handwritten Signature]*

NOTE: Calibration Tag (to be displayed on the instrument after calibration, along with Equipment Tag, CLL-MSP-5.5-01-05). Replace after every calibration. Attach the old calibration tags with its corresponding Coc (Certificate of Calibration).

Give Coc as follows:

CLL-3 Character Branch Code - Coc-3 Character Equipment ID / ddmmyyy

(For E.g. IDR-Coc-001/010415 represents calibration tag or Equipment ID '001' at CLL's Indore Branch, on 01 APRIL 2015).