

11/12/2021 11:35 PM

External Quality Assurance Scheme - Print Monthly Summary

LAB MONTHLY SUMMARY



Lab Name DR DHAM S DIAGNOSTIC CENTRE Lab No 11089
 Month October Year 2021
 Constituent Group Chemistry II



Date of Result Entered : 21/10/2021

Date of Report Published : 01/11/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Mindray	154	109.15	10.14	11.07	116.6 mg/dl	0.67	1.78
2	UREA II	Urease UV / GLDH II	Mindray	124	26.65	11.74	3.13	13.3 mg/dl	-4.27	0.56
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Mindray	95	1.33	11.04	0.15	1.33 mg/dl	0.00	0.03
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Mindray	131	7.14	11.17	0.80	6 mg/dl	-1.43	0.14
5	T-PROTEIN II	Biuret - Colorimetric II	Mindray	138	5.16	8.55	0.44	4.7 g/dl	-1.04	0.08
6	ALBUMIN II	BCG - Colorimetric II	Mindray	137	3.16	8.45	0.27	2.2 g/dl	-3.60	0.05
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Mindray	130	5.84	13.84	0.81	6.3 mg/dl	0.57	0.14
8	CHOLESTROL II	CHOD-PAP II	Mindray	143	104.86	9.87	10.35	132 mg/dl	2.62	1.73
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Mindray	116	99.58	12.21	12.15	84.6 mg/dl	-1.23	2.26

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss

CMC EQAS Co-Ordinator

Christian Medical College, Vellore

***** End of Report *****



Dr. Dham's Diagnostic Centre

A Specialist Centre

3856/23 Reghar Pura, Arya Samaj Road, Karol Bagh,
New Delhi-110005 Ph.: 011-25852890, 25852891

CAPA for the EQAS send in OCTOBER-2021

The parameters Urea , Albumin, Cholesterol were recalibrated.

QC were within normal range.

Raw data is attached.


ILC was done from NABL accredited lab GAGAN.

The reports are attached along.

DATED - 06/11/2021

Senior Lab Technician

Dr. GURMIT KAUR
MBBS, D. PATH (DMC No. 3825)
Consultant Biochemist


Dr. JASPREET KAUR
MBBS, MD (DMC No. 15396)
Consultant Biochemist

HOST

Admin

11/18 03:34 PM

Program

Levey-Jennings

Cumulative Sum

Twin-Plot

Results

Summary

QC Setup

Control MED P 971

QC Date 2021/11/06

Run Date/Time 2021/11/06 12:15:22

Lot No.	Chem	Sample Type	Lot No.	Result	Unit	Set Mean	Set SD	Flag	Run Date/Time
971	GLUC	Serum		247.1	mg/dL	202.31	13.12		2021/11/06 12:15:22
971	UREA	Serum		100.3	mg/dL	98.89	4.94		2021/11/06 12:10:16
971	SGPT	Serum		128.6	U/L	144.1	9.61		2021/11/06 12:11:45
971	SGOT	Serum		145.3	U/L	146.76	9.78		2021/11/06 12:12:03
971	CHLO	Serum		251.7	mg/dL	264.18	13.21		2021/11/06 12:13:27
971	HDL	Serum		79.09	mg/dL	86.51	5.77		2021/11/06 12:19:15
971	TC	Serum		173.53	mg/dL	170.43	8.52		2021/11/06 12:15:03
971	Ca +	Serum		13.2	mg/dL	10.61	0.53	2.7s	2021/11/06 12:14:45
971	Tp	Serum		9.95	g/dL	9.12	0.46		2021/11/06 12:19:51
971	ALB	Serum		5.56	g/dL	5.02	0.25		2021/11/06 12:16:53
971	BIL T LRB	Serum		6.11	mg/dL	4.97	0.41	2.7s	2021/11/06 12:21:21
971	BIL D LRB	Serum		1.61	mg/dL	1.59	0.13		2021/11/06 12:21:39

Search F1

Charts F2

Sort F3

Freeze Graph F4

Comments F5

Archive F6

Print F7

>

REDM NOTE 10 LIS communication error

Dr. Dham's Diagnostic Centre

A Specialist Centre

3856/23 Reghar Pura, Arya Samaj Road, Karol Bagh,
New Delhi-110005 Ph.: 011-25852890, 25852891

Date 06/11/2021
Name MS. SHERRIN
Ref. By SELF

Srl No. 29
Age 19 Yrs.
Address


Ref. No
Sex F

Test Name	Value	Unit	Normal Value
<u>BIOCHEMISTRY</u>			
BLOOD UREA	21.6	mg /dl	10 - 50
ALBUMIN	4.6	gm/dl	3.4 - 4.8
TOTAL CHOLESTEROL	164.00	mg/dL	123.0 - 199.0

**** End Of Report ****

Senior Lab Technician

Dr. GURMIT KAUR
MBBS, D. PATH (DMC No. 8825)
Consultant Biochemist


Dr. JASPREET KAUR
MBBS, MD (DMC No. 15396)
Consultant Biochemist

Ph. : 011-25852890, 25852891, E-mail : dr.kawaldeepdham123@gmail.com

(REPORTS ARE NOT FOR MEDICO LEGAL PURPOSE, IT IS ONLY A PROFESSIONAL OPINION)



GAGAN PATHOLOGY CENTRE

MRI, CT, ULTRASOUND, ECHO, TMT, EEG, EMG
MAMMOGRAPHY, NCV, DEXA SCAN, OPG, ECG
PFT, BERA, VEP, DIGITAL & PORTABLE X-RAYS

NABL Lab & NABH Accredited CT/MRI Imaging

Date 06/11/2021

Srl No. 30

Ref. No

Name MS. SHERRIN

Age 19 Yrs.

Sex F

Ref. By DR.DHAM

Address

Test Name	Value	Unit	Normal Value
-----------	-------	------	--------------

BIOCHEMISTRY

BLOOD UREA	22.0	mg /dl	10 - 50
ALBUMIN	4.8	gm/dl	3.4 - 4.8
TOTAL CHOLESTEROL	162.00	mg/dL	123.0 - 199.0

**** End Of Report ****

Dr. R.K. Garg
MBBS, MD
Senior Consultant Pathologist

Page 1 of 1

Dr. Ankur Garg
MBBS, MD
Consultant Pathologist

Expert and Reliable Opinion for over 25 Years

H.O. : F-26/21-22, Sector-7, Rohini, Delhi-110085 Email : gaganlab@yahoo.com website : www.gaganlab.com

On Panel : CGHS, D.G.E.H.S., DJB, NDPL, ECHS, DDA, ITPO, NDMC, MCD, LIC (TPA), ESI, DELHI UNIVERSITY

Terms & conditions governing the test report issued.

1. *The reported results are for information and interpretation of the referring doctor or such other medical professionals who understand reporting units, reference ranges and limitations of technologies. Gagan Pathology Centre will not be responsible for any wrong assumption or interpretation whatsoever.*
2. *It is presumed that the test(s) performed are on the specimen(s) sample(s) belonging to the patient named or identified and the verification of the particulars have been carried out by the patient or his/her representative at the point of generation of the said specimen(s) or sample(s).*
3. *The results of test may vary from lab to lab and also from time to time for the same parameters for the same patient. Assays are performed in accordance with standard procedures. The reported results are dependant on individual assay methods, equipment used, method specificity, sensitivity, drug interaction and the quality of the specimen(s), sample(s) received.*
4. *The report will only help in arriving at a diagnosis in conjunction with clinical presentation and related investigation. The report therefore, expected to be interpreted by taking into consideration the same.*
5. *Should the results indicate an unexpected abnormality, reconfirmation shall be sought. A repeat investigation may be entertained within 72 hours from the time of testing on request by the client.
Helpline Numbers : 011-27052705, 011-27048225, 9968198065, 9971709244, 9910353783.*
6. *Histopathology specimen(s) / sample(s) will be preserved for one month from the date of the testing. Other clinical specimen(s) / sample(s) will be discarded after testing unless otherwise specified by the client during sample registration, for preservation for longer period. Such preservation shall be subject to sample integrity feasibility and on payment of additional fee for the same.*
7. *Neither Gagan pathology Centre nor its directors / employees / representatives assume any liability, responsibility for any loss or damage that may be incurred by any person including the patient, as a result of assuming the meanings or contents of the report.*
8. *This report is not valid for medicolegal purpose.*
9. *All disputes are subject to Delhi, India Jurisdictions.*
10. *This is to certify that the results of this report are kept confidential and disclosed to only patient or referring doctor or relatives of patient on request.*

LABORATORY

- 1) Advia centaur (Siemens) XP
- 2) ROCHE COBAS 6000
- 3) Immulite 1000 (siemens)
- 4) PICTUS 500 Biochemistry Analyser
- 5) ROCHE COBAS INTEGRA
- 6) D-10 HPLC by Biorad
- 7) RAL electrophoresis system & many more

EQUIPMENTS USED

- 8) Sysmex 5 part XS-800i
- 9) Tulip Analyser
- 10) Elisa reader
- 11) Sysmex KX-21 cell counter
- 12) Bactac Blood culture (BD)
- 13) Bactac AFB culture (BD)
- 14) Real Time PCR-Exicycler

IMAGING

- 1) GE EXCITE- MRI
- 2) Multi Slice GE Lightspeed CT
- 3) GE LOGIQ P6 ULTRASOUND
2D, 3D, 4D, ECHO
- 4) 500mA Digital Xray
- 5) OPG, DEXA Mamography
- 6) ECG, EMG/NCV

15169

NEW LAB MONT

LAB MONTHLY SUMMARY

Lab Name **DR DHAM S DIAGNOSTIC CENTRE**
 Lab No **11089**
 Month **September**
 Year **2021**
 Constituent Group **Chemistry II**

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 20/09/2021

Date of Report Published : 05/10/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	GLUCOSE II	GOD-POD II	Mindray	143	403.99	11.17	45.12	490 mg/dl	1.91	7.55
2	UREA II	Urease UV / GLDH II	Mindray	112	91.16	8.16	7.44	94 mg/dl	0.38	1.41
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Mindray	87	3.35	10.93	0.37	2.4 mg/dl	-2.60	0.08
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Mindray	125	2.12	13.90	0.30	2.3 mg/dl	0.61	0.05
5	T-PROTEIN II	Biuret - Colorimetric II	Mindray	121	5.13	7.50	0.38	5.5 g/dl	0.96	0.07
6	ALBUMIN II	BCG - Colorimetric II	Mindray	124	3.21	8.83	0.28	2.2 g/dl	-3.56	0.05
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Mindray	122	8.39	12.38	1.04	9.9 mg/dl	1.45	0.19
8	CHOLESTROL II	CHOD-PAP II	Mindray	138	113.82	10.26	11.68	120 mg/dl	0.53	1.99
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Mindray	126	230.68	9.14	21.09	220 mg/dl	-0.51	3.76

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

Dr. Dham's Diagnostic Centre

A Specialist Centre

3856/23 Reghar Pura, Arya Samaj Road, Karol Bagh,
New Delhi-110005 Ph.: 011-25852890, 25852891

CAPA for the EQAS Send in SEPTEMBER-2021

The parameters Creatinine, Albumin, were recalibrated.

QC were within normal range.


Raw data is attached.

ILC was done from NABL accredited lab GAGAN.

The reports are attached along.

Senior Lab Technician

Dr. GURMIT KAUR
MBBS, D. PATH (DMC No. 3825)
Consultant Biochemist


Dr. JASPREET KAUR
MBBS, MD (DMC No. 15396)
Consultant Biochemist

Ph. : 011-25852890, 25852891, E-mail : dr.kawaldeepdham123@gmail.com

(REPORTS ARE NOT FOR MEDICO LEGAL PURPOSE, IT IS ONLY A PROFESSIONAL OPINION)



Dr. Dham's Diagnostic Centre

A Specialist Centre

3856/23 Reghar Pura, Arya Samaj Road, Karol Bagh,
New Delhi-110005 Ph.: 011-25852890, 25852891


Date	05/10/2021	Srl No.	69	Ref. No.	
Name	MR. OM PRAKASH	Age	60 Yrs.	Sex	M
Ref. By	SELF	Address			

Test Name	Value	Unit	Normal Value
<u>BIOCHEMISTRY</u>			
SERUM CREATININE	2.10	mg/dl	0.6 - 1.4
ALBUMIN	3.6	gm/dl	3.4 - 4.8

**** End Of Report ****

Senior Lab Technician

Dr. GURMIT KAUR
MBBS, D. PATR (DMC No. 3825)
Consultant Biochemist


Dr. JASPREET KAUR
MBBS, MD (DMC No. 15396)
Consultant Biochemist

Ph. : 011-25852890, 25852891, E-mail : dr.kawaldeepdham123@gmail.com

(REPORTS ARE NOT FOR MEDICO LEGAL PURPOSE, IT IS ONLY A PROFESSIONAL OPINION)



The Association of Practising Pathologists (Regd.)

GAGAN PATHOLOGY CENTRE

MRI, CT, ULTRASOUND, ECHO, TMT, EEG, EMG
MAMMOGRAPHY, NCV, DEXA SCAN, OPG, ECG
PFT, BERA, VEP, DIGITAL & PORTABLE X-RAYS

NABL Lab & NABL Accredited CT/MRI Imaging

Date: _____
Name: **MR. OM PRAKASH**
Ref. By: **DR. DHAM**

Sri No. **70**
Age **60 Yrs.**
Address _____

Ref. No _____
Sex **M**

Test Name

Value

Unit

Normal Value

SERUM CREATININE
ALBUMIN

BIOCHEMISTRY

2.30 mg/dl
3.7 gm/dl

0.6 - 1.4
3.4 - 4.8

**** End Of Report ****

Dr. R.K. Garg
MBBS, MD
Senior Consultant Pathologist

Page 1 of 1

Dr. Ankur Garg
MBBS, MD
Consultant Pathologist

NABL Accredited. Tests without # are out of Scope.

Expert and Reliable Opinion for over 25 Years

H.O : F-26/21-22, Sector-7, Rohini, Delhi-110085 Email : gaganlab@yahoo.com website : www.gaganlab.com

On Panel : CGHS, D.G.E.H.S., DJB, NDPL, ECHS, DDA, ITPO, NDMC, MCD, LIC (TPA), ESI, DELHI UNIVERSITY