

**CALIBRATION CERTIFICATE**

**Customer Name & Address:-**

**Date:- 02.09.2021**

Likhitha's Diagnostics and Speciality Lab,  
(A Unit of Pragnashree Health Care Private Limited)  
H. No. 1-8-65/A, North Kamalanagar, Above Andhra Bank,  
FCH, Hyderabad, Telangana State, Pin -500062

**Instrument Name:- Urine analyzer**

**Model:- QDX Urilyzer**

**Serial No:- NI3001001H1172SE**

**Power Supply Check List**

	Required Voltage	Observed Voltage
Point1	+5V	<input type="text" value="+5 V"/>
Point2	-12V	<input type="text" value="-12 V"/>
Point 3	+12V	<input type="text" value="+ 12 V"/>

**Main Board Check List**

	Required Voltage			Observed Voltage
Pin JP 3	+5V	-12V	+12V	<input type="text" value="+5 V"/> <input type="text" value="-12 V"/> <input type="text" value="+ 12 V"/>
Pin JP 10	+12V			<input type="text" value="+ 12 V"/>
Pin JP 8	+5V			<input type="text" value="+5 V"/>
Pin JP 6	+12V all Pins			<input type="text" value="+ 12 V"/>

**A.D Board Check List**

	Required Voltage	Observed Voltage
JP 1	+5V (First Pin)	+5 V

**Calibration**

**Successful**

**Not Successful**



For RAPHA DIAGNOSTICS PVT. LTD

Authorized Signatory



Next Calibration due on **01.09.2022**