



Calibration Certificate

This is to certify that the **Cobas E411** is a fully automated analyzer bearing the Serial Number **16HO – 21** at Likhitha's Diagnostics and Speciality Lab, H. No. 1-8-65/A ECIL, Hyderabad – 500062, has been calibrated on **19.08.2021**.

The calibration includes:

Adjustments: Checked and adjusted sample/Reagent Rotor, Incubator, Sample/Reagent probe, Bead Mixer and their respective home positions.

Incubation Disk: Range: 36.8°C to 37.2°C
Adjusted: 37°C

Detection Unit: Target: 28.0°C
Adjusted: 28.0°C

PC/CC: Target: 28°C
Adjusted: 28°C

Adding to this all the temperatures, PMT Voltage and volumes drawn by all pumps were checked and found they are OK.

Next calibration due on 18.08.2022

For Roche Diagnostics India Pvt Ltd.,

T R Siddardha,
Sr.Technical Service Specialist,
Hyderabad.